

**UNITED STATES DISTRICT COURT  
SOUTHERN DISTRICT OF WEST VIRGINIA  
AT CHARLESTON**

<b>IN RE: ETHICON, INC., PELVIC REPAIR SYSTEM PRODUCTS LIABILITY LITIGATION</b>  <b>THIS DOCUMENT RELATES TO WAVE 1 CASES</b>	<b>Master File No. 2:12-MD-02327</b>  <b>JOSEPH R. GOODWIN U.S. DISTRICT JUDGE</b>
---	--

**RULE 26 EXPERT REPORT OF MICHAEL THOMAS MARGOLIS, M.D.**

**I. BACKGROUND AND QUALIFICATIONS**

My curriculum vitae, attached hereto as Exhibit "A", more fully reflects my training, background, and publications, including those for the past ten years. I graduated from the University of Kansas Medical School in Kansas City, Kansas in 1987. I am licensed to practice medicine in the states of Wisconsin and California, where I also hold a valid California fluoroscopy x-ray supervisor and operator certificate. I did my internship in internal medicine for one year followed by three years of obstetrics and gynecology residency training under the direction of the father of urinary incontinence surgery, Dr. Kermit E Krantz, M.D. (co-founder of the MMK procedure which is predecessor to the Burch procedure(still considered the gold standard procedure for the treatment of urinary incontinence)). After completing my residency, I became one of only a handful of fellowship-trained pelvic surgeons in the country to complete the pelvic surgery fellowship at Emory University in Atlanta, Georgia where I trained under the World-renowned Johns Hopkins surgeon, Dr. Clifford Wheelless.

I currently hold hospital privileges at Mills Peninsula Medical Center in Burlingame, California, Good Samaritan Hospital in San Jose, ValleyCare Hospital in Pleasanton, Seton Medical Center in Daly City, Sequoia Hospital in Redwood City, El Camino Hospital, O'Connor Hospital in San Jose, Community Hospital of the Monterey Peninsula in Monterey, and several surgery centers. I currently serve as Vice Chairman in the Department of Obstetrics and Gynecology at Peninsula Hospital in Burlingame, California.

I am Board Certified in Obstetrics and Gynecology by the American Board of Obstetrics and Gynecology. I recently took and passed the first subspecialty board certification exam in the subspecialty of Female Pelvic Medicine and Reconstructive Surgery, thus becoming a member of the first class of Board-Certified Pelvic Surgeons in the country. I am a fellow in good standing of the American College of Obstetrics and Gynecology and the American College of Surgeons and I am a member of multiple societies including: the American Urogynecologic Society, the Society of Pelvic Reconstructive Surgeons, the Northern California Chapter of the American College of Surgeons and the Kermit E. Krantz Society.

I am District VII delegate to the California Medical Association. I have served at various hospitals as member of various medical executive committees including service as chairman of the OB/GYN department at Los Gatos Community Hospital, quality management committees at Northwestern University Hospital, Stanford Hospital, Peninsula Hospital and several hospitals in Wisconsin. I have served as president of the Peninsula Gynecologic Society and the Shufelt Gynecologic Society in San Jose, California.

I have held faculty positions in the Department of Obstetrics and Gynecology at Emory University in Atlanta Georgia, Northwestern University in Chicago, Illinois, and Stanford University in California, where I also served as chief of the division of gynecology and founder

and director of The Stanford Center for Pelvic Reconstructive Surgery and Urogynecology. I have held clinical faculty appointments in the Department of Obstetrics and Gynecology at Stanford University and the University of Wisconsin in Milwaukee where I received the teaching excellence award in 2009.

I have been and continue to be actively engaged in the education and training of medical students, resident physicians, fellows and faculty at numerous universities and I currently teach on a daily basis and hold a clinical faculty position in the Department of Obstetrics and Gynecology at the David Geffen School of Medicine at UCLA, the Department of Obstetrics and Gynecology at Mbarara University in Uganda (where I also hold the title of visiting lecturer), and I am currently gynecology rotation site director at Valley Medical Center., Department of Obstetrics and Gynecology residency program in San Jose, California. I have performed animal testing on many novel surgical instruments and I have performed field testing in humans for Boston Scientific sling anchor instruments in the past.

I have published numerous peer-reviewed and non-peer-reviewed publications in the field of obstetrics and gynecology and have written several books chapters in the field as well. These publications including those in the past 10 years are listed in my C.V. (Exhibit A). I have lectured extensively throughout the US and abroad on numerous subjects involving sling and mesh complications and advanced pelvic surgery, including the treatment of prolapse and incontinence. I have presented original scientific peer reviewed articles at the Society of Gynecologic Surgeons and I have presented surgical videos at the American College of Obstetricians and Gynecologists annual meeting. I have served as reviewer for the Journal of Reproductive Medicine and the International Journal of Obstetrics and Gynecology.

Pursuant to a request by the FDA, on September 8, 2011, I testified at the FDA hearing for the Obstetrics and Gynecology Devices Panel in Gaithersburg, Maryland on the serious issue of complications of transvaginal synthetic mesh placement. My address was based on my knowledge, experience, education, and training as a pelvic surgeon and on observations made during scores of salvage operations I have performed on women who have experienced mesh and sling complications since 1996. During my address, I made the analogy that extirpation of vaginal mesh was akin to using a hammer and chisel and trying to remove the rebar from a sidewalk while leaving the cement otherwise intact and not damaging the water mains and power lines below. Since my FDA testimony and based upon my further experience with slings and mesh, I have found that it may even be more dramatic and challenging than that.

I am medical director of Medlend Medical Missions to Developing Countries, a nonprofit organization that provides free surgical care to women in need around the globe. I have participated in IO different surgical missions throughout Ghana, Peru, and Uganda. I am team leader of the yearly fistula mission to Uganda where I direct trainees from around the world in the repair of complex gynecologic injuries, often related to pelvic organ prolapse and incontinence.

I am in private practice of Pelvic Surgery and Urogynecology in the Bay Area of California. I perform on average 10 major surgical procedures per week. I have performed over 25,000 surgical procedures in my professional career, including approximately 1,500 organic (non-synthetic) sling procedures. The majority (90%) of my patients are referred to me by other gynecologists for complications associated with prior gynecologic surgery, prolapse, incontinence, fistulas and polypropylene or other synthetic mesh and sling related injuries. Since 1996, I have removed over 300 mesh or sling systems in patients who have suffered numerous

complications. Although when removing mesh, I do not always know what particular mesh was implanted. In the early years, as my expertise at explant surgery developed, in order to assist in the removal, I would gain access to the implant records and to learn which particular synthetic mesh was involved. In this regard, I would also review the applicable Instructions for Use (IFU). Frequently, the product has been one of the TVT products manufactured by J&J/Ethicon, including the product at issue in this case.

Within the last 1 to 2 years, the number of patients with synthetic mesh or sling related complications presenting to me for definitive surgical treatment has increased dramatically. I now remove on average two sling or mesh systems per week and the number continues to rise. Patients are referred to me for sling and mesh complications from across the country. These patients include women with the TVT Obturator device ("TVTO").

In order to more fully understand how to deal with the complications of these synthetic mesh systems and to remove them as effectively as possible, I have personally observed numerous sling and mesh insertion procedures by colleagues of mine and through industry videos from a variety of manufacturers, including J&J/Ethicon for their TVT products. As well, I have studied textbooks, publications, IFU's (including those from J&J/Ethicon for TVT and TVTO), surgical videos, cadaver dissections and countless operative reports as part of my study of sling and mesh surgeries. I have conferred with colleagues at conferences and in person throughout the country and abroad on synthetic mesh related complications.

Based on this education and experience, I am uniquely qualified and experienced as an expert in synthetic sling and mesh related issues and complications and can provide absolute standard of care testimony on all clinical aspects of this subject matter.

## II. OVERVIEW OF INCONTINENCE AND TREATMENT HISTORY

The average age of menopause in US women is 50 years and the life expectancy is 80 years. Thus, women live more than one third of their lives in the postmenopausal years. US Census data from 2006 show that of the 300 million US citizens one half were women 16 million of whom were women age of 45 and older. Between 1995 and 2020, the Census Bureau predicts an increase in the number of women 45 years old and over of more than 25 million. Since the majority of incontinence and prolapse occurs in the late and post-reproductive age group pelvic floor defects and incontinence will continue to affect a large number of women now and well into the future. This population provides an attractive target for marketing campaigns by device manufacturers seeking to capture a high market share for devices used in the treatment of these conditions.<sup>1</sup>

Almost 200 procedures have been described in the literature for the treatment of stress incontinence alone. These procedures can be broken down into seven major categories including:

1. Anterior repair
2. Needle suspension procedures
3. Retropubic urethropexy - Open vs. scope
4. Sling operations - Traditional vs. new
5. Perry urethral injections
6. Artificial sphincter
7. Urinary diversion

The first sling procedure for SUI was described in 1907. This procedure, later called the Goebell-Frankenheimer-Stoeckel procedure utilized the patient's own ligaments to support the return. Its only indication was for type III stress incontinence which is also known as intrinsic sphincter deficiency (ISD). ISD is characterized by an absence of urethral hypermobility, low urethral pressures and is typically seen in patients with a history of previous urethral surgery. By

---

<sup>1</sup> Wall, The Perils of Commercially Driven Surgical Innovation, Am J. of Obstetrics and Gyne Jan 2010; 202:30e1-4.

definition, synthetic slings are obstructive procedures which are always done blindly, thus increasing the risk of injury to adjacent organs.

In 1998, Ethicon's TVT Obturator device ("TVTO") was cleared for marketing. TVTO is a suburethral sling used for the treatment of female stress urinary incontinence (SUI). The TVTO kit includes the actual TVTO device, which is a polypropylene mesh sling attached to needles, and surgical tools for implanting the sling. The TVTO sling itself is made from Prolene® polypropylene mesh, which is constructed of knitted filaments of extruded polypropylene strands.

### **III. COMPLICATIONS ASSOCIATED WITH THE TVTO**

In my opinion from my professional experience and research, there are a number of serious and often permanent complications associated with implanting a polypropylene synthetic mesh like the TVTO to treat incontinence. I have reached these opinions based on my personal experience explanting meshes, through my personal research, through my review of Ethicon's internal documents, and my review of the depositions and experts reports that have been provided to me. Unfortunately, it is now well known that the use of polypropylene mesh slings like the TVTO are associated with serious and often permanent complications. These complications can include:

1. Development of abnormal tissue between the posterior urethra and the anterior vaginal wall;
2. Thickening and disfigurement of the vaginal wall;
3. Scar plate formation;
4. Distal urethral narrowing;
5. Temporary and permanent foreign body giant cell reaction;
6. Temporary and permanent chronic, debilitating pelvic pain;
7. Temporary and permanent dyspareunia;
8. Urinary problems, including: Urethral injury, Voiding dysfunction, De novo detrusor instability or urgency, Urinary retention/obstruction, Urinary retention (temporary or permanent), Urinary tract infection, Dysuria, Hematuria, and Worsening or recurrence of incontinence;

9. Erosions, including into surrounding tissue and organs, that can recur and require numerous surgeries;
10. Perforation of the surrounding organs;
11. Injury to partner during intercourse;
12. Serious and chronic inflammation, that is not slight or transient;
13. Chronic foreign body response;
14. Permanent nerve injury;
15. Defecatory problems, including chronic debilitating pain;
16. Mesh degradation;
17. Cytotoxic;
18. Toxic shock syndrome;
19. Difficulty treating infections, including de novo urinary tract infections that do not resolve with aggressive treatment or removal of the mesh;
20. Vaginal scarring which is greater than disclosed in its IFU;
21. Sarcomas at the site of implantation;
22. Permanent and debilitating leg pain.

When these complications arise, which they frequently do, additional surgery may be needed. A revision surgery can be as simple as cutting extruded mesh or much more invasive such as an explant surgery. Each of these additional carries additional risks for the patient, especially due to the scarring that has resulted from the initial implant procedure. As a result of the implantation and subsequent removal of a TVTO sling, a patient can be injured in many ways, including the need for additional surgeries, increased scarring, damage and deformity to her vagina, voiding symptoms, and recurrence of her SUI or de novo incontinence.

Synthetic sling and mesh procedures expose patients to triple the surgical risk of non-mesh repairs. Compared to a repair surgery that does not use a synthetic mesh, when complications arise, the patient with a synthetic sling implant has to then undergo another risky surgery to remove the mesh.

Revision surgeries further complicate a bad situation. Even if the mesh is completely removed, which is difficult, the patient then still suffers from the original problem of incontinence but with the addition of a damaged and scarred pelvic region. For subsequent surgery under these circumstances, a surgeon faces an operative field that is damaged, thus



increasing the risk of complications and a higher failure rate of repair. In his textbook, *Reoperative Gynecologic Surgery*, 1991, David Nichols states the obvious, “It is important that every effort be made to ensure the success of the first operative procedure undertaken as treatment of SUI. This is especially true since it is the first operative procedure that is most likely to cure a patient's SUI. Subsequent procedures have failure rates that increase in proportion to their number.”

Scar tissue is a permanent condition and will probably be problematic for the rest of the patient's life. A patient who has undergone an implant surgery and then subsequent revision surgery due to complications from the implant will likely suffer from a permanently disfigured vagina with permanently diminished function. For the patient, this translates into never regaining function of her vagina, never being completely pain free during intercourse, and never being completely free of pelvic pain.

These risks are well known now. Numerous Ethicon employees have testified that they were aware of the risks discussed above, including Ethicon Medical Directors Dr. Piel Hinoul, Dr. David Robinson, Dr. Martin Weisberg, and Axel Arnaud.<sup>2</sup> Long before the TVTO was launched, internal documents reflect that all of the risks associated with implanting a synthetic mesh such as the TVTO were known. In fact, Ethicon updated the warnings in the TVTO Instructions for Use in 2015 to include many of these risks.

#### **IV. CAUSES OF THE COMPLICATIONS**

In my opinion from my professional experience and research, the above adverse events are caused by one or a combination of defects in the mesh or techniques associated with the use of Ethicon's TVTO as described below. I have reached these opinions based on my personal

---

<sup>2</sup> (Robinson (July 25, 2013) 448:4 - 451:12, (September II, 2013) 28:1-288:20; 293:21-293: 11 298; 18-24, 305:20-306:2, 306:7-23;329: 12-331:20, 1136:23 - 1140: 17; Hinoul (June 27, 2013) 547:11 - 557:7; Weisberg (August 9, 2013) 930: 3-8, 951:6-10, 951:11-16; 944; 16-945:5; 947:4-19; 968:12-972:21, Arnaud (July, 19, 2000) 66:8 to 66:19, 114:21 to 127:11); ETH.MESH .06372356 - ETH.MESH .06372363.

experience explanting meshes, through my personal research, through my review of Ethicon's internal documents, and my review of the depositions and experts reports that have been provided to me.

**A. The TVTO Causes a Chronic Foreign Body Response**

Mesh, when implanted into the vagina, is a foreign body. It is the human body's natural response to react to the presence of the mesh. A foreign body response is characterized by inflammation, soreness and pain at the site of the foreign body – here, the implanted mesh. The Ethicon Instructions for Use (“IFU”) recognize that this reaction is likely to occur. The IFU, however, described this response as “transient.”<sup>3</sup> This is not consistent with what I have seen in my professional experience, nor with Ethicon's own internal documents.<sup>4</sup>

I have treated a number of women who have suffered from chronic long-term foreign body response to implanted mesh. These women often present with severe inflammation at the site of the implant that is persistent and painful. This inflammatory response can, in turn, lead to deformation of the mesh and damage to the tissues surrounding the mesh. Severe inflammation also leads to shrinkage and contraction of the mesh and the surrounding tissues and erosion.<sup>5</sup>

Severe inflammation can also lead to degradation of the mesh itself.<sup>6</sup> Ethicon's scientists have demonstrated in animal studies that the mesh used in the TVTO products degrades for years

---

<sup>3</sup> Until 2010, the IFU stated that the mesh was: “non-reactive with a minimal and transient foreign body reaction.” In 2010, Ethicon removed the word “transient.”

<sup>4</sup> ETH.MESH.00211259 (“the foreign body reaction is not transitory – it doesn't ever go away, but decreases over time to a minimal level.”).

<sup>5</sup> See, e.g., Vailhe Dep. (6/21/13) 383:8-19; Klinge, U., et al., *Shrinking of Polypropylene Mesh In Vivo: An Experimental Study in Dogs*, Eur J Surg 1998, 164: 965-969; Klinge, U., *Foreign Body reaction to Meshes Used for the Repair of Abdominal Wall Hernias*, Eur J Surg 1998, 164:951–960. Note that I refer to documents and scientific articles concerning hernia meshes because it is my understanding from Ethicon's internal documents that the mesh used in the TVT devices is the same as that used in Ethicon's early hernia meshes. See Smith Dep. (2/3/2014) 723:9-724:6 (“The first generation (old, old) mesh is utilized currently in the TVT product...”); ETH.MESH.10633520 at 3522.

<sup>6</sup> Burkley Dep. (5/22/13) 184:17-24 (admitting that severe inflammation can lead to degradation of the mesh, shrinkage and contraction).

after implantation.<sup>7</sup> Degradation of Ethicon's mesh has been well described in a number of scientific articles, internal and external Ethicon studies.<sup>8</sup> Degradation can also occur from the natural ecosystem in the vagina, including hydrogen peroxide and other compounds that can degrade polypropylene.<sup>9</sup> However, the TVTO IFU states, incorrectly, that the mesh is inert and not subject to degradation.<sup>10</sup> In my opinion, this statement is false and misleading to physicians and patients.

This, unfortunately, establishes a vicious cycle of inflammatory foreign body response leading to mesh degradation leading to enhanced inflammatory foreign body response, etc. In the end, these women present with chronic pain, difficulty urinating, painful sexual encounters, tissue degradation, erosions, recurrent and difficult to eradicate infections and loss of sexual motivation or desire. Many of these problems are chronic and do not subside with removal of the mesh and often lead to the need for additional surgeries all of which increases the risk of and can lead to permanent damage to the urinary sphincter, worsening or de novo incontinence, scarring and chronic, debilitating pain.

It is my opinion that the Prolene mesh used in Ethicon's TVT products, including the TVTO, is not suitable for implantation in the vagina. Despite developing newer, safer polypropylene meshes that had larger pores and were lighter weight, Ethicon continued to use the old construction Prolene mesh which is associated with a number of problems as discussed in

---

<sup>7</sup> Burkley Dep. (5/23/13) 315:8-13 (discussing tests of mesh in dogs).

<sup>8</sup> ETH.MESH.02589032 and ETH.MESH.07192929 (May 18, 2011) (PA Consulting Report: Investigating Mesh Erosion in Pelvic Floor Repair and PowerPoint presentation stating: "[p]olypropylene can suffer from degradation following implant... a process which initiates after a few days post implantation in animal studies."; Clave, A., *Polypropylene as a Reinforcement in Pelvic Surgery is Not Inert: Comparative Analysis of 100 Explants*, I Urogynecol J 2010 21:261-270; Costello C., et al., "Characterization of Heavyweight and Lightweight Polypropylene Prosthetic Mesh Explants from a Single Patient," Surgical Innovation, 2007, 143:168- 176).

<sup>9</sup> Sunoco MSDS; see also EB-405, *The Durability of Polypropylene Geotextiles for Waste Containment Application* (2011) ("[P]olypropylene is vulnerable to the following substances: highly oxidized substances such as (peroxide), certain chlorinated hydrocarbons (halogenated hydrocarbons), and certain aromatic hydrocarbons.").

<sup>10</sup> ETH.MESH.02340406 (TVT IFU (May 2015) claiming mesh "is not absorbed, nor is it subject to degradation or weakening by the action of enzymes.").

this report. Ethicon's TVTO mesh causes chronic foreign body responses and severe inflammation; a response Ethicon falsely described as "transient" in its TVTO IFU. Ethicon's TVTO mesh degrades in the vagina which in turn leads to enhanced foreign body response and more inflammation. The increased inflammation further increases the degradation of the mesh – and the vicious cycle continues.

As a treating physician it would be important for me to know that Ethicon's mesh could degrade in the vagina and that it would produce a severe inflammatory response. This information would inform my treating decision when recommending treatment to a woman suffering from SUI. In addition, it would inform my treatment of a woman experiencing the adverse effects described above. Physicians expect and rely upon a medical device manufacturer to truthfully report the known qualities and risks of its products, and Ethicon failed to do so. Ethicon's TVTO products are not suitable for their intended use to treat SUI.

#### **B. The TVTO Causes Fibrotic Bridging & Scar Plate Formation**

The pore size of the mesh used in Ethicon's TVTO device is best described as "heavy weight, small-pore."<sup>11</sup> Small pore mesh has been demonstrated to increase fibrotic bridging which in turn leads to increased scar plate formation, contracture, shrinkage and effectively a stiffer mesh.<sup>12</sup> The fibrotic bridging and scar plate prevents tissue in-growth and causes complications, including, among other things, pain with the rigid mesh, shrinkage or contraction of the mesh, erosions due to mechanical irritation in the tissue of a rigid, nerve entrapment, chronic pain, recurrence, de novo or worsening incontinence, dyspareunia, wound infection, defecatory dysfunction, vaginal scarring, wound healing problems, abscess formation, and the need for additional surgeries.

---

<sup>11</sup> ETH.MESH.05918776; Batke Depo. (Aug. 1, 2013) at 87:12 - 88:10, 113:3 - 114:3, 257:23 - 259:13; Holste Depo. (July 29, 2013) at 51:3 - 53:6, 55:22-57:4; Vailhe Depo (June 20, 2013) at 182:2 185:5.

<sup>12</sup> See, e.g., B. Todd Heniford 2007 "The benefits of lightweight meshes in Ventral Hernia Repair in Ventral Hernia Repair" Video produced by Ethicon.

Ethicon's documents admit that curling/roping, frayed edges and inadequate pore size of mesh can lead to the harms of erosion, recurrence, and pain.<sup>13</sup> Former Medical Director, David Robinson, testified that the pore size of mesh "[c]ould reduce, the tissue might not encapsulate . . . the tissue might not grow through the mesh. It can become encapsulated and then it could cause . . . a rejection of the mesh" which could lead to erosion.<sup>14</sup> In fact, Ethicon developed lighter weight, large pore meshes in order to minimize the complications seen with heavyweight meshes like the Prolene used in TVTO.<sup>15</sup>

I personally have explanted meshes that have exhibited the characteristics described in these videos, e-mails and studies. I have treated these women for the very complications listed above – all of which were discussed extensively in Ethicon internal e-mails and studies.

Based on my experience, review of the scientific literature and Ethicon's internal documents, it is my opinion that the Prolene mesh used in the TVTO is not suitable for its intended application as a permanent prosthetic implant for stress urinary incontinence in women. In addition, Ethicon failed to inform physicians and patients that its mesh was susceptible to fibrotic bridging which could lead to painful erosions, infections, nerve injury, dyspareunia, the need for multiple surgeries, chronic inflammation and pain among other complications.

This information is important for the treating physician who is attempting to have an informed conversation about the risks and benefits of various treatments for her SUI. In addition, it is critical information for the physician who is treating a woman who presents with these complications. Just as important, and as discussed in more detail below, Ethicon does not provide a treating physician with any guidance regarding how to treat a woman who is experiencing a mesh that has effectively become one large scar-plate. In sum, Ethicon failed to

---

<sup>13</sup> ETH.MESH.01218019.

<sup>14</sup> Robinson 9/11/13 1070:23-1072:25.

<sup>15</sup> Holste 7/29/13 51:3-53:6.

inform physicians about these risks associated with fibrotic bridging (as a result of small-pore mesh), which means a physician could not properly discuss these risks with his or her patients, and Ethicon failed to provide any guidance or training on how to treat a patient who presented with fibrotic bridging or scar-plate formation. In my experience and my review of other IFUs, physicians expect and rely upon the manufacturer to provide this information as the manufacturer designed the device and has the most experience available to it to properly report on the risks with the device.

**C. The TVTO Causes Infections and Greatly Enhances the Probability of Experiencing Serious, Resistant Infections**

As noted above, the TVTO mesh is the same mesh as that originally used in the hernia application. However, one of the major differences between the use of mesh for hernia repair versus in the vagina is the nature of the surgical field. When placing a transvaginal mesh, the device is necessarily being placed in a contaminated surgical field. This alone can lead to severe debilitating infections, chronic foreign body response, severe inflammation, pain, painful sex, erosions, and the need for additional surgeries.<sup>16</sup>

However, the probability of severe infection is further increased due to the nature of the mesh itself. That is, the weave of the mesh effectively creates perfect, protected sites for bacteria to colonize. In addition, the microscopic flaws and degradation of the mesh fibers themselves provide fertile areas for bacterial growth. As discussed below, when mesh is implanted it can fray and deform creating additional surfaces ripe for bacterial colonization. These bacteria can secrete an encasing biofilm which shields the bacteria from destruction by the body's natural defenses.<sup>17</sup>

---

<sup>16</sup> Shah, K, et al., Bacteriological Analysis of Explanted Transvaginal Meshes (Abstract 1144); ETH.MESH. 00006636.

<sup>17</sup> Osterberg, B., et al., *Effect of Suture Materials on Bacterial Survival in Infected Wounds: An Experimental Study*, Acta. Chir. Scand 1979, 145:7 431-434; Merritt, K., *Factors Influencing Bacterial Adherence to Biomaterials*, J

Hence, implantation of the mesh in a contaminated field can lead to infection and an inflammatory chronic foreign body response. These effects are enhanced by the very nature of the mesh itself – the microscopic defects in the fibers, the frayed and deformed mesh, the weave of the mesh (particularly if the mesh is small-pored as is the TVTO mesh). These design defects create fertile ground for bacterial colonization and severe infection and inflammation. The mesh also allows the bacteria to shield itself from the body’s immune response by providing protected pockets for colonization of the bacteria, thus, creating a biofilm.

In its IFU, Ethicon states that the mesh may “potentiate infection.” As a treating physician, this statement is misleading. In fact, the very use of the device itself in a contaminated field likely will *cause* infection – not merely potentiate an already existing infection. This infection likely will be further exacerbated by the design of the mesh – the inherent microscopic defects in the fibers, the weave of the mesh (small pores), the defects of fraying and mesh deformation and the creation of biofilm in the pockets created by the mesh. Mesh-induced infections can be very difficult to eradicate and lead to chronic pain and inflammation, erosion (spreading the infection to other surrounding areas),<sup>18</sup> dyspareunia, the need for removal of the mesh (in one or multiple surgeries), tissue damage and death, urinary pain and dysfunction and de novo incontinence.

Ethicon failed to properly inform physicians and patients of the risk of infection caused by the TVTO procedure particularly when combined with the design of the mesh itself. I have personally seen the terrible outcomes associated with these severe infections, including the

---

Biomat Appl 1991, 5:185-203; An, Y., *Concise Review of Mechanisms of Bacterial Adhesion to Biomaterial Surfaces*, J Biomed Mater Res (Appl Biomater) 1998, 43:338-348; The TVM Group: J. Berrocal, et al., *Conceptual advances in the surgical management of genital prolapsed*, J Gynecol Obstet Biol Reprod 2004, 33:577-587.

<sup>18</sup> ETH.MESH.02089392 (“Consensus: Erosion is a risk. Erosion, possibly an infection response. Typically seen by 3 mos, usually by 6-12 mos. Can present late, 3 years. To vagina-not a good situation. To bladder, urethra or rectum-a very bad situation.”).

chronic foreign body response that persists even when the mesh is explanted, severe inflammation, persistent, chronic pain, painful intercourse, erosions and repeated erosions, persistent infections that are difficult to fully eradicate, urinary tract infections, pain during urination and the need for multiple surgeries to achieve mesh removal.<sup>19</sup>

#### **D. Problems Attributable to Ethicon's Cutting Processes**

Ethicon cuts its TVTO mesh to size by one of two methods: by machine (mechanically-cut) or with a laser (laser-cut). The method by which mesh is cut significantly impacts how that mesh will perform during the implantation surgery and while implanted in a woman's body.<sup>20</sup> Despite these significant differences, Ethicon chose to inform physicians that there was no real difference in performance between the meshes.<sup>21</sup> This was misleading.

##### **1. Mechanically Cut Mesh**

When it is cut to size by mechanical means, the edges of the mesh are sharp and have a tendency to cut into the surrounding tissue. In a presentation by Ethicon's engineer, Gene Kammerer, significant degradation in the form of fraying of the cut edges of the of the mesh and a loss of some of the edge particles was observed in the mechanically-cut mesh (MCM) after it had been elongated by 50%, while the mesh that was laser-cut (LCM) demonstrated less particle loss, fraying and degradation.<sup>22</sup>

---


<sup>19</sup> The PROLENE IFU by comparison states as follows: "PROLENE Mesh in contaminated wounds should be used with the understanding that subsequent infection may require removal of the material." ETH.MESH.02342102.

<sup>20</sup> ETH.MESH.00687819; Depo. Robinson (July 25, 2013) 585:12-23.

<sup>21</sup> ETH.MESH .00858252.

<sup>22</sup> ETH.MESH.08334245.





In addition to fraying and resultant particle loss, the mechanically-cut TVTO mesh also tends to more drastically rope, curl and deform when it is stretched or placed under tension. In the same presentation by Kammerer, significant deformation, in the form of roping/twisting, and curling was also observed in the mechanically-cut mesh (MCM) after it had been elongated by 50%, while the mesh that was laser-cut (LCM) suffered some of the same deformation issues, but to a lesser degree.

Indeed, I personally have witnessed the same type of deformation of the mesh material, *i.e.*, curling, roping/twisting, narrowing and fraying when I have been called upon to remove mesh months or years after implantation. It is my opinion that TVTO mesh can rope, curl and deform. Deformation of mesh can create the perfect habitat for bacteria leading to severe infections, inflammation and chronic foreign body response and all of the associated adverse events, such as chronic debilitating pain, difficulty urinating, painful sex, erosions, needing mesh

removal (including multiple surgeries), and urinary incontinence. Deformation of mesh by tensioning can also make the pores of the already small-pore TVTO mesh even smaller, creating a higher probability and exacerbation of the adverse outcomes discussed above.<sup>23</sup>

## 2. Laser Cut Mesh Is Stiffer and Less Elastic

The laser cutting technique was developed presumably in an effort to address the chronic problems of particle loss, fraying, sharp edges and other issues inherent in mechanically-cut mesh. However, laser cut products have their own problems.

Ethicon's own testing revealed that laser cutting the mesh caused it to behave differently than mechanically-cut mesh, particularly with regard to its stiffness and elasticity. The studies showed that laser-cut mesh was indeed substantially stiffer and less elastic.<sup>24</sup> Because of its relative lack of elasticity, it is more difficult to achieve proper tensioning with the stiffer laser-cut mesh, and the tendency is to over-tension the product during implantation.<sup>25</sup> In addition, stiff mesh can lead to pain, damage to the urethra, and erosions.<sup>26</sup>

Despite having different mechanical characteristics, Ethicon did not tell doctors via the Instructions for Use for TVTO that the laser cut mesh acted differently than the mechanical mesh. Therefore, implanting surgeons would not have known that they needed to tension the TVTO sling differently when using a laser cut mesh versus a mechanical mesh. This failure to notify doctors about the differences combined with the fact that the TVTO is always implanted via a blind procedure would lead to an increased likelihood that the laser cut TVTO would not be implanted tension free, which leads to increased complications as discussed in this report.

<sup>23</sup> Depo. Smith (February 3, 2014) 816:5-15.

<sup>24</sup> In an internal memo documenting the results of TM403-477 (Ethicon's standard method for measuring tensile properties of TVT mesh), Ethicon reported that the laser cut mesh was approximately three times stiffer than mechanically-cut mesh after 20% elongation. ETH.MESH.00302181.

<sup>25</sup> ETH.MESH.04048515 at 516 (Professor Carl G. Nilsson stated he "will not use Laser-cut mesh!!" as it "does not have the same stretch profile of Mechanical-cut mesh").

<sup>26</sup> The Failure Modes and Effect Analysis for laser-cut mesh specifically notes that mesh that is "too stiff" can cause harms such as pain, damage to the urethra, urethral impingement, and damage to the bladder."

**E. Cytotoxicity**

Cytotoxicity means toxicity to the cells causing cell injury or death.<sup>27</sup> In a May 26, 2000, Ethicon Memo titled “Review of biocompatibility on the tension-free vaginal tape (TVT) system for compliance to FDA,”<sup>28</sup> the review contains a “Cytotoxicity Risk Assessment for the TVT (Ulmsten) Device” from August 8, 1997.<sup>29</sup> The Cytotoxicity Assessment states “there is some evidence to suggest that the PP [polypropylene] mesh from the sterile Ulmsten device may have cytotoxic potential.”<sup>30</sup> In addition, ISO Elution testing “resulted in marked cytotoxicity in tests conducted at Ethicon (Scotland).” Because the TVT and TVTO slings are both made from the same Prolene mesh, the TVTO is also cytotoxic.

It is my opinion based on my professional experience and the research that I have conducted that the adverse outcomes I have seen or researched can be caused and, in some of my patients, were likely caused by the cytotoxicity of the mesh.

**F. Polypropylene Is Incompatible with the Vaginal Ecosystem**

I have been provided with the Material Safety Data Sheets (“MSDS”) for the polypropylene used in the Prolene mesh for Ethicon’s TVTO devices. I have also reviewed a number of other MSDS’s for other similar products. For a number of reasons discussed below, the MSDSs of Phillips Sumika and Chevron Corporation specifically warned that the polypropylene material should never be permanently implanted in the human body.

According to the MSDS for the polypropylene resin used in the Prolene for the TVTO devices, the polypropylene is incompatible with “strong oxidizers” such as “peroxides.” Peroxides, particularly hydrogen peroxide, are a critical part of maintaining the vaginal ecosystem. As early as 1987, Ethicon scientists concluded that Prolene was susceptible to

---

<sup>27</sup> Robinson Depo. (September 11, 2013) 1091:11-21.

<sup>28</sup> ETH.MESH.06852118 at 2118-2129 (5/26/2000 Biocompatibility Review).

<sup>29</sup> ETH.MESH.06852120 (8/8/1997 Cytotoxicity Risk Assessment).

<sup>30</sup> *Id.* and Robinson Depo. (September 11, 2013) 1098:23-1099:9.

oxidation which led to degradation of the mesh in vivo. Hence, according to the Prolene MSDS, the polypropylene is incompatible for use in the vagina.

I have also reviewed MSDSs for the polypropylene resin used in other pelvic floor meshes. These MSDSs similarly warn about the incompatibility of the polypropylene with peroxides and discuss how such interactions could lead to degradation of the polypropylene product, such as mesh when exposed to peroxides. Some expressly state that the product should not be used for permanent implantation in the human body.

In my opinion, this information would be important to a treating physician when making the decision whether or not to use a TVTO device and also when treating a woman who has suffered from complications from a TVTO device. For example, when explanting mesh, it often is very difficult to remove all of the mesh as often it has frayed, degraded or some portions have been fully in-grown. However, understanding that the mesh is incompatible with the very space where it is implanted and that leaving any mesh in place might exacerbate or even continue to cause the initial problem might militate in favor of taking a more aggressive approach to removal.

In my review of the documents provided to me for this report, it is clear that Ethicon never informed physicians of this incompatibility and the potential adverse outcomes that could arise. It is also clear from the depositions I reviewed that Ethicon failed to test whether the polypropylene used in the TVTO devices was incompatible with the peroxides and other vaginal chemistry and the potential adverse effects that could result.

In addition to these undisclosed risks of incompatibility, the MSDSs I have reviewed expressly discuss a risk of cancer associated with the polypropylene. According to Ethicon employee depositions I have reviewed, Ethicon reproduced this cancer risk in animal studies.

### **G. TVT-O Causes Severe and Debilitating Groin, Leg and Lower Body Pain**

The TVT-O IFU stated, “Transient leg pain lasting 24-48 hours may occur and can usually be managed with mild analgesics.” This statement is very misleading. Numerous studies have demonstrated that TVT-O causes severe debilitating leg, groin and lower body pain at a very high frequency.<sup>31</sup> Moreover, the studies demonstrate that often the pain is not transient, but long-lasting and in some cases permanent.<sup>32</sup> The cause of the pain appears to be multi-factorial and related to the surgical path necessitated by the TVT-O procedure, the location of the mesh once it is implanted, nerve damage or entrapment, chronic foreign body response, and/or scar formation.<sup>33</sup> In fact, Ethicon and DeLeval, the inventor of TVT-O, designed the TVT Abbrevio in an effort to lessen the pain syndromes caused by TVT-O.

### **V. THE TVTO INSTRUCTIONS FOR USE ("IFU") IS INADEQUATE**

In my opinion based on my professional experience, the instructions for use (“IFU”) plays an important role in the informed consent process because it provides information for doctors to learn about the risks and benefits associated with a medical implant, such as the TVTO. It also provides insights a manufacturer has gained from the worldwide experience with the use of the product, including how to address unique patient attributes and deal with typical complications that may be seen with the device. In my experience, physicians rely upon the device manufacturers to accurately and truthfully describe the known risks and complications associated with the device to allow them to make informed treatment decisions and to allow

---

<sup>31</sup> See, e.g., Teo, et al., Randomized Trial of TVT and TVT-O for the Treatment of Urodynamic Stress Incontinence in Women, 2011 J Urol. (investigators stopped trial due to excess pain reports in the TVT-O arm (26.4% pain reported at 6 months)); Lim J., et al., Clinical and quality of life outcomes in women treated by TVT-O procedure. BJOG (2006) 113:1315–1320 (24.4% groin discomfort within the first 6 months; 3.6% had persistent groin pain unresolved by 1 year); Collinet P., et al. The safety of the insideout transobturator approach for transvaginal tape (TVT-O) treatment in stress urinary incontinence: french registry data on 984 women. Int J Pelvic Floor Dysfunct (2008) 19:712–715., requiring eventual surgery.

<sup>32</sup> *Id.*

<sup>33</sup> ETH.MESH.04050265.

them to properly consent their patients. In addition, physicians rely upon the IFU to help treat patients who have presented with complications from the device. During the course of my career, I have regularly reviewed IFUs for medical products, including IFUs for mesh slings to treat stress urinary incontinence such as the TVTO.

When a patient undergoes a surgical procedure, the surgeon first discusses the risks and benefits with the patient so that the patient may make an informed decision to undergo the procedure. This process is called an informed consent. For a general pelvic surgery, the basic hospital informed consent would cover risks common surgical risks, such as damage to the pelvic structures (GI/GU/vessels/nerves), hemorrhage, infection, urinary incontinence, retention, dyspareunia and recurrent prolapse. If a device is to be used, the surgeon would then rely upon information provided by the manufacturer in the IFU to learn about the warnings specific to that product.

From my extensive experience practicing in medicine and gynecology, talking with other doctors in my field, and speaking at and attending medical conferences, many of the doctors who implanted these products were not aware of many of the risks or their frequency or severity, as discussed above. The TVTO IFU is inadequate because it omits and understates the serious complications discussed above. It also fails to properly inform a physician how to treat a patient who has presented with the above complications. My opinion is further supported by my review of internal Ethicon documents, including statements by Ethicon employees such as Meng Chen and Dr. Aaron Kirkemo.<sup>34</sup>

#### **A. The IFU Omits Many of the Significant Complications**

Many of the risks identified in the previous sections are not included in the TVTO IFU, such as:

---

<sup>34</sup> ETH.MESH.04092868; ETH.MESH.0494863.

1. TVTO was capable of causing chronic, permanent debilitating pain;
2. TVTO could cause lifelong risk of erosions;
3. Erosions can be severe, untreatable and incurable;
4. Complications from TVTO could cause the patient to need lifelong surgeries to treat mesh erosions and degradation;
5. Serious and chronic inflammation could occur with the use of TVTO, and that this complication was not slight or transient;
6. TVTO was capable of causing permanent dyspareunia;
7. Eroded mesh posed an injury risk to the woman's partner;
8. TVTO mesh was capable of causing Urinary problems, including: Urethral injury, Voiding dysfunction, De novo detrusor instability or urgency, Urinary retention/obstruction, Urinary retention (temporary or permanent), Urinary tract infection, Dysuria, Hematuria, and Worsening or recurrence of incontinence;
9. Mesh was capable of causing permanent nerve injury;
10. Mesh pores could collapse under strain and cause unwanted fibrotic bridging, which was capable of causing painful scarring, permanent tissue damage, nerve pain and entrapment, dyspareunia, urinary and defecatory problems, and chronic debilitating pain;
11. The polypropylene TVTO mesh was capable of degrading;
12. TVTO mesh was cytotoxic;
13. TVTO could cause toxic shock syndrome;
14. TVTO can cause serious, difficult to treat infections, de novo urinary tract infections, inflammation and that the changes to the vaginal ecosystem from these imbalances can potentiate mesh degradation and chronic foreign body reaction;
15. Particle loss was a known problem and that it contributed to further unwanted foreign body reaction;
16. Ethicon possessed evidence that the risk of vaginal scarring was greater than disclosed in its IFU;
17. Use of TVTO could cause narrowing of the vaginal wall;
18. Ethicon did not have any procedure or Professional Education program to teach doctors how to properly remove TVTO mesh slings when known complications occurred;
19. Manufacturers of polypropylene resin stated that it should not be used for medical purposes or permanent implantation in the human body;
20. That the polypropylene resin used for the mesh in the TVTO induced sarcomas at the site of implantation.
21. The TVTO could cause permanent and debilitating leg, groin or thigh pain.

**B. The IFU Understates the Severity and Frequency of the Complications**

The IFU also uses language that understates the known risks listed in the IFU, such as “transitory” or “rare” or “typical”. This language understates the frequency of the adverse outcomes, the severity of those complications, and the duration of those complications. In my experience and my review of the medical literature, and as admitted in numerous Ethicon



documents described above, the complications from TVTO are far from transitory, and that the complications and injuries suffered by patients are often long term and permanent.

**C. The IFU Does Not Adequately Discuss Methods to Treat Complications, Including the Safest Methods to Remove a TVTO**

The TVTO IFU did not inform physicians about techniques to treat complications from TVTO mesh. From my experience, it is important for doctors to know what to look for and where to look when performing a revision surgery. Ethicon has more information from its worldwide database and experience from sales representatives than anyone else regarding the most effective and safest methods to treat complications. Further, Ethicon did not provide information in the IFU about the significant dissection associated with surgery to explant the TVTO mesh. Ethicon did not warn doctors that it is almost impossible to fully remove all of the mesh once it has been integrated into the tissue or degraded. Finally, Ethicon did not warn in the IFU that revision surgery further increases scar tissue, leading to increased complications such as pain, disfigurement, and dyspareunia.

**D. The IFU Does Not Address Special Patient Populations**

The TVTO IFUs offer no guidance for physicians on use of the TVTO mesh products in special patient populations. For example, studies have shown the TVTO device is not as effective and can cause more complications in women who are obese, diabetic, smokers, older women, women with pre-existing or who might develop pelvic pain or myalgias and younger athletic women. Further, research has shown that there are anatomical differences between individuals and between different ethnicities generally. Even with highly skilled surgeons, it is impossible to fully predict and account for anatomical differences when implanting the TVTO mesh via the blind passage. Yet, the IFU does not inform physicians (who in turn cannot inform patients) of these known risks.



### **E. Bi-Directional Properties of the Mesh**

Ethicon states in the TVTO Instructions for Use (“IFU”) that Prolene mesh is knitted by a process “which interlinks each fiber junction and which provides for elasticity in both directions. This bi-directional elastic property allows adaptation to various stresses encountered in the body.”<sup>35</sup> Ethicon has never performed a study to determine what bi-directional forces are present in the pelvic region, or tested how its TVTO product is able to meet those requirements. Instead, and as discussed above, the implanted mesh creates a chronic foreign body reaction, which can lead to the formation of a rigid scar plate and contraction of the mesh implant. As such, the mesh is not capable of adapting to various, bi-directional stresses in the body. In my opinion, this information would be important to a treating physician when making the decision whether or not to use a TVTO device and also when treating a woman who has suffered from complications from a TVTO device.

### **VI. THE SIGNIFICANT RISKS OUTWEIGH THE POTENTIAL BENEFIT**

In my experience and professional opinion, the benefits of Ethicon’s TVTO product are outweighed by the significant risks that the products pose to the health of the patients in the long term. As discussed above, there are numerous serious complications associated with implanting the TVTO sling for treatment of stress urinary incontinence, including death, permanent disfigurement, and chronic pain. Based on my surgical experience in removing mesh from women and years of experience reviewing mesh related complications, I have reached the opinion that polypropylene mesh has harmful effects for women when permanently implanted in the pelvic region. I have seen how the polypropylene mesh reacts with major organs in the pelvic region. It is my opinion that polypropylene mesh permanently implanted in the pelvic

---

<sup>35</sup> ETH.MESH.00353639, ETH.MESH.00015699 –00015706; ETH.MESH.00013506; ETH.MESH.00922443-00922445; ETH-00938; Walji Deposition p471-472; Robinson Deposition 3-14, p683-684; Kirkemo Deposition 4-18, p246-247, Ciarrocca Deposition 3-29, p264


region has a propensity to shrink, contract, erode, degrade, sustain particle loss and fraying, and cause a chronic foreign body reaction, fibrotic bridging and scar plating. All of these propensities result in harmful complications for the patient.

Studies have shown that the TVTO device is no more efficacious at treating stress urinary incontinence than other procedures that do not involve mesh. Multiple alternative therapies and procedures (surgical and non-surgical) are available for the treatment of SUI, including but not limited to Burch procedures (which use the patient's own ligaments to support the prolapse), removable pessary devices, urinary seals, urethral inserts, bulking agent injections, medications and pelvic muscle exercises (or "Kegel"). When weighing the risks and benefits with the TVTO product, it is important to note that stress urinary incontinence is not a life threatening condition and does not cause chronic and debilitating pain.

In further support of my opinion that the risks outweigh the benefits, it is commonly accepted that the true rate of risks is likely higher than what published in the scientific literature. Research has shown that adverse events are commonly underreported, including complications with mesh slings for treatment of stress urinary incontinence. Additionally, many of the safety and efficacy studies for these products were performed by persons with a financial interest in selling these products. Accordingly, the true rate of complications is likely understated and the actual rate of success is likely overstated, both of which further support my opinion that the risks are outweighed by the benefits.

I reserve the right to update my opinions if additional materials become available.

February 1, 2016



Michael Thomas Margolis, M.D.

## **VII. EXHIBITS**

My current curriculum vitae is attached hereto as Exhibit "A"

All exhibits that will be used to support my findings and opinions or documents that I have reviewed are referred to herein or in Exhibit "B"

## **VIII. RECENT TESTIMONY**

In the previous four years, I have testified, by deposition or in trial, in the following cases:

- Coleen Perry v. Luu, et al.
- Martha Salazar v. Lopez, et al.
- Linda Batiste v. Johnson & Johnson and Ethicon, Inc.
- Elizabeth Ayerdis v. Hawaii Permanente Medical Group
- Carpenter v. Hardesty, et al; Superior Court of the State of California, San Bernadino County
- Sabrina Gaines v. Kendal Freeman, et al.
- Gross v. Gynecare, New Jersey Superior Court, Atlantic County Division
- Harrison v. Eberts
- Lorie Koop v. George Stankevych
- Tracey Patane v. Ardent Health Services, et al.
- Pamela Russell v. Neil Harrison, et al.
- Rizzo, et al. v. C.R. Bard, Inc., United Stated District Court, Northern District of Georgia
- Scorpio v. Noone
- Waddoups v. Barry Noorda, et al., United States District Court, District of Utah
- Nancy & Gary Fleming v. Boston Scientific
- Mildred Blankenship v. Boston Scientific
- Penny Barton v. Boston Scientific
- Catherine Pickle v. Boston Scientific

- Esbarda Chapa v. Boston Scientific
- Gwendolyn Gravitt v. Boston Scientific
- Hilda Escalante v. Bard
- Glory Lewis v. Bard
- Rebecca George –2:12-cv-03585
- Gilbert v C.R. Bard, Inc.: Case No.: 2:14-cv-00404
- Sherrer v Boston Scientific Corporation, et al: Case No.: 1216-cv27879
- Carlino v Ethicon (Ongoing in Philadelphia right now)

# **EXHIBIT B**

**Dr. Michael Thomas Margolis**

List of Prior Testimony

In the previous four years, I have testified, by deposition or in trial, in the following cases:

- Coleen Perry v. Luu, et al; Case No. 1500-CV-279123; Superior Court of the State of California, Kern County
- Martha Salazar v. Lopez, et al; Cause No. DC-12-14349; In the 160<sup>th</sup> Judicial District Court of Dallas County, TX
- Linda Batiste v. Johnson & Johnson and Ethicon, Inc; Cause No. DC-12-14350; IN the 134<sup>th</sup> Judicial District Court of Dallas County, TX
- Elizabeth Ayerdis v. Hawaii Permanente Medical Group
- Carpenter v. Hardesty, et al; Case No. CIVDS-1007094; Superior Court of the State of California, San Bernadino County
- Sabrina Gaines v. Kendal Freeman, et al.
- Gross v. Gynecare, Case No. AT1-L-6966-10; New Jersey Superior Court, Atlantic County Division
- Harrison v. Eberts
- Lorie Koop v. George Stankevych
- Tracey Patane v. Ardent Health Services, et al.
- Pamela Russell v. Neil Harrison, et al.
- Rizzo, et al. v. C.R. Bard, Inc., United Stated District Court, Northern District of Georgia
- Scorpio v. Noone
- Waddoups v. Barry Noorda, et al., United States District Court, District of Utah
- Nancy & Gary Fleming v. Boston Scientific
- Mildred Blankenship v. Boston Scientific
- Penny Barton v. Boston Scientific
- Catherine Pickle v. Boston Scientific
- Esbarda Chapa v. Boston Scientific
- Gwendolyn Gravitt v. Boston Scientific
- Hilda Escalante v. Bard
- Glory Lewis v. Bard
- Rebecca George –2:12-cv-03585
- Gilbert v C.R. Bard, Inc.: Case No.: 2:14-cv-00404
- Sherrer v Boston Scientific Corporation, et al: Case No.: 1216-cv27879
- Carol Cavness v. Kowalczyk, et al; Cause No. DC-14-04220; In the 95<sup>th</sup> District Court of Dallas County, Texas
- Sharon and Charles Carolina v. Ethicon; Case No. 1306-03470; Philadelphia Court of Common Pleas
- Carolyn Lewis, et al. v. Ethicon; Cause No. 2:12-cv-04301; United District Court for the Southern District of West Virginia

- Jennifer Ramirez v. Reyes, et al; Cause No. 2012-CI-18690; In the 438<sup>th</sup> Judicial District Court of Bexar County, Texas
- Josephine Rabiola v. Antoni, et al; Cause No. D-1-GN-13-002039; In the 53<sup>rd</sup> Judicial Court of Travis County, Texas

**Expert Report of Dr. Michael Thomas Margolis**  
**List of Materials Reviewed**

**Scientific Literature**

Abbott S, et al., Am J Obstet Gynecol. 2014 Feb;210(2):163 Evaluation and management of complications from synthetic mesh after pelvic reconstructive surgery: a multicenter study.
Adloff, M., and JP Arnaud, [Experimental study of resistance to and biologic tolerance of prosthetic materials used in the repair of losses of substance of the abdominal wall, 102 Chirurgie 390-6 (1976) (Article in French).
Adverse Events over Two Years after Retropubic or Transobturator Midurethral Sling Surgery: Findings from the Trial of Midurethral Sling (TOMUS) Study, Am J Obstet Gynecol. 2011 November; 205(5): 498.e1–498.e6.
Altman, D., et al., Anterior colporrhaphy versus transvaginal mesh for pelvic-organ prolapse, 364 N. Engl. J. Med. 1826-36 (2011).
Amid PK, 2D sheets shrink 20% in surface area, Hernia 1997
American College of Obstetricians and Gynecologists, Pelvic Organ Prolapse Practice Bulletin (Feb. 2007).
American College of Obstetricians and Gynecologists, February Committee on Gynecologic Practice (2012).
Amid P.K. and IL Lichtenstein, [Current assessment of Lichtenstein tension-free hernia repair], 68 Journal Hernia 959-64 (1997) (Article in German).
An, Y.N., and RJ Friedman, Concise review of mechanisms of bacterial adhesion to biomaterial surfaces, 43 J. Biomed. Mater. Res. 338 (1998).
Anderson, A., et al., Foreign Body Reactions to Biomaterials, 20 Semin. Immunol. 86-100 (2008).
Andonian, S, et al, Randomized clinical trial comparing SPARC and TVT: one year results. Eur Urol. 2005 Apr;47(4):537-41
Balakrishnan, S., et al., Sling Distress: A Subanalysis of the IVS Tapes from the SUSPEND Trial, 47 Aust. N. Z. J. Obstet. Gynaecol. 496-8 (2007).
Beck, RP et al., The Fascia Lata Sling Procedure for Treating Recurrent Genuine Stress Incontinence of Urine, Obstet Gynecol 72:699, 1988s
Berard F, Gandon J, ACS-classification of operative wounds based on degree of microbial contamination, Ann Surg 1964
Berard, F., and Gandon, J. Postoperative Wound Infections: The Influence of Ultraviolet Irradiation of the Operating Room and of Various other Factors, 160 Ann. Surg. 1-192 (1964) (Adapted).
Blaivas, J., <i>Pubovaginal Slings</i> , in Female Urology 235 (ED Kursh & EJ McGuire eds., 1994).
BMJ. May 14, 2005; 330(7500): 1121. The gold standard: not a golden standard Jurgen AHR Claassen
Bournager, et al., Bacteriological analysis of meshes removed for complications after SUI & prolapse surgery, 19 Int Urogynecol J. 827-31 (2008).



Boyles & Edwards, Incidence of Complications After TOT Sling Procedures, The Society of Gynecologic Surgeons 31st Annual Scientific Meeting (2005).
Brink, D.M. Bowel injury following insertion of TVT, 90 S. Afr. Med. J. 450, 452 (2000).
Bustan, M., et al., [Burch laparoscopic procedure for repairing proven stress incontinence- report of 32 cases] 139 Harefuah 350-2, 407 (2000) (Article in Hebrew).
Chmielewski, L., et al., Reanalysis of a randomized trial of 3 techniques of anterior colporrhaphy using clinically relevant definitions of success, 205 Am. J. Obstet. Gynecol. 1-8 (2011).
Clave, et al., Polypropylene as a reinforcement in pelvic surgery is not inert: Comparative analysis of 100 explants Int Urogynecol J 2010
Clave, A. et al., Polypropylene as a reinforcement in pelvic surgery is not inert: Comparative analysis of 100 explants, 21 Int. Urogynecol J. 261-70 (2010).
Coda, A. et al., Structural alterations of prosthetic meshes in humans, 7 Hernia 29-34 (2003).
Committee on Gynecologic Practice, Vaginal placement of synthetic mesh for pelvic organ prolapse, 18 Female Pelvic Med Reconstr Surg. (2012).
Committee on Practice Bulletins-Gynecology, Practice Bulletin No. 79: Pelvic Organ Prolapse, 109 Obstet. Gynec. 461-73 (2007).
Cooper, M., et al., A Review of Results in a Series of 113 Laparoscopic Colposuspensions, 36 Aust. N.Z. J. Obstet. Gynecol. 44-48 (1996).
Council for International Organizations of Medical Sciences Working Group V, Current Challenges in Pharmacovigilance: Pragmatic Approaches (2001)
Cruse, P.J., and R Foord, The epidemiology of wound infection - A 10-year prospective study of 52,939 wounds, 60 Surg. Clin. North Am. 27-39 (1980).
Cruse and Foord, Rate of infection (1980/1992)
Culligan, P., et al., Bacterial Colony Counts During Vaginal Surgery, 11 Infect. Dis. Obstet. Gynecol. 161 (2003).
Cvach, K., et al., Abdominal Sacral Hysteropexy: A Pilot Study Comparing Sacral Hysteropexy to Sacral Colpopexy with Hysterectomy, 18 Female Pelvic Med. Reconstr. Surg. 286-90 (2012).
Cvach K, Geoffrion R, Cundiff GW. Female Pelvic Med Reconstr Surg, 2012 Sep-Oct; 18(5):286-90
Elliott, CS, MD et al., Might Pelvic Surgeons Be Unaware of Their Surgical Failure's? Patient Reporting & Perceptions After Failed Incontinence or POP Surgery, FPM&RS Sept/Oct 2015 Issue 5 p 298-300
Elmer, C., et al., Histological inflammatory response to transvaginal polypropylene mesh for pelvic reconstructive surgery, 181 J Urol. 1189 (2009).
E-mail from James McDivitt to Thomas Barbolt dated 8/16/2001 re: Autoclaving Prolene
Falagas, M.E et al., Mesh-related infections after pelvic organ prolapse repair surgery 134 Eur. J. Obstet. Gynecol. Reprod. Biol. 147-56 (2007).
Fan X, et al., Comparison of polypropylene mesh and porcine-derived, cross-linked urinary bladder matrix materials implanted in the rabbit vagina and abdomen, Int Urogynecol J. Nov 29, 2013

Fatthy, H., et al., Modified Burch colposuspension: laparoscopy versus laparotomy, 8 J. Am. Assoc. Gynecol. Laparosc. 99-106 (2001).
Feola, A et al., Deterioration in Biomechanical Properties of the Vagina Following Implantation of a High-Stiffness Prolapse Mesh. BJOG. 2013 Jan;120(2):224-32
Female Pelvic Medicine & Reconstructive Surgery. 18(1):5-9, January/February 2012
Frostling, H. et al., Analytical, occupational and toxicologic aspects of the degradation products of polypropylene plastics, 10 Scand. J. Work Environ. Health 163 (1984).
Gandi, S., et al., A prospective randomized trial using solvent dehydrated fascia lata for the prevention of recurrent anterior vaginal wall prolapse, 192 Am. J. Obstet. Gynecol. 1649-54 (2005).
Garcia-Urena, M.A., et al., Differences in polypropylene shrinkage depending on mesh position in an experimental study 193 Am. J. Surg. 538-42 (2007).
Glavind, K., and E. H. Larsen, Results and Complications of TVT for Surgical Treatment of Female SUI, 12 Int. Urogyn. J. (2001).
Gutman, et al., Three-Year Outcomes of Vaginal Mesh for Prolapse-A RCT, Obstet Gynecol Vol. 122, No. 4, October 2013
Hervas, M., et al., Analysis of the efficacy and complications of the surgical treatment of SUI: Retropubic and Transobturator TVT. Our experience, 60 Arch. Esp. Urol. 1091-8 (2007).
Hviid, U., et al., Porcine skin collagen implants for anterior vaginal wall prolapse: a randomised prospective controlled study, 21 Int. Urogynecol. J. 529-34 (2010).
Iglesia 2010
"Industry Sponsorship and Research Outcome", The Cochrane Library, 12Dec2012
Jacome, E.G., et al., Laparoscopic Burch urethropexy in a private clinical practice, 6 J. Am. Assoc. Gynecol. Laparosc. 39-44 (1999).
Jacquelin, B., and Cosson, M. Complications of vaginal Mesh: Our Experience, 20 Int. Urogyn. J. 893-96 (2009).
Jenkins, T.R. and CY Liu, Laparoscopic Burch colposuspension, 19 Curr. Opin. Obstet. Gynecol. 314-8 (2007).
Jongebloed, W.L. and JF Worst, Degradation of polypropylene in the human eye: a SEM study. 64 Doc Ophthalmol 143-52 (1986).
Karram, M., et al., Complications and Untoward Effects of the TVT Procedure, 101 Obstet. Gynecol. 929-32 (2003).
Karram 2012
Kaelin-Gambirasio, I., et al., Complications Associated with Transobturator Sling Procedures: Analysis of 233 Consecutive Cases with a 27 months Follow-up. 9 BMC Womens Health 28 (2009).
Klinge, U., et al., Modified mesh for hernia repair that is adapted to the physiology of the abdominal wall 164 Eur. J. Surg. 965 (1998).

Klinge U, Klosterhalfen B, et al. Shrinking of polypropylene mesh in vivo: an experimental study in dogs. Eur J Surg. 1998 Dec; 164(12):965-9
Klinge, U., et al., Do multifilament alloplastic meshes increase the infection rate? Analysis of the polymeric surface, the bacteria adherence. and the in vivo consequences in a rat model, 63 J. Biomed. Mater. Res. 765 (2002).
Korda, A., et al., Experience with Silastic Slings for Female Urinary Incontinence, 29 Aust. N.Z. J. Obstet. Gynecol. 150-154 (1989).
Kuhlmann-Capek, Maggie MD, et al., Enmeshed in Controversy: Use of Vaginal Mesh in the Current Medicolegal Environment. FPM&RS Sep/Oct 2015 issue 5 p 241-43
Kursh and McGuire, <i>Female Urology</i> 1994
Latini, J.M., et al., Efficacy and morbidity of autologous fascia lata sling cystourethropexy, 171 J. Urol. 1180-4 (2004).
Lee, C.L., et al., Extraperitoneal approach to laparoscopic Burch Colposuspension, 8 J. Am. Assoc. Gynecol. Laparosc. 374-7 (2001).
Liang, R et al., Vaginal degeneration following implantation of synthetic mesh with increased stiffness, BJOG 2013 Jan;120(2):233-43
Lundh, A., et al., Industry Sponsorship and Research Outcome, 12 Cochrane Database Syst. Rev. MR0000033 (2012).
Mahmoud, W. J. 7 Biomat. Sci. Polymer Ed 751 (1996).
Manodoro, S et al., Graft-related complications and biaxial tensiometry following experimental vaginal implantation of flat mesh of variable dimensions, BJOG. 2013 Jan;120(2):244-50
Mazouni, C et al., Urinary complications and sexual function after the TVT procedure, Acta Obstetrica et Gynecologica Scandinavica, Vol. 83, Issue 10, pages 955-961, Oct 2004
Merritt, K., and CC Chang, Factors influencing bacterial adherence to biomaterials, 5 J. Biomater. Appl. 185 (1991).
Mohide, E.A., The Prevalence and Scope of Urinary Incontinence, 2 Clin. Geriatr. Med. 639-55 (1986).
Neuman, M., TVT and TVT-Obturator: comparison of two operative procedures, 131 Eur.J.Obstet. Gynecol. Reprod. Biol. 89-92 (2007).
Nichols, David H., Reoperative Gynecologic Surgery (1991 ).
Nieminen 2010
Niemczyk et al., U.S. experience with TVT procedure for SUI: assessment of safety and tolerability Tech Urol 2001 Dec;7(4):261-5.
Maher, et al April 30, 2013, Cochrane Summaries: Surgical Management of POP
Maher C., et al., Surgical management of pelvic organ prolapse in women, 4 Cochrane Database Syst. Rev. CD004014 (2013).
Marshall, V.F., et al., The correction of stress incontinence by simple vesicourethral suspension, 88 Surg. Gynecol. Obstet. 509-18 (1949).

MAUDE Database Report Counts-SUI & POP Mesh (1999-Sep, 2014)
Mazouni, C et al., Urinary complications and sexual function after the TVT procedure, Acta Obstetricia et Gynecologica Scandinavica, Vol. 83, Issue 10, pages 955-961, Oct 2004
Medline Search 4/8/2014 for “gold standard”
Meschia, M., et al., Porcine skin collagen implants to prevent anterior vaginal wall prolapse recurrence: a multicenter. randomized study, 177 J. Urol. 192-5 (2007).
Moore, R.D., et al., Laparoscopic Burch colposuspension for recurrent stress urinary incontinence, J. Am. Assoc. Gynecol. Laparosc. 389-92 (2001).
Niemczyk, P., et al., U.S. experience with TVT procedure for SUI: assessment of safety and tolerability, 7 Tech. Urol. 261-5 (2001).
Oktay Demirkisen, et al., Comparison of Dyspareunia Between TVT v Burch Patients, International Braz J Urol, Vol 32 (2):214-219, March-April, 2008
Olsen, AL et al., Epidemiology of surgically managed pelvic organ prolapse and urinary incontinence. 89 Obstet Gynecol. 501-6 (1997).
Osterberg, B. and B. Blomstedt, Effect of suture materials on bacterial survival in infected wounds. An experimental study, 145 Acta. Chir. Scand. 431(1979).
Ostergard, D, Polypropylene Vaginal Mesh Grafts in Gynecology, Ob & Gyn, Oct. 2010-Vol. 116-Issue 4-pp 962-966
Ostegard, Donald R. and Alfred E. Bent, UROGYNECOLOGY AND URODYNAMICS: THEORY AND PRACTICE 456-7 (3rd ed. 1991).
Olsen et al., Obstet Gynecol 1997;89:501-6
Ou, C.S., and R. Rowbotham, Five-year Follow-up of Laparoscopic Bladder Neck Suspension using Synthetic Mesh and Surgical Staples, 9 J. Laparoendosc. Adv. Surg. Tech. A. 249-52 (1999).
Palma, P.C., Which Sling for Which Patient?, Int. Urogynecol. J. 15 Pelvic Floor Dysfunction 369-72 (2004).
Patel H, Effect of the Material, Int Urogynecol J (2012) 23:669-679
Patel, H., Polypropylene Mesh and the Host Response, 23 Int. Urogynecol. J. 669-79 (2012).
Petit, J., et al., [Experimental evaluation of tissue reactions around prostheses of the abdominal wall made of dacron gauze as a function of the duration of implantation and of its location in depth], 107 J. Chir. (Paris) 667-72 (1974) (Article in French).
Peyrat, L., et al., Intestinal perforation as a complication of TVT procedure for UI, 39 Eur. Urol. 603-5 (2001).
Rapoport, D., et al., Reported Complications of TVT Procedures: A Review, 49 BCMJ 465-524 (2007).
Steege, J., and D. Zolnoun, Evaluation and Treatment of Dyspareunia, 113 Obstetrics & Gynecology 1124-1136 (2009).
Rapoport, D., et al. Reported Complications of TVT Procedures: A Review. BCMJ, Vol. 49, No. 9, November 2007, pages 465-524

Rardin, C.R., et al., Tension-Free Vaginal Tape: Outcomes Among Women With Primary Versus Recurrent Stress Urinary Incontinence, 100 Obstet. Gynecol. 893-97 (2002).
Ross, J.W., Comparison of two techniques of laparoscopic burch repair for stress incontinence, 2 J. Am. Assoc. Gynecol. Laparosc. S47 (1995).
Ross, J.W., Laparoscopic Burch colposuspension and overlapping sphincteroplasty for double incontinence, 5 JSLS 203-9 (2001).
Ross, S et al., TOT COMPARED WITH TVT FOR STRESS INCONTINENCE-A RCT, Ob/Gyn, Vol. 114, No. 6, Dec. 2009 pp. 1287-1294
Scott, HD, et al. Rhode Island physicians' recognition and reporting of adverse drug reactions. RIMed J. 1987;70:311316
Shah H.N. and GH Badlani, Mesh complications in female pelvic floor reconstructive surgery and their management: A systematic review, 28 Indian J. Urol. 129-53 (2012).
Smith, TM, Delancey, JO et al., Pathologic evaluation of explanted vaginal mesh: interdisciplinary experience from a referral center, Female Pelvic Med Reconstr Surg. 2013 Jul-Aug;19(4):238-41
Stanford, E.J., and MF Paraiso, A Comprehensive Review of Suburethral Sling Complications, 15 J. Minim. Invasive Gynecol. 132-145 (2008).
Steege, John F MD; Zolnoun, Denniz MD, MPH, Evaluation and Treatment of Dyspareunia, Obstetrics & Gynecology: May 2009-Vol. 113-Issue 5-pp 1124-1136
Sternschuss, G. et al., Post-implantation alterations of polypropylene in the human, 188 J. Urol. 27-32 (2012).
Sweet, Vaginal ecosystem, 1990
Tseng, LH, et al Int Urogyn J (2005) 16: 230-235
Tunitsky-Biton Elena MD et al., Ultrasound Evaluation of Midurethral Sling Position & Correlation to Physical Examination & Patient Symptoms. FPM&RS Sep/Oct 2015 issue 5 p 263-68
Tunn, R., et al., Sonomorphological evaluation of polypropylene mesh implants after vaginal mesh repair in women with cystocele or rectocele, 29 Ultrasound Obstetrics Gynecol. 449-52 (2007).
Ulmsten U, Falconer C, Johnson P, Jomaa M, Lannér L, Nilsson CG, Olsson I. Int Urogynecol J Pelvic Floor Dysfunct. 1998;9(4):210-3. A multicenter study of tension-free vaginal tape (TVT) for surgical treatment of stress urinary incontinence.
Vierhout, M.E. Severe Hemorrhage Complicating TVT: a case Report, 12 Int. Urogynecol. J. Pelvic Floor Dysfunct. 139-40 (2001).
Vollebregt, A., et al., Bacterial colonisation of collagen-coated polypropylene vaginal mesh: are additional intraoperative sterility procedures useful?, 20 Int Urogynecol J Pelvic Floor Dysfunct. 1345 (2009).
Wall, LL & D. Brown, The Perils of Commercially Driven Surgical Innovation, 202.30e Am J. Obstet Gynecol. 1-4 (2010).
Walters & Karram, Problems With Synthetic Slings, Urogynecology & Reconstructive Pelvic Surgery 3rd Ed. 2007
Walters, M.D., et al., Vascular injury during TVT for SUI, 98 Obstet. Gynecol. 957-9 (2001).

Wei JT, et al, A midurethral sling to reduce incontinence after vaginal prolapse repair. N Engl J Med. 2012 Jun 21;366(25):2358-67
Withagen 2011
Yahim Y. et al., 18 Int. Urogyn. J. Sl 49 (2007).
Zilbert, A.W. & Farrell, S.A. External iliac artery laceration during TVT procedure, 12 Int. Urogyn. J. Pelvic Floor Dysfunction. 141-3 (2001).
TRC First Amended Ethicon LLC Specific 30 B 6 dated 4-8-13
Dow Product Safety Assessment
Sunoco PP MSDS
Secant Val Proto TVT Blue
Secant from Cornelia 2004 Validation
History of Mesh Production and Processing
Erosion Extrusion Issue Reports 2001-2002
Sunoco PP MSDS
Secant Val Proto TVT Blue
History of Mesh Production and Processing
TVT Patient Brochure Index 8-26-13
Mesh Complications Review
ACOG Committee on Gynecologic Practice
Chevron Philips, Material Safety Data Sheet 1/28/2004
Bloomberg Businessweek, June 26, 2013
October 2008 FDA Warning
July 13, 2011 FDA News Release
TeLinde's Operative Gynecology 2014
AUGS 2014
U.S. Census Data 2006
US Bureau of the Census 1994
<a href="http://www.fda.gov/Medical Devices">www.fda.gov/Medical Devices</a>
1/4/2016 FDA News Release
FDA Executive Summary 2011
FDA Public Health Notification (October 20, 2008).
FDA Safety Communication (July 13, 2011).

21 C.F.R. Part 807

<b>Author</b>	<b>Name</b>	<b>Journal Citation</b>
Abbott S, Unger CA	Abbott [2014] Evaluation and management of complications from synthetic mesh after pelvic reconstructive surgery	Am J obstet Gynecol 2013;210:163.e1-8
Abdel-Fattah M, Barrington JW, et al	Pelvicol pubovaginal sling versus tension-free vaginal tape for treatment of urodynamic stress incontinence: a prospective randomized three- year follow-up study	Eur Urol. 2004 Nov;46(5):629-35
Abdel-fattah M, et al	Primary and repeat surgical treatment for female pelvic organ prolapse and incontinence in parous women in the UK: a register linkage study	BMJ Open 2011; 1:e000206
Abdel-Fattah M, et al	Lower urinary tract injuries after transobturator tape insertion by different routes: a large retrospective study	GJOG 2006; 113: 1377-1381
Abdel-Fattah M, Familusi A, et al	A randomised prospective single-blinded study comparing inside-out versus outside-in transobturator tapes in the management of female stress urinary incontinence (E-TOT study): 3 years follow-up.	Neurourol Urodyn 2011;30:825–826.
Abdel-Fattah M, Ford JA, et al	Single-Incision Mini-Slings Versus Standard Midurethral Slings in Surgical Management of Female Stress Urinary Incontinence: A Meta Analysis of Effectiveness and Complications.	European Urology ,60: 468-80,24-May-11
Abdel-Fattah M, Mostafa A, et al	Prospective randomised controlled trial of transobturator tapes in management of urodynamic stress incontinence in women: 3-year outcomes from the Evaluation of Transobturator Tapes study.	Eur Urol. 2012 Nov;62(5):843-51. doi: 10.1016/j.eururo.2012.04.021. Epub 2012 Apr 14.
Abdel-fattah M, Ramsay I, et al	Evaluation of transobturator tapes (E-TOT) study: randomised prospective single-blinded study comparing inside-out vs. outside-in transobturator tapes in management of urodynamic stress incontinence: short term outcomes.	Eur J Obstet Gynecol Reprod Biol. 2010 Mar;149(1):106-11. doi: 10.1016/j.ejogrb.2009.11.023. Epub 2009 Dec 24



Abdel-Fattah M, Ramsay I, et al	Randomised prospective single-blinded study comparing 'inside-out' versus 'outside-in' transobturator tapes in the management of urodynamic stress incontinence: 1-year outcomes from the E-TOT study.	BJOG,117:870-8 ,12-Apr-10
Abdel-Fattah M, Sivanesan K, et al	How common are tape erosions? A comparison of two versions of the transobturator tension-free vaginal tape procedure.	BJU Int. 2006 Sep;98(3):594-8.
Abdelnaser KH, et al	The use of polypropylene mesh as a transobturator sling for the treatment of female stress urinary incontinence (early experience with 40 cases)	Int Urogynecol J 2008; 19: 833-838
Abdelwahab O, Shedid I, et al	Tension-free vaginal tape versus secure tension- free vaginal tape in treatment of female stress Urinary Incontinence2	Current Urology(2010) 4 (2): 93-98
Abed H, Rahn DD, et al	"Incidence and Management of Graft Erosion, Wound Granulation, and Dyspareunia Following Vaginal Prolapse Repair with Graft Materials: A Systematic Review "	The International Urogynecology Journal,22:789- 98,22-Mar-11
Abouassaly R, Steinberg JR, et al	Complications of tension-free vaginal tape surgery: a multi-institutional review.	BJU International,94:110-3 ,01-Jul-04
Abrams P et al	Synthetic Vaginal Tapes for Stress Incontinence: Proposals for Improved Regulation of New Devices in Europe	European Urology 60 (2011) 1207-1211
Abrams T	The Regulation of Prescription Drug Promotion	In Ethics and the Pharmaceutical Industry, 2005; 153-169
Achtari C, et al	Anatomical study of the obturator foramen and dorsal nerve of the clitoris and their relationship to minimally invasive slings	Int Urogynecol J (2006) 17: 330-334
ACOG	Patient Safety in Obstetrics and Gynecology	ACOG Committee opinion number 447, December 2009
ACOG	ACOG Committee Opinion 439 Informed Consent	ACOB 439 (2009)
Adile B, Granese R, Lo Bue A, et al	A prospective randomized study comparing laparoscopic Burch versus TVT: short and long term follow-up.	<a href="http://www.ics.org/Abstracts/Publish/41/000550.pdf">http://www.ics.org/Abstracts/Publish/41/000550.pdf</a>
Afonso JS, et a	Structural and thermal properties of polypropylene mesh used in treatment of stress urinary incontinence	Acta of Bioengineering and Biomechanics (2009) 11(3):
Agarwala N.	A randomized comparison of two synthetic mid- urethral tension-free slings.	UroToday Int J. 2008 Oct;1(4)



Agnew G, Dwyer PL, et al	Functional outcomes following surgical management of pain, exposure or extrusion following a suburethral tape insertion for urinary stress incontinence <sup>2</sup>	Int Urogynecol J (2014) 25:235-239
Aigmueller, T.	Reason for dissatisfaction ten years after TVT procedure	Int Urogynecol J (2014) 25;213-217
Alajmo F	Polypropylene Suture Fracture <sup>2</sup>	Ann Thorac Surg 1985 39.4: 400
Albo ME, et al	Treatment success of retropubic and transobturator mid urethral slings at 24 months	J urology 2012 (188) 2281-2287
Albo ME, Richter HE, et al	Burch colposuspension versus fascial sling to reduce urinary stress incontinence <sup>2</sup>	New England Journal of Medicine,356:2143- 55,21-May-07
Alcalay, Menachem	Burch colposuspension: a 10-20 year follow up	British Journal of Obstetrics and Gynaecology September 1995, Vol. 102, pp. 740-745
Aldrete V	Polpropylene Suture Fracture	Ann Thorac Surg 1984 Mar; 37(3):264
Alinsod R, et al	Initial outcomes of a stabilized adjustable minisling for female urinary stress incontinence	Abstract 865 (2009)
Al-Omary R, et al	Long term patient satisfaction after suburethral sling operation for stress incontinence	Abstract 381 Int Urogynecol J (2011) ss (Suppl 3): S1769- S2008
Altman AJ, Gorn RA, et al	The breakdown of polypropylene in the human eye: is it clinically significant?	Ann Ophthalmol 1986 May; 18(5) 182-5
Altman D, Väyrynen T, et al	Anterior colporrhaphy versus transvaginal mesh for pelvic-organ prolapse.	New England Journal of Medicine,364:1826-36, 5/12/2011
Altuna S, et al	Lower urinary tract injuries associated with the out-in transobturator tape: is cystoscopy required? An argentinean multicenter experience	Int Urogynecol J (2007) 18 (Suppl I) S163-S164
Amaro, JL, et al	Clinical and qualify-of-life outcomes after autologus faxcial sling and tension-free vaginal tape: A prospective randomized trial	International Braz U Urol (2009) 35(1): 60-67
Amat I Tardiu L, Martínez et al	Contasure-Needleless compared with transobturator-TVT for the treatment of stress urinary incontinence.	Int Urogynecol J. 2011 Jul;22(7):827-33. doi: 10.1007/s00192-011-1380-9. Epub 2011 Mar 2.
An, YH	Concise Review of Mechanisms of Bacterial Adhesion to Biomaterial Surfaces	J Biomed Mater Res (Appl Biomater) 1998;43:338-48
Anderson, HA	Utilization of Adipose Tissue Biopsy in Characterizing Human Halogenated Hydrocarbon Exposure	Environ Health Perspect. 1985 May; 60:127-131

Anderson-Smits C	Developing of the Pelvic Floor Disorder Registry	www.fda.gov
Andonian S, Chen T, et al	Randomized Clinical Trial Comparing Suprapubic Arch Sling (SPARC) and Tension-Free Vaginal Tape (TVT): One-Year Results	European Urology,47:537-41,13-Jan-05
Andonian S, St-Denis B, et al	Prospective clinical trial comparing Obtape and DUPS to TVT: one-year safety and efficacy results.	Eur Urol. 2007 Jul;52(1):245-51. Epub 2007 Jan 10
Andrada Hamer M, Larsson PG, et al	One-year results of a prospective randomized, evaluator-blinded, multicenter study comparing TVT and TVT Secur.	Int Urogynecol J. 2013 Feb;24(2):223-9
Andrada Hamer M, Larsson PG, et al	Short-term Results of a Prospective Randomized Evaluator Blinded Multicenter Study Comparing TVT and TVT-Secur.	International Urogynecology Journal,22:781- 787,16-Apr-11
Anger JT, Litwin MS, Wang Q, et al	Complications of sling surgery among female Medicare beneficiaries <sup>2</sup>	Obstetrics & Gynecology,109:707-14,01-Mar-07
Anger JT, Weinberg AE, et al	Trends in surgical management of stress urinary incontinence among female Medicare beneficiaries	Urology,74:283-7,07-Jun-09
Angioli R, Plotti F, Muzii L, et al	"Tension-free vaginal tape versus transobturator suburethral tape: five-year follow-up results of a prospective, randomised trial"	European Urology 58 (2010) 671-677
Aniulienė R.	Tension-free vaginal tape versus tension-free vaginal tape obturator (inside-outside) in the surgical treatment of female stress urinary incontinence.	Medicina (Kaunas). 2009;45(8):639-43
Ansquer Y, et al	The suburethral sling for female stress urinary incontinence: a retropubic or obturator approach?	J Am Assoc Gynecol Laparosc 2004, 11(3):353- 358
Anthanasiou S, Grigoriadis T, et al	Mixed Urodynamic Incontinence: TVT or TVT-O?	Int Urogynecol J (2009) 20 (Suppl 2): S73-S239
Apple DJ, Mamalis N, et al	Biocompatibility of implant materials: a review and scanning electron microscopic study	J Am Intraocul Implant Soc 1984 Winter; 10(1):53-66
Apte, et al	Pain after suburethral sling insertion for urinary stress incontinence <sup>2</sup>	Pain Practice 2012; 12(2):88-110
Araco F., et al	TVT-O vs TVT a randomized trial in patients with different degrees of urinary stress incontinence.	Int Urogynecol J 2008; 19:917-926

Arunkalaivanan A, et al	Efficacy and safety of transbutorator tape (Obtryx™) in women with stress urinary incontinence and intrinsic sphincter deficiency: results from international Obtryx™ registry	ICS 2009 Abst. 778
Arunkalaivanan AS and Barrington JW	Randomized trial of porcine dermal sling (Pelvicol implant) vs. tension-free vaginal tape (TVT) in the surgical treatment of stress incontinence: a questionnaire-based study	Int Urogynecology J (2003) 14: 17-23
Asicioglu, O	A 5-year follow-up study comparing Burch colposuspension and transobturator tape for the surgical treatment of stress urinary incontinence.	Int J Gynecol Obstet.( 2014) April 125(1): 73-77
Atassi Z, Reich A, Rudge A, et al	Haemorrhage and nerve damage as complications of TVT-O procedure: case report and literature review	Arch Gynecol Obstet 2008, 277:161-4
Athanasίου S, Grigoriadis T, et al	Seven years of objective and subjective outcomes of transobturator (TVT-O) vaginal tape: Why do tapes fail?	Int Urogynecol J. 2013 Jul 27. [Epub ahead of print]
Atherton MJ and Stanton SL.	The tension-free vaginal tape reviewed: an evidence-based review from inception to current status.	BJOG,112:534-46, 5/1/2005
Atis G, Arisan S, Ozagari A, et al	Tissue reaction of the rat urinary bladder to synthetic mesh materials.	Scientific World Journal,9:1046-51,02-Oct-09
AUA	Position Statement on the Use of Vaginal Mesh for the Surgical Treatment of Stress Urinary Incontinence	
AUA	Guideline for the Surgical Management of Female Stress Urinary Incontinence: 2009 Update	
AUA	AUA Newsletter	13-Oct
AUGS	Position Statement on Restriction of Surgical Options for Pelvic Floor Disorders	
AUGS	Corporate Members"	<a href="http://www.augs.org/p/cm/ld/fid=24">http://www.augs.org/p/cm/ld/fid=24</a>
AUGS/SUFU	Position Statement on Mesh Midurethral Slings for Stress Urinary Incontinence	
Baessler K, et al	Severe mesh complications following intravaginal slingplasty	Obstet Gynecol 2005; 106:713-6
Bai SW, Sohn WH, Chung DJ, et al	"Comparison of the efficacy of Burch colposuspension, pubovaginal sling, and tension- free vaginal tape for stress urinary incontinence"	Int J Gynaecol Obstet. 2005 Dec;91(3):246-51. Epub 2005 Oct 20

Ballard, J., et al	Harnessing a Health Information Exchange to Identify Surgical Device Adverse Events for Urogynecologic Mesh	Abstract
Banks CL, et al	Abscess formation following trans-obturator tape procedures	Int Urogynecol J (2006) 17 (Suppl 2):S171-S359 Abst 204
Barber M, et al	A multicenter randomized trial comparing the transobturator tape with tension-free vaginal tape for the surgical treatment of stress urinary incontinence	Abst 113
Barber MD, Brubaker L, et al	Defining success after surgery for pelvic organ prolapse.	Obstetrics and Gynecology, 114:600-9, 01-Sep-09
Barber MD, Kleeman S, et al	Transobturator tape compared with tension-free vaginal tape for the treatment of stress urinary incontinence: a randomized controlled trial.	Obstetrics & Gynecology, 111:611-21, 01-Mar-08
Barber MD.	"Cleveland Clinic leads multicenter trial of single-incision 'minisling' for stress urinary incontinence."	Ob/GYN & Women's Health Research Perspectives. Summer 2011. <a href="http://clevelandclinic.org/obgyn">clevelandclinic.org/obgyn</a>
Barber, M., et al	Single-Incision Mini-Sling Compared With Tension-Free Vaginal Tape for the treatment of Stress Urinary Incontinence	Obstetrics & Gynecology, Vol. 119, No. 2, Part 1 (Feb 2012)
Barone WR, et al	The impact of boundary conditions on surface curvature of polypropylene mesh in response to uniaxial loading	J Biomechanics, Accepted 28 Feb 2015, In Press
Barone, DVM, MS, Mark A, et al	Determinants of Postoperative Outcomes of Female Genital Fistula Repair Surgery.	Obstet Gynecol, 2012 September; 120(3); 524-531, doi:10.1097/AOG.0b013e31826579e8.
Barry C, Lim YN, Muller R, et al	A multi-centre, randomised clinical control trial comparing the retropubic (RP) approach versus the transobturator approach (TO) for tension-free, suburethral sling treatment of urodynamic stress incontinence: the TORP study	The International Urogynecology Journal, 19:171-8, 19-Jul-07
Basok EK, Yildirim A, Atsu N, et al	Cadaveric fascia lata versus intravaginal slingplasty for the pubovaginal sling: surgical outcome, overall success and patient satisfaction rates.	Urol Int 2008;80(1):46-51
Basu M, et al	Three-year results from a randomised trial of a retropubic mid-urethral sling versus the Miniarc single incision sling for stress urinary incontinence	Int Urogynecol J 2013

Basu M, et al	A randomised trial of a retropubic tension-free vaginal tape versus a mini-sling for stress incontinence	BJOG 2010; 117:730-735
Bekelman JE, et al	Scope and Impact of Financial Conflicts of Interest in Biomedical Research A Systematic Review	JAMA, January 22/29, 2003 Vol 289 No. 4, 454-465
Bemelmans BLH, Chapple CR.	Are slings now the gold standard treatment for the management of female urinary stress incontinence and if so which technique?	Curr Opin Urol 2003;13:301-307
Bendavid R Iakovlev V, et al	Mesh-Related SIN Syndrome. A Surreptitious Irreversible Neuralgia and Its Morphologic Background in the Etiology of Post-Herniorrhaphy Pain2	Int J Clinical Med (2014)(5) 799-810
Bezerra CA and Bruschini H.	Suburethral sling operations for urinary incontinence in women.	Cochrane Database Syst Rev.
Bianchi-Ferraro AM, Jarmy-Di et al	Single-incision sling compared with transobturator sling for treating stress urinary incontinence: a randomized controlled trial.	Int Urogynecol J. 2013 24: 1459-1465
Binnebosel M, et al.	Biocompatibility of prosthetic meshes in abdominal surgery	Semin Immunopathol. 2011; [Deleted Object]:235-243
Birolini C, Minossi JG, et al	Mesh Cancer: Long-Term Mesh Infection Leading to Squamous-Cell Carcinoma of the Abdominal Wall	Hernia,19-Apr-13
Blaivas J, Purohit R, et al	Salvage Surgery after Failed Treatment of Synthetic Mesh Sling Complications	J Urology, 2013 Oct, Vol 190, 1281-1286
Blaivas JG and Chaikin DC.	Pubovaginal fascial sling for the treatment of all types of stress urinary incontinence: surgical technique and long-term outcome.	The Urologic Clinics of North America,38:7-15,00- Jan-00
Blaivas JG, et al	Safety considerations for synthetic sling surgery	Nat Rev Urol 2015 Aug 18 doi: 10.1038/nrurol.2015.183 [Epub ahead of print]
Blandon, R.E., Gebhart, J.B., et al	(2009). Complications from vaginally placed mesh in pelvic reconstructive surgery.	Int. Urogynecol J Pelvic Floor Dysfunct, 20(5), 523-531. doi: 10/1007/s00192-009-0818-9.
Bodelsson G, et al	Short term complications of the tension free vaginal tape operation for stress urinary incontinence in women	BJOG (2002) 109: 566-569
Bonnet P, Waltregny D, et al	Transobturator vaginal tape inside out for the surgical treatment of female stress urinary incontinence: Anatomical considerations	The Journal of Urology,173:1223-8,01-Apr-05
Bouillot JL, et al	Parietal mesh abscess as an original presentation of cancer of the caecum	Digestive Surgery (1999) 16(2) 158-160

Bowler S	Preparing articles for publication in peer-reviewed journals	American College of Preventive Medicine www.acpm.org
Boyles SH, et al	Complications associated with transobturator sling.	Int Urogynecol J (2007) 18:19-22
Brill AI.	The hoopla over mesh: what it means for practice.	Obstet Gynecol News,47:14-5 ,01-Jan-12
Brubaker L,	5-Year Continence Rates, Satisfaction and Adverse Events of Burch Urethropexy and Fascial Sling Surgery for Urinary Incontinence <sup>2</sup>	The Journal of Urology
Brubaker L, Cundiff GW, et al	Abdominal sacrocolpopexy with Burch colposuspension to reduce urinary stress incontinence.	New England Journal of Medicine,354:1557- 66,13-Apr-06
Brubaker L, et al	Adverse events over two years after retropubic or transobturator midurethral sling surgery findings from the Trial of Midurethral Slings (TOMUS) study	Am J Obstet Gynecol 2011; 205: 498.e1-6
Brubaker L, Maher C, et al	Surgery for pelvic organ prolapse.	Female Pelvic Medicine & Reconstructive Surgery,16:9-19,01-Jan-10
Brubaker L, Nygaard I, et al	Two-year outcomes after sacrocolpopexy with and without burch to prevent stress urinary incontinence.	Obstetrics and Gynecology,112:49-55, 7/1/2008
Brubaker, L.	(2006). Editorial: partner dyspareunia (hispareunia).	Int Urogynecol J Pelvic Floor Dysfunt, 17(4), 311. doi: 10.1007/s00192-006-0097-7
Brubaker, L., et al	Missing data frequency and correlates in two randomized surgical trials for urinary incontinence in women	Int Urogynecol J Published online 03/24/2015
Burton C, et al	Comments on Araco et al.: TVT-O vs TVT: a randomized trial I patients with different degrees of urinary stress incontinence	Int Urogynecol J 2009;20:369
Burns-Heffner, C., et al	Testing Compliance of Surgical Meshes Fabricated from Different Polymeric Biomaterials	Abstract #877 2014 Society for Biomaterials
But I and Faganelj M.	Complications and short-term results of two different transobturator techniques for surgical treatment of women with urinary incontinence: a randomized study.	The International Urogynecology Journal,19:857- 61,01-Jun-08
Cadish L, Hacker M, et al	Characterization of Pain After Inside-Out Transobturator Midurethral Sling	Female Pelvic & Reconstructive Surgery Vol 20, (2) March/April 2014
Calhoun TR, Kitten	Polypropylene suture -- Is it safe?	J Vasc Surg 1986; 4:98-100



DM		
Calvo JJ, Alfara AH, et al	"Stress urinary incontinence surgery with MiniArc sling system: Our experience."	Actas Urologicas Espanolas 2010;34(4):372-377
Cameron AP, et al	The treatment of female stress urinary incontinence: an evidenced-based review	Open Access J Urology 2011;3 109-120
Campeau L, Tu LM, Lemieux MC, et al	A Multicenter, Prospective Randomized Clinical Trial Comparing Tension-Free Vaginal Tape Surgery and No Treatment for the Management of Stress Urinary Incontinence in Elderly Women	Neurourology and Urodynamics 26:990-994 (2007)
Cao J., et al	In Vitro Study: Synthetic Prosthetic Meshes for Inguinal Hernia Repair	Presented at AATCC International Conference May 18-20 in Atlanta, GA AATCC Review November/December 2011
Capobianco G, Dessole M, et al	TVT-Abbrevio: efficacy and two years follow-up for the treatment of stress urinary incontinence	Clin Exp Obstet Gynecol (2014); 41(4): 445-447
Caquant, F., et al.	Safety of Trans Vaginal Mesh procedure: retrospective study of 684 patients.	J Obstet Gynaecol Res. 2008; 34(4):449-456
Carey M, Higgs P, Goh J, et al	Vaginal repair with mesh versus colporrhaphy for prolapse: a randomised controlled trial.	BJOG,116:1380-6,07-Jul-09
Carey M, Slack M, et al	Vaginal surgery for pelvic organ prolapse using mesh and a vaginal support device	BJOG 2008; 115:391-397
Caruso S, Rugolo S, et al	Clitoral Blood Flow Changes After Surgery for Stress Urinary Incontinence: Pilot Study on TVT Versus TOT Procedures	Urology 70: 554-557, 2007
Casiano ER, et al	Does concomitant prolapse repair at the time of midurethral sling affect recurrent rates of incontinence	Int Urogynecol J (2011) 22:819-825
Cattoni,Elena; Serati M, et al	ISU Abs 5 Can preoperative overactive bladder (OAB) symptoms influence TVT-O outcome?	Neurolurol and Urodyn; International Society of Urology Conference abstracts 2012
Cayan, R.	Sexual function after surgery for stress urinary incontinence: vaginal sling versus Burch colposuspension.	Gynecol Obstet. 2008 Jan; 277(1):31-6 Epub 2007 Jul 25.
Celebi I, et al	Results of the tension-free vaginal tape procedure for treatment of female stress urinary incontinence: a 5-year follow-up study.	Arch Gynecol Obstet (2009) 279: 463-467

Cervigni M, Natale F, et al	Surgical Correction of Stress Urinary Incontinence Associated with Pelvic Organ Prolapse: Trans- Obturator Approach (Monarc) Versus Retropubic Approach (TVT)	Int Urogynecol J 2006;17(suppl. 2):S215 abstract 280. 31st annual IUGA meeting, Athens, Greece, 6-9 September 2006
Chae HD, Kim SR, Jeon GH, et al	A comparative study of outside-in and inside-out transobturator tape procedures for stress urinary incontinence.	Gynecol Obstet Invest. 2010;70(3):200-5. doi: 10.1159/000318866. Epub 2010 Jul 17.
Challoner, D, Korn, D	Medical Devices and the Public's Health The FDA 510(k) Clearance Process at 35 Years <sup>2</sup>	Institute of Medicine of the National Academies Report Brief July 2011
Challoner, DR,	2011 - IOM Committee on the Public-Health Effectiveness of the FDA 510(k) Clearance Process - letter to FDA re 510k	IOM 2011
Challoner, DR, et al	Medical Devices and Health -- Creating a New Regulatory Framework for Moderate-Risk Devices	N Engl J Med 365:11: 977-979
Chapple C, Raz S, Brubaker L, et al	Mesh Sling in an Era of Uncertainty: Lessons Learned and the Way Forward	European Urology 64 (2013) 525-529
Charalambous S, Touloupidis S, et al	Transvaginal vs transobturator approach for synthetic sling placement in patients with stress urinary incontinence.	Int Urogynecol J Pelvic Floor Dysfunct. 2008 Mar;19(3):357-60. Epub 2007 Aug 29.
Chen X, Tong X, Jiang M, et al	A modified inexpensive transobturator vaginal tape inside-out procedure versus tension-free vaginal tape for the treatment of SUI: a prospective comparative study	Archives of Gynecology and Obstetrics, 284:1461- 6 ,22-Mar-11
Chen YH, Wang YJ, Li FP, and Wang	Efficacy and postoperative complication of tension-free vaginal tape-Secur for female stress urinary incontinence	Chinese Medical Journal, 124:1296-9 ,01-May-11
Chen Z, Chen Y, Du GH, et al	Comparison of three kinds of mid-urethral slings for surgical treatment of female stress urinary incontinence	Urologia. 2010 Jan-Mar;77(1):37-41; discussion 42.
Cheng D, Liu C.	Tension-free vaginal tape-obturator in the treatment of stress urinary incontinence: a prospective study with five-year follow-up.	Eur J Obstet Gynecol Reprod Biol. 2012 Apr;161(2):228-31. doi: 10.1016/j.ejogrb.2012.01.011. Epub 2012 Feb 13.
Chinthakanan O, et al	Mesh removal following sling/mesh placement: a multicenter study <sup>2</sup>	2014-A-1236-AUGS/IUGA
Chinthakanan O, et al	Indication and surgical treatment of midurethral sling complications: a multicenter study	2014-A-1239-AUGS/IUGS
Chinthakanan O, Miklos JR, et al	Mesh Removal Following Sling/Mesh Placement: A Multicenter Study	Int Urogynecol J(2014)25(Suppl 1):S139-S140



Chinthakanan O, Miklos JR, et al	Indication and surgical treatment of midurethral sling complications	Int Urogynecol J (2014) 25 (Suppl 1):S144-S143
Cho MK, Kim CH, Kang WD, et al	Comparison of the clinical and quality-of-life outcomes after the inside-out TVT-O procedure with or without concomitant transvaginal gynaecological surgery.	J Obstet Gynaecol. 2012 Apr;32(3):280-4. doi: 10.3109/01443615.2011.654290
Choe JH, Kim JH, Na TG, et al	Comparative Study of Tension-Free Vaginal Tape (TVT) and Suprapubic Arc (SPARC) Sling Procedure for Female Stress Urinary Incontinence	International Urogynecology Journal 2005
Choi J, et al.	Use of Mesh During Ventral Hernia Repair in Clean Contaminated and Contaminated Cases	Annals of Surgery, Vol 255, Number 1, Jan 2012
Cholhan HJ, et al	Pre-pubic approach to mid-urethral slings: 3-month interim report on peri-operative experience and complications	2007 Abstract
Cholhan HJ, Hutchings TB, et al	Dyspareunia associated with paraurethral banding in the transobturator sling <sup>2</sup>	Am J Obstet Gynecol (2010);202:481.e1-5
Chrysostomou A	The management of stress urinary incontinence using transobturator tapes in a tertiary hospital in South Africa	Abst 0177 Oral presentations/Int J Gynecol & Obstet 10752 (2009) S93-S396
Chung C, Kingman T, Tsai L, et al	Serious Complications From a Single-Incision Midurethral Sling Placement	Obstet & Gynecol, February 2012, Vol 119, No. 2 Part 2, 464-466
Clave', A., et al	Polypropylene as a Reinforcement in Pelvic Surgery in Not Inert: Comparative Analysis of 100 Explants.	Int Urogyn J 2010; 21:261-270
Claymen HM	Polypropylene	Ophthalmology 1981 88:959-976
Cobb WS, et al	The argument for lightweight polypropylene mesh in hernia repair	Surg Innov 2005 12: 63-69
Cobb, W., et al.	The Argument for Lightweight Polypropylene Mesh in Hernia Repair	Surgical Innovation 2005, 12(1):T1-T7
Coda, A., et al	Structural Alterations of Prosthetic Mesh in Humans	Hernia. 2003; Mar;7(1):29-34
Collinet P, Ciofu C, et al	Safety of inside-out transobturator approach for urinary stress incontinence treatment: prospective multicentric study of 994 patients – French TVT-O registry	Int Urogynecol J (2008) 19:711-715
Colombo M, Vitabello D, et al	Randomized study to compare pereyra and TVT procedures for women with stress urinary incontinence and advanced urogenital prolapse	International Urogynecology Journal August 2005, Volume 16, Issue 2 Supplement, pp S35- S130
CommonHealth	Surgery Under Scrutiny: What Went Wrong With Vaginal	<a href="http://commonhealth.wbur.org/2011/11">http://commonhealth.wbur.org/2011/11</a>

Constantini E, Lazzeri M, Kocjanci	ICS Abs 3 Prolonged follow-up shows continence deterioration after transobturator tape: Results from a randomised controlled trial.	International Continence Society Mtg 2013
Contemporary OB/GYN Staff	American Urogynecologic Society voices opposition to restrictions on transvaginal mesh	Contemporary OB/GYN May 01, 2013
Cornel G	Fracture of Polypropylene Suture	Ann Thorac Surg 1982; [Deleted Object]:641
Cornelis R, Hogewoning C, et al	The introduction of mid-urethral slings: an evaluation of literature (Update)	Int Urogynecol J (2015) 26: 229-234
Cornelis R, Hogewoning C, et al	The introduction of mid-urethral slings: an evaluation of literature	Int Urogynecol J (Published online 21 August 2014))
Cornu J-N, Haab F.	Mini-slings for female stress urinary incontinence: Not yet at the age of reason.	Eur Urol 2011;60:481-483
Cornu, JN, et al	Midterm prospective evaluation of TVT-Secur reveals high failure rate	European Urology 58 (2010) 157-161
Corona, R., et al	Tension-free Vaginal Tapes and Pelvic Nerve Neuropathy	J Min Invas Gynecol 2008; 15(3):262-267
Corporate ACTION Network	Putting Women at Risk, The Case Against Pelvic Mesh Report	Corporate Action Network August 2014
Corton M	Critical anatomic concepts for safe surgical mesh	Clin Obstet Gynecol 2013 Jun; 56(2):247-56
Coskun B, Zimmers PE, et al	Minislings can cause complications	Int Urogynecol J (2015) 26:557-562
Cosson M, et al	Mechanical properties of synthetic implants used in the repair of prolapse and urinary incontinence in women: which is the ideal material?	Int Urogynecol J (2003) 14: 169-178
Costa P	Comparisons of Safety and Efficacy of the Obtryx Sling and Advantage Mid-Urethral Sling for the Treatment of Stress Urinary Incontinence: Propensity Matching Results in a Large International Registry	Abstract 614 (2010)
Costa P, Ballanger P, et al	Transobturator Tape for SUI. Preliminary Results of Prospective Multicenter Register	ICS Abstract #379,05-Oct-03
Costa P, Grise P, et al	Surgical treatment of female stress urinary incontinence with a trans-obturator-tape (T.O.T.) Uratape: short term results of a prospective multicentric study	European Urology

Costello CR, et al	Materials Characterization of Explanted Polypropylene Hernia Meshes	J Biomed Mater Res Part B: Appl Biomater 83B: 44-49, 2007
Costello, C., et al	Characertization of Heavyweight and Lightweight Polypropylene Prosthetic Mesh Explants from a Single Patient.	Surgical Innovation. 2007; 14(3): 168-176
Cox A, Herschorn S, et al	Surgical management of female SUI: Is there a gold standard?	Nat Rev Urol 2013;10:78-89
Cozad MJ, et al	Materials characterization of explanted polypropylene, polyethylene terephthaiate, and expanded polytetrafluoroethylene composites: Spectral and thermal analysis	J Biomed Mater Res Part B: Appl Biomater 94B: 455-462, 2010
Crosby E, Abernethy M, et al	Symptom Resolution After Operative Management of Complications From Transvaginal Mesh2	Obstet & Gynecol January 2012, Vol 123, No. 1, 134-139
Cundiff GW, Varner E, et al	Risk factors for mesh/suture erosion following sacral colpopexy.	The American Journal of Obstetrics & Gynecology,199:688.e1-5,31-Oct-08
Curfman, GD., et al	Medical Devices -- Balancing Regulation and Innovation	N Engl J Med 365; 11: 975-977
Daher N, et al	Pre-pubic TVT: an alternative to calssic TVT in selected patients with urinary stress incontinence	European J Obstet & Gynecol and Reproduct Biology 107 (2003) 205-207
Daneshgari F, Kong W, et al	Complications of Mid Urethral Slings: Important Outcomes for Future Clinical Trials.	The Journal of Urology Vol 180, 1890-1897 November 2008.
Darai E, Frobert JL, et al	Functional results after the suburethral sling procedure for urinary stress incontinence: a prospective randomized multicentre study comparing the retropubic and transobturator routes.	Eur Urol 2007 Mar;51(3):795-801; discussion 801-2. Epub 2006 Sep 8
Das, N., et al	Review Article: Microbial Degradation of Petroleum Hydrocarbon Contaminants: An Overview	Biotechnology Res Int 2011: 1-13 Article ID 941810
Dati S, Cappello S, et al	Single-incision minsling (AJUST®) vs Obturator Tension-free vaginal shortened tape (TVT- Abbrevo™) in surgical management of female stress urinary incontinence	Dati poster XX FIGO World Congress of Gynecology and Obstetrice Rome, 7-12 October 2012
Dati S, et al	Obtryn (TM) system: transobturator out-in sling in the treatment of isolated or pop-associated urinary incontinence (Abstract only)	Int Urogynecol J (2007) 18 (Suppl I): S152

Dati S, Rombola P, et al	Single-Incision Minisling (Ajust) vs Obturator Tension-Free Vaginal Shortened Tape (TVT- Abbrevio) in Surgical Management of Female Stress Urinary Incontinence	Poster presentations/ Int J Gynecol & Obstet 119S3 (2012) S670
David-Montefiore E, et al	Peri-operative complications and pain after the suburethral sling procedure for urinary stress incontinence: a French prospective randomised multicentre study comparing the retropubic and transobturator routes.	European Urology,49:133-8 ,02-Nov-05
de Leval J, Thomas A, and Waltregn	The original versus a modified inside-out transobturator procedure: 1-year results of a prospective randomized trial	The International Urogynecology Journal,22:145- 56,21-Sep-10
de Leval J.	Novel surgical technique for the treatment of female stress urinary incontinence: Transobturator Vaginal Tape Inside-Out	European Urology,44:724-30,02-Oct-03
de Oliveira LM, Girao MJBC, et al	Comparison of Retro-Pubic TVT, Pre-Pubic TVT and TVT Transobturator in Surgical Treatment of Women With Stress Urinary Incontinence	International Urogynecology Journal
De Ridder D, et al	Single incision mini-sling versus a transobutator sling: a comparative study on MiniArc and Monarc slings	Int Urogynecol J (2010) 21:773-778
De Souza A	Sexual function following retropubic TVT and transobturator Monarch sling in women with intrinsic sphincter deficiency: a multicentre prospective study .	Int Urogynecol J (2012) 23:153-158
de Tayrac R, Deffieux X, et al	A transvaginal ultrasound study comparing transobturator tape and tension-free vaginal tape after surgical treatment of female stress urinary incontinence.	Int Urogynecol J (2006) 17: 466-471
de Tayrac R, Deffieux X, et al	A prospective randomized trial comparing tension- free vaginal tape and transobturator suburethral tape for surgical treatment of stress urinary incontinence.	American Journal of Obstetrics and Gynecology,190:602-8 ,01-Mar-04
de Tayrac R, Droupy S, et al	A Prospective Randomized Study Comparing TVT and Transobturator Suburethral Tape (TOT) for the Surgical Treatment of Stress Incontinence	ICS Abstract #344,,05-Oct-03
de Tayrac, R et al.	Long-term anatomical and functional assessment of trans- vaginal cystocele repair using a tension- free polypropylene mesh.	Int Urogynecol J (2006) 17: 483-488.

de Tayrac, R. & Letouzey, V.	Basic science and clinical aspects of mesh infection in pelvic floor reconstructive surgery.	Int Urogynecol J, 22(7), 775-780. doi: 10.1007/s00192-011-1405-4.
Debodinance P, Legrange E, et al	TVT Secur: Prospective Study and Follow up to 1 Year about 150 Patients	International Urogynecology Journal,19(Supp. 1):211-S12 ,00-Jan-00
Debodinance P.	Trans-obturator urethral sling for surgical correction of female stress urinary incontinence: Outside-in (Monarc) versus inside-out (TVT-O). Are the two ways reassuring?	Eur J Obstet Gynecol Reprod Biol,133:232-8
Deffieux X, Daher N, et al	Transobturator TVT-O versus retropubic TVT: results of a multicenter randomized controlled trial at 24 months follow-up	Int Urogynecol J. 2010 Nov;21(11):1337-45.
Delorme E	Transobturator urethral suspension: mini-invasive procedure in the treatment of stress urinary incontinence in women	Prog Urol 2001 Dec; 11(6):1306-13
Delorme E, Droupy S, et al	Transobturator tape (Uratape): a new minimally- invasive procedure to treat female urinary incontinence	European Urology,45:203-7 ,01-Feb-04
Demirkesen, O	Does vaginal anti-incontinence surgery affect sexual satisfaction? A comparison of TVT and Burch –colposuspension.	Int Braz J Urol. 2008 Mar- Apr; 34 (2):214-9
Demirkesen, O, et al	Are the effectiveness and complication rates of transoburator and tension-free vaginal tape similar?	Türk Üroloji Dergisi: 34(4): 456-462, 2008
Deng M, Chen G, et al	A Study on In Vitro Degradation Behavior of a Poly(glycolide-co-L-lactide) Monofilament	Acta Biomaterialia,4:1382-91,11-Apr-08
Deng, DY, et al	Presentation and management of major complications of midurethral slings: are complications under-reported?	Neurourol and Urodyna (2007) 26: 46-52
Deprest, et al.	The need for preclinical research on pelvic floor reconstruction	BJOG 2013; 120: 141-143
Descazeaud A, Salet- Lizée, et al	Traitement de l'incontinence urinaire d'effort par bandelette TVT-O : résultats immédiats et à un an	Gynécologie Obstétrique & Fertilité
DeSouza R, Shapiro A, et al	"Adductor brevis myositis following transobturator tape procedure: a case report and review of the literature."	Int Urogyn J 2007;18:817-820

Detollenaere RJ, De Boon J, et al	Short term anatomical results of a randomized controlled non inferiority trial comparing sacrospinous hysteropexy and vaginal hysterectomy in treatment of uterine prolapse stage 2 or higher	Int Urogynecol J (2013) 24 (Suppl 1): S1-S152
Detollenaere, RJ., et al	Short Term Anatomical Results of a Randomized Controlled Non Inferiority Trial Comparing Sacrospinous Hysteropexy and Vaginal Hysterectomy in Treatment of Uterine Prolapse Stage 2 or Higher2	Int Urogynecol J (2013) 24 (Suppl 1): S1-S152
Deval B and Haab F.	Management of the complications of the synthetic slings.	Current Opinion in Urology
Deval B, Birsan A, et al	Objective and Subjective Cure Rates After Tension- free Vaginal Tape for Treatment of Urinary Incontinence.	Urology,58: 702-706 ,01-Nov-01
Di Piazza L, Piroli Torelli D,et al	Complications in short suburethral sling positioning.	Int Urogyn J 2009;20(Suppl 3):S403-404 IUGA Abst 430
Diallo S, Cour F, et al	Evaluating single-incision slings in female stress urinary incontinence: The usefulness of the CONSORT statement criteria.	Urol 2012;80:535-541
Dietz HP, Vancaillie P, et al	Mechanical Properties of Urogynecologic Implant Materials	The International Urogynecology Journal,14:239- 43; discussion 243,05-Aug-03
Dobson A, et al	Trans-obturator surgery for stress urinary incontinence: 1-year follow-up of a cohort of 52 women	Int Urogynecol J (2006) Accepted 6 March 2006
Drahoradova P,	Longitudinal trends with improvement in quality of life after TVT, TVT O and Burch colposuspension procedures.	Med Sci Monit. 2011; 17(2):CR67-72
Drahoradova P, Masata J, et al	Comparative Development of Quality of Life Between TVT and Burch Colposuspension	Proceedings of the 34th Annual Meeting of the International Continence Society and the International UroGynecological Association August 25-27; Paris, France (, 2004) Abstract 278
Drews RC	Polypropylene in the human eye	Am Intra-Ocular Implant Soc J 1983 Spring 9:137-142
Duckett J, Baranowski A	Pain after suburethral sling insertion for urinary stress incontinence	Int Urogynecol J (2013) 24:195-201
Duckett JR, Jain S	Groin pain after a tension-free vaginal tape or similar suburethral sling: management strategies	BJU Int. 2005 Jan;95(1):95-7



DukeMedicine	Alert - Pelvic floor disorders Incontinence and vaginal prolapse	DukeMedicine
Dunn GE, et al	Changed Women The Long-Term Impace of Vaginal Mesh Complications	Female Pelvic Med Reconstr Surg 2014;20: 131- 136
Dunn GE, et al	Changed Women: The Long-Term Impace of Vaginal Mesh Complications	Female Pelvic Med Reconstr Surg 2014;20: 131- 136
Duong TH, Taylor D, et al	A Multicenter Study of Vesicovaginal Fistula Formation Following Cystotomy During Hysterectomy for Benign Indications	Felmale Pelvic Med & Reconstruc Surgery (2010)16(2) Supp S5-S43
Dwyer, PL	Editorial The 75% rule: all stress incontinence procedures are alike	Int Urogynecol J (2011) 22:769-770
Dyrkorn OA, Kulseng- Hanssen et al	TVT compared with TVT-O and TOT: results from the Norwegian National Incontinence Registry.	Int Urogynecol J. 2010 Nov;21(11):1321-6. doi: 10.1007/s00192-010-1195-0. Epub 2010 Jun 18.
EAU	Guidelines on Surgical Treatment of Urinary Incontinence	Published online 6 November 2012
EAU 2009:	Single-incision mid-urethral sling has high cure rate for stress urinary incontinence.	Medscape Mar 31, 2009.
Eland IA, et al	Attitudinal surgey of volunatry reporting of adverse drug reactions	Br J Clin Pharmacol (1999) 48, 623-627
El-Barky E, El-Shazly A, et al	Tension free vaginal tape versus Burch colposuspension for treatment of stress urinary incontinence.	Int Urol Nephrol. 2005;37(2):277-81
El-Hefnawy AS, Wadie BS, et al	TOT for treatment of stress urinary incontinence: how should we assess its equivalence with TVT?	Int Urogynecol J. 2010 Aug;21(8):947-53
Ellington DR and Richter HE.	Indications, Contraindications, and Complications of Mesh in Surgical Treatment of Pelvic Organ Prolapse	Clinical Obstetrics & Gynecology,56:276-88,01- Jun-13
Elser, D	New Options for Stress Urinary Incontinence	The Female Patient, Supp. Aug. 2009
Elzevier, Henk Willem	Female Sexual Function after Surgery for Stress Urinary Incontinence: Transobturator Suburethral Tape vs. Tension-Free Vaginal Tape Obturator	J Sex Med 2008; 5: 400-406
Endo M, et al	Mesh contraction: in vivo documentation of changes in apparent surface area utilizing meshes visible on magnetic resonance imaging in the rabbit abdominal wall model	Int Urogynecol J (2014) 25:737-743
Enzelsberger H, Schalupny J, et al	TVT versus TOT - A prospective randomized study for the treatment of female stress urinary incontinence at a followup of 1 year [German].	Geburtsh Frauenheilk 65:506-511

Ethicon, iNC. 2012	Sacrocolpopexy with ARTISYN™ Y-Shaped Mesh, Sales Process	ETH.MESH.08114911 - ETH.MESH.08114934
Falconer, C., et al	Influence of Different Sling Materials on Connective Tissue Metabolism in Stress Urinary Incontinent Women	Int Urogynecol J (2001) (Suppl 2): S19-S23
Farrell SA,	The evaluation of stress incontinence prior to primary surgery	J Obstet Gynaecol Can 2003;25(4):313-8
FDA	FDA Public Health Notification: Serious Complications Associated with Transvaginal Placement of Surgical Mesh in Repair of Pelvic Organ Prolapse and Stress Urinary Incontinence	20-10-08
FDA	801.63 Medical devices; warning statements for devices containing or manufactured with chlorofluorocarbons and other class I ozone-depleting substances; Subpart D-Exemptions from Adequate Directions for Use	Food and Drug Administration 21 CFR Ch. I (4-1- 14 Edition)
FDA	§801.63 Medical devices; warning statements for devices containing or manufactured with chlorofluorocarbons and other class I ozone-depleting substances	2014 Food and Drug Administration, HHS §801.63
FDA	FDA's Role and Activities	FDA webpage 4/29/2014
FDA	Surgical mesh for treatment of women with pelvic organ prolapse and stress urinary incontinence	FDA Executive Summary 2011
Feifer A and Corcos J.	The use of synthetic sub-urethral slings in the treatment of female stress urinary incontinence.	Int Urogynecol J (2007) 18:1087-1095
Feiner B and Maher C.	Vaginal mesh contraction: definition, clinical presentation, and management.	Obstetrics & Gynecology, 115:325-30 ,01-Feb-10
Feng CL, Chin HY, Wang KH.	Transobturator vaginal tape inside out procedure for stress urinary incontinence: results of 102 patients.	Int Urogynecol J Pelvic Floor Dysfunct. 2008 Oct;19(10):1423-7. doi: 10.1007/s00192-008- 0658-z. Epub 2008 May 31.
Feola A, Moalli PA, et al	Stress-Shielding the impact of Mesh Stiffness on Vaginal Function	Female Pelvic Med Reconstr. Surgery (2011) 17(5): S54-S110
Feola, A.	Deterioration in biomechanical properties of the vagina following implantation of a high-stiffness prolapse mesh	BJOG An International Journal of Obstetrics and Gynaecology 2013
Firoozi F and Goldman HB.	Pure transvaginal excision of mesh erosion involving the bladder	The International Urogynecology Journal,,04-Apr- 13



Fischer A, Fink T, et al	Comparison of retropubic and outside-in transoburator sling systems for the cure of female genuine stress urinary incontinence	European Urology,48:799-804,15-Aug-05
Flam F, Boijssen M, and Lind F.	Necrotizing fasciitis following transobturator tape treated by extensive surgery and hyperbaric oxygen.	Int Urogynecol J (2009) 20:113-115
Fletcher AP	Spontaneous adverse drug reaction reporting vs event monitoring: a comparison	J Royal Society Med (1991) 84: 341-344
Flock F, Reich A, et al	Hemorrhagic complications associated with tension-free vaginal tape procedure.	Obstetrics & Gynecology,104:989-94,01-Nov-04
Flood CG, et al	Anterior colporrhaphy reinforced with Marlex mesh for the treatment of cystoceles	Int Urogynecol J Pelvic Floor Dysfunct 1998; 9(4): 200-4
Food and Drug Administration (FDA).	FDA executive summary. Surgical mesh for the treatment of women with pelvic organ prolapsed and stress urinary incontinence.	Obstetrics and Gynecology Devices Advisory Committee Meeting,, 9/8/2011
Food and Drug Administration (FDA).	FDA Safety Communication: Update on Serious Complications Associated with Transvaginal Placement of Surgical Mesh for Pelvic Organ Prolapse.	13-07-11
Foote J, et al	Referral Patterns and Complications of Midurethral Slings	Abstract 843
Ford, AA, Rogerson L, et al	Mid-urethral sling operations for stress urinary incontinence in women (Review)	The Cochran Collaboration, The Cochran Library 2015, Issue 7
Freeman R, Holmes D, et al	What patients think: patient-reported outcomes of retropubic versus trans-obturator mid-urethral slings for urodynamic stress incontinence - a multi- centre randomised controlled trial	Int Urogynecol J. 2011 Mar;22(3):279-86
Frenkl TL, Rackley RR, et al	Management of Iatrogenic Foreign Bodies of the Bladder and Urethra Following Pelvic Floor Surgery	Neuro and Urodynam (2008)27:491-495
Funk MJ, Siddiqui NY, et al	Long-term Outcomes After Stress Urinary Incontinence Surgery	Obstetrics & Gynecology,120:83-90,01-Jul-12
Gamble T, et al	Predicting persistent dstrusor overactivity after sling procedures	Int Urogynecol J (2008) 19 (Suppl 1): S64-S65
GAO	Medical Devices FDA Should Take Steps to Ensure That High-Risk Device Types Are Approved through the Most Stringent Premarket Review Process	GAO Report to Congressional Addressees January 2009
Garber, AM	Modernizing Device Regulation	N Engl J Med (2010) 362:13: 1161-1163

Garcia-Urena, M.A., et al	2007. Differences in polypropylene shrinkage depending on mesh position in an experimental study.	Am J Surg, 193(4), 538-542. doi: 10.1016/j.amjsurg.2006.06.045
Gebhart JB, Dixon DA, et al	Three-year outcomes of Uretex Urethral Support System for treatment of stress urinary incontinence	The International Urogynecology Journal,19:1075-9,28-Feb-08
Geoffrion R, et al	Closing the Chapter on Obtape: a case report of delayed thigh abscess and a literature reivew	J Obstet Gynaecol Can 2008; 30(2): 143-147
Gerstenbluth, RE., et al	Simultaneous Urethral Erosion of Tension-Free Vaginal Tape and Woven Polyester Pubovaginal Sling	J Urol. 2003, Aug; 170 (2 Pt 1): 525-6
Gilberti C, et al	Transobturator Tape for Treatment of Female Stress Urinary Incontinence: Objective and Subjective Results After a Mean Follow-up of Two Years.	Urology 2007;69:703-707
Gomelsky A, et al	Biocompatibility Assessment of synthetic sling materials for female stress urinary incontinence	J Uro (2007)178:1171-1181
Greca FH	The influence of differing pore sizes on the biocompatibility of two polypropylene meshes in the repair of abdominal defects. Experimental study in dogs.	Hernia. (2001); 5: 59-64.
Green J, Buencamino D, et al	A comparison of the transobturator tape and transabdominal tension free vaginal tape procedures for the surgical treatment of stress urinary incontinence	Presented at the 35th Annual Meeting of the International Continence Society, 28th August- 2nd September 2005, Montreal, Canada.
Greenwald D, Shumway S, et al	Mechanical Comparison of 10 Suture Materials Before and After in Vivo Incubation	J Surg Research 1994; 56:372-377
Gristina, A.G.	(1987). Biomaterial-centered infection: microbial adhesion versus tissue integration.	Science, 237:1588-1595.
Groutz A, Cohen A, et al	The safety and efficacy of the ""inside-out"" trans- obturator TVT in elderly versus younger stress- incontinent women: a prospective study of 353 consecutive patients.	Neurourol Urodyn. 2011 Mar;30(3):380-3. doi: 10.1002/nau.20976.
Groutz A, Levin I, Gold R, et al	""Inside-out"" transobturator tension-free vaginal tape for management of occult stress urinary incontinence in women undergoing pelvic organ prolapse repair."	Urology. 2010 Dec;76(6):1358-61. doi: 10.1016/j.urology.2010.04.070. Epub 2010 Oct 25.

Groutz A, Rosen G, Gold R, et al	Long-term outcome of transobturator tension-free vaginal tape: efficacy and risk factors for surgical failure.	J Womens Health (Larchmt). 2011 Oct;20(10):1525-8. doi: 10.1089/jwh.2011.2854. Epub 2011 Aug 5.
Guerrero KL, Emery SJ, et al	"A randomised controlled trial comparing TVT, Pelvicol and autologous fascial slings for the treatment of stress urinary incontinence in women "	BJOG 2010 Nov;117(12):1493-502.
Guidoin R, Chakfe N	Aneurysmal Deterioration of Arterial Substitutes	Current Therapy in Vascular Surgery 2: 324-328
Haferkamp, A., et al	Urethral Erosion of Tension-Free Vaginal Tape	J Urol. 2002;167(1): 250
Hammad FT, et al	Erosions and Urinary Retention Following Polypropylene Synthetic Sling: Australasian Surgery	Eur Urol 47 (2005) 641-647
Hammett J, et al	Short-term surgical outcomes and characteristics of patients with mesh complications from pelvic organ prolapse and stress urinary incontinence surgery <sup>2</sup>	Int Urogynecol J (2014) 25:465-470
Han WHC.	Burch colposuspension or tension-free vaginal tape for female urinary incontinence? (Abstract).	International Urogynecology Journal and Pelvic Floor Dysfunction
Handa VL, et al	Banked human fascia lata for the suburethral sling procedure: a preliminary report	Obstet Gynecol 88(6): 1045-9 1996 Dec
Hansen BL, et al	Long-Term Follow-up of Treatment for Synthetic Mesh Complications	Female Pelvic Med Reconstr Surg 2014;20: 126- 130
Harding CK, et al	A prospective study of surgeon and patient-perceived outcome following transobturator tape insertion for treatment of urodynamic stress incontinence	Br J Med and Surg Urol (2009) 2, 197-201
Hartung DM, et al	Reporting Discrepancies between clinicaltrials.gov results database and peer-reviewed publications	Ann Intern Med 2014;160:477-483
Hassan MF, et al	Treatment success of transobturator tape compared with tension free vaginal tape for stress urinary incontinence at 24 months: a randomized controlled trial	Open J Obstet and Gynecol, 2014, 4, 169-175

Haylen, BT, et al	An International Urogynecological Association (IUGA)/International Continence Society (ICS) joint terminology and classification of the complications related directly to the insertion of prostheses (meshes, implants, tapes) and grafts in female pelvic fl	Neurourology and Urodynamics,30:2-12,01-Jan- 11
Hazell L, et al	Under-reporting of adverse drug reactions a systematic review	Drug Safety 2006; 29(5) 385-396
Hazewinkel MH, Hinoul P, et al	Persistent groin pain following a trans-obturator sling procedure for stress urinary incontinence: a diagnostic and therapeutic challenge	The International Urogynecology Journal,20:363- 5,04-Sep-08
Health Canada	Health Canada Warns of Serious, Life-Altering Complications Associated with Transvaginal Mesh _ NewsInferno	<a href="http://www.newsinferno.com/health-canada-warns">http://www.newsinferno.com/health-canada-warns</a>
Heinonen, P.	Tension-free vaginal tape procedure without preoperative urodynamic examination: Long-term outcome	International Journal of Urology (2012) 19, 1003-1009
Heise, CP, et al	Mesh Inguinodynia: A New Clinical Syndrome After Inguinal Herniorrhaphy?	J Am Coll Surg. 1998 Nov; 187(5): 514-8
Heniford, B.T.	""The benefits of lightweight meshes in Ventral Hernia Repair in Ventral Hernia Repair""	Video produced by Ethicon. 2007
Hernández-Gascón B, et al	Mechanical behaviour of synthetic surgical meshes: Finite element simulation of the herniated abdominal wall	Acta Biomaterialia 7 (2011) 3905-3913
Hilton P	Long-term follow-up studies in pelvic floor dysfunction: the Holy Grail or a realistic aim?	BJOG 2008; 115: 135-143
Hilton, P., et al	Postural Perineal Pain Associated With Perforation of the Lower Urinary Tract Due to Insertion of a Tension-Free Vaginal Tape	BJOG (2003) 110: 79-82
Hines, Jonas Zajac, et al.	Left to Their Own Devices: Breakdowns in United States Medical Device Premarket Review	PLOS Med 7(7): e1000280 July 13, 2010
Hinoul P, Bonnet P, Krofta L,et al	An anatomic comparison of the original versus a modified inside-out transobturator procedure	Int.Urogynecol J(2011) 22(8) 997-1004
Hinoul P, et al	Anatomical variability in the trajectory of the inside-out transobturator vaginal tape technique (TVT-O)	Int Urogynecol J (2007) 18:1201-1206

Hinoul P, et al	A Randomized, Controlled Trial Comparing an Innovative Single Incision Sling With an Established Transobturator Sling to Treat Female Stress Urinary Incontinence	J Urology 2011 (185): 1356-1362
Hinoul P, et al	TVT obturator system versus TVT secur: arandomized controlled trial, short term results	Int Urogynecol J (2009) 20:S213
Hinoul P, Roovers JP, et al	Surgical management of urinary stress incontinence in women: A historical and clinical overview.	Eur J Obstet Gyn Reprod Biol 2009;145:219-225
Hogston P	Single surgeon experience with 125 trans-obturator sling procedures	Int Urogynecol J (2011) 22 (Suppl 3) S1947
Hokenstad ED, et al	Health-related quality of life and outcomes after surgical treatment of complications from vaginally placed mesh	Female Pelvic Med Reconstr Surg 2015;21: 176- 180
Holmgren, C	Quality of life after tension-free vaginal tape surgery for female stress incontinence	Scandinavian Journal of Urology and Nephrology, 2006; 40: 131-137
Hong MK, Liao CY, et al	Internal Pudendal Artery Injury during Prolapse Surgery Using Nonanchored Meesh	J Min Invas Gynecol (2011) 18, 678-681
Hota LS, Hanaway K, et al	TVT-Secur (Hammock) Versus TVT-Obturator: A Randomized Trial of Suburethral Sling Operative Procedures.	Female Pelvic Reconstructive Surgery,18:41- 45,01-Jan-12
Hou JC, et al	Outcome of Transvaginal Mesh and Tape Removed for Pain Only	The Journal of Urology Vol 192, 856-860, September 2014
Houwert RM, Renes-Zijl C, et al	TVT-O versus Monarc after a 2-4-year follow-up: a prospective comparative study.	Int Urogynecol J Pelvic Floor Dysfunct. 2009 Nov;20(11):1327-33. doi: 10.1007/s00192-009- 0943-5. Epub 2009 Jul 14.
Howden NS, Zyczynski HM, et al	Comparison of autologous rectus fascia and cadaveric fascia in pubovaginal sling continence outcomes	The American Journal of Obstetrics & Gynecology,194:1444-9,31-Mar-06
Hsiao SM, et al	Sequential comparisons of postoperative urodynamic changes between retropubic and transobturator midurethral tape procedures	World J Urol 2008 Dec;26(6):643-8
Hubka P, Masata J, Nanka O, et al	Possible complications of the TVT-S vaginal tape in the H-position.	ICS Abstract #286,,00-Jan-00
Hubka P, Nanka O, Martan A, et al	Anatomical study of position of the TVT-O to the obturator nerve influenced by the position of the legs during the procedure: based upon findings at formalin-embalmed and fresh-frozen bodies.	Archives of Gynecology and Obstetrics

Hurtado EA, Appell RA	Management of complications arising from transvaginal mesh kit procedures: a tertiary referral center's experience	Int Urogynecol J (2009) 20:11-17
Iakovlev V	Explanted surgical meshes: what pathologists and industry failed to do for 50 years	Virchows Arch (2014) 465 (Suppl 1):S1-S379
Iakovlev V, et al	Pathological Findings of Transvaginal Polypropylene Slings Explanted for Late Complications Mesh Is Not Inert	ICS Abstract 228, 2014
Iakovlev V, Guelcher S, et al	In vivo degradation of surgical polypropylene meshes: A finding overlooked for decades	Virchows Arch (2014) 465 (Suppl 1): S35
Iakovlev V, Mekel G, Blaivas J	Pathological Findings of Transvaginal Polypropylene Slings Explanted for Late Complications: Mesh Is Not Inert	ICS.org abst 228 Study St Michael's Hospital, Univ. Toronto
Iakovlev VV, Carey ET, Steege J	Pathology of Explanted Transvaginal Meshes2	Accepted Abstract
Iakovlev VV, et al	Pathology of Explanted Transvaginal Meshes	Int J Med Health Pharmaceu and Biomedical Engineering 2014, (8)9: 510-513
Iakovlev, V.,	PS-24-006 Explanted surgical meshes: what pathologists and industry failed to do for 50 years	Virchows Arch (2014) 465 (Suppl 1): S337
Iakovlev, V., et al	OFP-13-001 In vivo degradation of surgical polypropylene meshes: A finding overlooked for decades	Virchows Arch (2014) 465 (Suppl 1): S35
ICS	ICS Fact Sheets A Background to Urinary and Faecal Incontinence Prepared by the Publications & Communications Committee	13-Jul
Iglesia CB	Pelvic Organ Prolapse Surgery: Long-term Outcomes and Implications for Shared Decision Making	The Journal of American Medical Association,309:2045-6,15-May-13
Iglesia CB, et al	The use of mesh in gynecologic surgery	Int Urogynecol J (1997)8:105-115
Iglesia CB, Sokol AI, et al	Vaginal mesh for prolapsed: a randomized controlled trial.	Obstetrics & Gynecology,116:293-303,01-Aug-10
Iglesia, CB	Stop Using Synthetic Mesh for Routine Repair of Pelvic Organ ProlapseStart Performing Native Tissue Repairs and Reserve Mesh for Selective Cases	OBG Management 2013; 25: 24-25
IUGA	Stress Urinary Incontinence A Guide for Women	IUGA 2011

Jaburek L, Jaburkova J, et al	Risk of haemorrhagic complications of retropubic surgery in females: anatomic remarks	Biomedical Papers of the Medical Faculty of the University Palacky, Olomouc, Czech Republic,155:75-7,01-Mar-11
Jacquetin, B., et al	Complications of Vaginal Mesh: Our Experience	Int Urogyn J 2009; 20:893-6
Jelovsek JE, Barber MD, et al	Randomised trial of laparoscopic Burch colposuspension versus tension-free vaginal tape: long-term follow up	BJOG,115:219-25; discussion 225,01-Jan-08
Jha, Swati	Impact of Incontinence Surgery on Sexual Function: A Systematic Review and Meta-Analysis	J Sex Med 2012; 9: 34-43
Jones, Keisha, Beola A, et al	Tensile Properties of Commonly Used Prolapse Meshes	Int Urogynecol J Pelvic Floor Dysfunct (2009); 20(7); 847-853
Jongebloed, WL, et al.	Degradation of Polypropylene in the Human Eye: A Sem-Study	Doc Ophthalmol, Vol 64, No. 1, pp. 143-152, 1986
Jonsson Funk M, Siddiqui NY, et al	Sling revision/removal for mesh erosion and urinary retention: long-term risk and predictors	Am J Obstet Gynecol 2013; 208:73.e1-7
Juang CM, Yu KJ, Chou P, et al	Efficacy analysis of trans-obturator tension-free vaginal tape (TVT-O) plus modified Ingelman- Sundberg procedure versus TVT-O alone in the treatment of mixed urinary incontinence: a randomized study	European Urology,51:1671-8,16-Jan-07
Julia JJ, et al	Long term experinece in 72 patients with the Advantage® sling system	www.bostonscientific.com/gynecology
Julian TM.	The efficacy of Marlex mesh in the repair of severe, recurrent vaginal prolapse of the anterior midvaginal wall.	The American Journal of Obstetrics and Gynecology,175:1472-5,01-Dec-96
Juma S and Brito CG.	Transobturator tape (TOT): Two years follow-up	Neurourol.Urodynam. (2007) 26:37-41
Jung HC, Kim JY, Lim HS, et al.	Three-year Outcomes of the IRIS procedure for Treatment of Female Stress Urinary Incontinence: Comparison with TVT Procedure.	J Korean Med Sci. (2007); 22: 497-501.
Junge, K., Rosch, R. Klinge, U. et a	Risk factors related to recurrence in inguinal hernia repair: a retrospective analysis.	Hernia (2006)10: 309-315.
Kaelin-Gambirasio I, Jacob S, et al	Complications associated with transobturator sling procedures: analysis of 233 consecutive cases with a 27 months follow-up.	BMC Womens Health. 2009 Sep 25;9:28. doi: 10.1186/1472-6874-9-28.



Kane AR, et al	Midurethral Slings for Stress Urinary Incontinence	Clinical Obstetrics and Gynecol (2008) 51(1): 124-134
Kang D., et al	MP75-16 Patient Quality of Life After Removal of Vaginal Mesh	Urology J Vol 191, No. 4S e879
Karateke A, Haliloglu B, Cam C, et al	Comparison of TVT and TVT-O in Patients with Stress Urinary Incontinence: Short-term Cure Rates and Factors Influencing the Outcome. A Prospective Randomised Study.	Australian and New Zealand Journal of Obstetrics and Gynaecology 2009; 49:99-105
Karram MM, Segal JL, et al	Complications and untoward effects of the tension-free vaginal tape procedure.	Obstetrics and Gynecology, 101:929-32, 01-May-03
Karsenty G, et al	Severe soft tissue infection of the thigh after vaginal erosion of transobturator tape for stress urinary incontinence	Int Urogynecol J 2007;18:207-221
Katz, S.	Bacterial adherence to surgical sutures: a possible factor in suture induced infections.	Ann Surg, 194:35-41
Kavvadias T, Klinge U, Schuessler	"Ch. 56 Alloplastic Implants for the Treatment of Stress Urinary Incontinence"	Hernia Repair Sequelae, editors, V. Schumpelick, RJ Fitzgibbons, Springer-Verlag publ, 2010, pp. 439-444
Kennelly MJ, et al	Prospective evaluation of a single incision sling for stress urinary incontinence	J Urology (2010) 184: 604-609
Kessler, DA	Introducing MEDWatch A New Approach to Reporting Medication and Device Adverse Effects and Product Problems	JAMA (1993) 269 (21): 2765-2768
Khandwala S, Lucent V, et al	Preliminary Results of Peri-Operative and 3-Month Outcomes From a World-Wide Observational Registry of Tension-Free Vaginal Tapes in Women with Stress Urinary Incontinence	ICS Abstract 493
Khanuengkitkong S, et al	Delayed vaginal and urethral mesh exposure: 10 years after TVT surgery	Int Urogynecol J (2013) 24: 519-521
Kim J, Na Y, Lee J, Seo J. et al.	Comparative Study of Tension-free Vaginal Tape (TVT) and Suprapubic arc (SPARC) Sling Procedure for Female Stress Urinary Incontinence.	V. Schumpelick, RJ Fitzgibbons, Springer-Verlag publ, 2010, pp. 439-444
Kim JY, Jung HC, Moon KH, et al	Comparisons of IRIS, TVT and SPARC Procedure for Stress Urinary Incontinence.	European Urology Supplement
Kjollhede, Preben	Prognostic factors and long-term results of the Burch colposuspension	Acta Obstet Gynecol Scand 1994; 73: 642-64
Kjollhede, Preben	Long-term efficacy of Burch colposuspension: a 14 year follow-up study	Acta Obstet Gynecol Scand 2005; 84:767-772



Klein-Patel M, et al	Ultra-lightweight synthetic mesh has similar cellular response but increased tissue ingrowth relative to heavier weight prototype	Female Pelvic Med Reconstr. Surgery (2011) 17(5): S54-S110
Klinge U, et al	High structural stability of textile implants prevents pore collapse and preserves effective porosity at strain	BioMed Research Internal (2015) Art ID 953209; 1-7
Klinge U, Klosterhalfen B, et al	Foreign Body Reaction to Meshes Used for the Repair of Abdominal Wall Hernias	Eur J Surg 1999; 165: 665-673
Klinge, U., et al	Do Multifilament Alloplastic Meshes Increase the Infection Rate? Analysis of the Polymeric Surface, the Bacteria Adherence, and the In Vivo Consequences in a Rat Model	J Biomed Mater Res 2002; 63:765-771
Klinge, U., et al	Shrinking of Polypropylene Mesh In Vivo: An Experimental Study in Dogs	Eur J Surg. 1998; 164; 965-969
Klosterhalfen B and Klinge U.	The lightweight and large porous mesh concept for hernia repair	Expert Rev. Med. Devices, 2005; 2(1)
Kobashi KC	Management of Erosion of Graft Materials in Pelvic Floor Reconstruction	The Scientific World J (2009) 9, 32-26
Kobashi, K, et al	Management of Vaginal Erosion of Polypropylene Mesh Slings	The Journal of Urology, Vol. 169, 2242-2243 (June 2003)
Kobashi, K, et al	Erosion of Woven Polyester Pubovaginal Sling	The Journal of Urology, Vol. 162, 2070-72 (Dec 1999)
Kociszewski J	Tape Functionality: Sonographic Tape Characteristics and Outcome After TVT Incontinence Surgery	Neurourol Urodynam. 27.6 (2008): 485-490
Koelbl H, Halaska M, et al	Burch Colposuspension and TVT: Perioperative Results of a Prospective Randomized Trial in Patients with Genuine Stress Incontinence.	Neurourol Urodyn 22 (2003): 327.
Kondo A, Isobe Y, Kimura K, et al	Efficacy, Safety and Hospital Costs of Tension-free Vaginal Tape and Pubovaginal Sling in the Surgical Treatment of Stress Incontinence.	J Obstet Gynaecol Res. 2006 Dec;32(6):539-44.
Koops, S, et al	What determines a successful tension-free vaginal tape? A prospective multicenter cohort study: Results from The Netherlands TVT database	Am J Obstet Gynecol (2006) 194. 65-74
Krause H, et al	Biomechanical properties of raw meshes used in pelvic floor reconstruction	Int Urogynecol J (2008) 19: 1677-1681

Krauth JS, Rasoamiaramanana H,et al	Sub-urethral tape treatment of female urinary incontinence--morbidity assessment of the trans- obturator route and a new tape (I-STOP): a multi- centre experiment involving 604 cases	European Urology
Krofta L, et al	TVT and TVT-O for Surgical Treatment of Primary Stress Urinary Incontinence Prospective Randomized Trial.	Int Urogynecol J 2010;21:141-148
Krofta L, Feyereisl J, et al	"TVT-S for Surgical Treatment of SUI: Prospective Trial, 1-Year follow-up"	Int Urogynecol J (2010)21:779-85,
Krofta L, Feyereisl J, et al	TVT and TVT-O for Surgical Treatment of Primary Stress Urinary Incontinence: Prospective Randomized Trial.	Int Urogynecol J.(2010) 21:141-8
Kuhn, A.	Sexual function after suburethral sling removal for dyspareunia	Surg Endosc (2009) 23:765-768
Kuo HC.	Comparison of video urodynamic results after the pubovaginal sling procedure using rectus fascia and polypropylene mesh for stress urinary incontinence	The Journal of Urology,65:163-8,01-Jan-01
Kuuva N and Nilsson CG.	A nationwide analysis of complications associated with the tension-free vaginal tape (TVT) procedure	Acta Obstetricia et Gynecologica Scandinavica ,81:72-7,01-Jan-02
Kuuva, N, et al.	Long-term results of the tension-free vaginal tape operation in an unselected group of 129 stress incontinent women	Acta Obstetricia Gynecologica Scandanavica 2006(85: 4 482-87)
Kwon SY, Latchamsetty KC, et al	Inflammatory Myofibroblastic Tumor of the Urinary Tract Following a TVT	Female Pelvic Medicine & Reconstructive Surgery,18:249-51,01-Jul-12
Labrie J, van der Graaf Y,et al	Protocol for physiotherapy OR TVT Randomised efficacy trial (PORTRET): a multicenter randomized controlled trial to assess the cost- effectiveness of the tension free vaginal tape versus pelvic floor muscle training in women with symptomatic moderate to s	BMC Women's Health 2009;9:24-32
Lapitan MC, Cody JD, and et al	Open retropubic colposuspension for urinary incontinence in women.	Cochrane Database of Systematic Reviews ,CD002912,21-Jan-09
Latthe PM, Foon R, and Tooze-Hobso	Transobturator and retropubic tape procedures in stress urinary incontinence: a systematic review and meta-analysis of effectiveness and complications	BJOG 2007; 114:522-531

Latthe PM, Singh P, Foon R, et al	Two routes of transobturator tape procedures in stress urinary incontinence: a meta-analysis with direct and indirect comparison of randomized trials.	BJUI 2009; 106:68-76
Laurikainen E, et al	Retropubic TVT Compared with Transobturator TVT (TVT-O) in Treatment of Stress Urinary Incontinence Five-year Results of a Randomized Trial.	Öbstet Gynecol 2007;109:4-11
Laurikainen E, Valpas A, et al	Retropubic compared with transobturator tape placement in treatment of urinary incontinence: a randomized controlled trial	Obstetrics & Gynecology
Lee D, Bacsu C, et al	Meshology: a fast-growing field involving mesh and/or tape removal procedures and their outcomes	Expert Rev. Med Devices Early online 1-16 (2014)
Lee D, Dillon B, Lemack G, et al	"Transvaginal Mesh Kits-- How ""Serious"" Are the Complications and Are They Reversible?"	Urology 81: 43-49, 2013
Lee KS, Choo MS, Lee YS, et al	Prospective comparison of the 'inside-out' and 'outside-in' transobturator-tape procedures for the treatment of femals stress urinary incontinence	Int Urogynecol J. (2008) 19:577-582.
Lee KS, Han DH, Choi YS, et al	A prospective trial comparing tension-free vaginal tape and transobturator vaginal tape inside-out for the surgical treatment of female stress urinary incontinence: 1-year follow up.	J Urology (2007) 177: 214-218
Levin I, Groutz A, Gold R, et al	Surgical Complications and Medium-term Outcome Results of Tension-free Vaginal Tape: A Prospective Study of 313 Consecutive Patients.	Neurology and Urodynamics ,23:7-9,01-Jan-04
Levy, B., et al	Best Options, Techniques, and Coding Tips for Pelvic Prolapse Repair	OBG Management Supp.Sept. 2007; S1-S12
Li B, Zhu L, Lang J, Fan R,et al	Long-term Outcomes of the Tension-Free Vaginal Tape Procedure for Female Stress Urinary Incontinence: 7-Year Follow-up in China	J Min Invas Gynecol 2012; 19(2):201-205
Li X, et al.	Characterizing the ex vivo mechanical properties of synthetic polypropylene surgical mesh	J Mechanical Behavior Biomedical Materials (2014) 48-55
Liang R, Abramowitch S, et al	Vaginal Degeneration Following Implantation of Synthetic Mesh With Increased Stiffness	BJOG (2013); 120(2): 233-243

Liang R, et al	Increasing stiffness of synthetic mesh negatively impacts vaginal connective tissue	Female Pelvic Med Reconstr. Surgery (2011) 17(5): S54-S110
Liang R, et al	Vaginal degeneration following implantation of high stiffness mesh	Female Pelvic Med Reconstr. Surgery (2011) 17(5): S54-S110
Liang R, et al	Impact of prolapse meshes on the metabolism of vaginal extracellular matrix in rhesus macaque	Am J Obstet Gynecol 2015;212:174.e1-7.
Liapis A, Bakas P, and Creatsas G.	Monarc vs. TVT-O for the treatment of stress incontinence: a randomized study	Int Urogynecol J (2008) 19:185-190
Liapis A, Bakas P, Creatsas G.	Efficacy of inside-out transobturator vaginal tape (TVTO) at 4 years follow up.	Eur J Obstet Gynecol Reprod Biol. 148 (2010) 199-201.
Liapis A, Bakas P, Creatsas G.	Burch Colposuspension and Tension-free Vaginal Tape in the Management of Stress Urinary Incontinence in Women.	Eur Urol. 41 (2002):469-473.
Liapis A, Bakas P, et al	Tension-free Vaginal Tape for Elderly Women with Stress Urinary Incontinence.	International Journal of Gynecology & Obstetrics (2006) 92, 48-51
Liapis A, Bakas P, et al	The use of oestradiol therapy in postmenopausal women after TVT-O anti-incontinence surgery.	Maturitas. 2010 May;66(1):101-6. doi: 10.1016/j.maturitas.2010.01.020. Epub 2010 Mar 16.
Liapis A, Bakas P, Giner M, et al	Tension-free vaginal tape versus tension-free vaginal tape obturator in women with stress urinary incontinence	Gynecol Obstetric Invest 62:160-164
Liebert T., et al	Subcutaneous Implants of Polypropylene Filaments	J Biomed Mater Res. 1976; 10:939-951
Lim JL, Cornish A, et al	Clinical and quality-of-life outcomes in women treated by the TVT-O procedure.	BJOG. 2006 Nov;113(11):1315-20.
Lim JL, Quinlan DJ.	Safety of a new transobturator suburethral synthetic sling (TVT-O) procedure during the training phase.	J Obstet Gynaecol Can. 2006 Mar;28(3):214-7.
Lim YN, et al	Do the Advantage slings work as well as the tension-free vaginal tapes?	Int Urogynecol J (2010) 21:1157-1162
Lim YN, Muller R, et al	"Suburethral slingplasty evaluation study in North Queensland, Australia: the SUSPEND trial."	Australian and New Zealand Journal of Obstetrics and Gynaecology, 45:52-9, 24-Feb-05
Lin ATL, Wang SJ, et al	In vivo tension sustained by fascial sling in pubovaginal sling surgery for female stress urinary incontinence	J Urology (2005) 173: 894-897
Litwiller JP, et al	Effect of lithotomy positions on strain of the obturator and lateral femoral cutaneous nerves	Clin Anat 2004 Jan; 17(1): 45-9

Litwiller, SE et al	Presentation Number: Poster 109 Long term efficacy and safety of the Obtryx™ (Boston Scientific Corp.) sling for treatment of stress urinary incontinence in a community setting: an analysis of outcomes and quality of life	J Pelvic Med & Surg (2009) 12(5) 353
Liu PE, Su CH, Lau HH, et al	Outcome of tension-free obturator tape procedures in obese and overweight women.	Int Urogynecol J. 2011 Mar;22(3):259-63. doi: 10.1007/s00192-010-1311-1. Epub 2010 Nov 12.
Lord HE, Taylor JD, et al	A randomized controlled equivalence trial of short- term complications and efficacy of tension-free vaginal tape and suprapubic urethral support sling for treating stress incontinence.	BJU International,98:367-76 (2006)
Lowry, F	EAU 2009 Single-incision mid-urethral sling has high cure rate for stress urinary incontinence	Medscape. Mar 31, 2009
Lucas M, Emery S, Alan W, et al	Failure of porcine xenograft sling in a randomised controlled trial of three sling materials in surgery for stress incontinence (Abstract).	Proceedings of the International Continence Society (34th Annual Meeting) and the International Urogynecological Association. 2004.
Lucas, M, Bosch, R, et al.	EAU Guidelines on Surgical Treatment of Urinary Incontinence	European Urology 62 (2012) 1118-1129
Maher C, Qatawneh A, et al	Laparoscopic Colposuspension or Tension-free Vaginal Tape for Recurrent Stress Urinary Incontinence and/or Intrinsic Sphincter Deficiency in a Randomized Controlled Trial.	Neurourol Urodyn 2004; 23:433-434
Mahmoud, W.M., Vieth, R.F., et al.	Migration of bacteria along synthetic polymeric fibers.	J Biomater Sci Polym Ed, 4(6):567-578
Mansoor A, Védrine N et al	Surgery of female urinary incontinence using transobturator tape (TOT): a prospective randomised comparative study with TVT (Abstract).	Neurourology and Urodynamics 22.5 (2003): 88-88.
Marcus-Braun N, Bourret A, et al	Persistent pelvic pain following transvaginal mesh surgery: a cause for mesh removal	Euro J Obstet & Gynecol and Reprod Bio 162 (2012) 224-228
Marcus-Braun N, von Theobald P	Mesh removal following transvaginal mesh placement: a case series of 104 operations.	Int Urogynecol J (2010) 21:423-430
Marks B, Goldman H	Controversies in the Management of Mesh-Based Complications: A Urology Perspective	Urol Clin N Am 39 (2012) 419-428
Marsh F and Rogerson L.	Groin abscess secondary to trans obturator tape erosion: case report and literature review.	Neurol Urodynam 26:543-546, 2007

Martinez-Fornes	A three year follow-up of a prospective open randomized trial to compare tension-free vaginal tape with Burch colposuspension for treatment of female stress urinary incontinence	Actas Urol Esp. 2009 Nov;[Deleted Object](10):1088-96
Mary, Celine, et al	Comparison of the In Vivo Behavior of the Polyvinylidene Floride and Polypropylene Sutures Used In Vascular Surgery	ASAIO Journal 1998: 199-206
Masata J, Svabik K, et al	Randomized Prospective Trial of a Comparison of the Efficacy of TVT-O and TVT SECUR System in the Treatment of Stress Urinary Incontinent Women – Comparison of the Long- and Short- Term Results	Neurourol Urodyn 2011;30:805-806
Masata J, Svabik K, Hubka P, et al	ICS Abs 6 Is the fixation of single incision tape (TVT-S) as good as a transobturator tape (TVT-O)? An ultrasound study,results from randomized trial.	International Continence Society Mtg 2012
Masata J, Svabik K, Zvara K, et al	Randomized trial of a comparison of the efficacy of TVT-O and single-incision tape TVT SECUR systems in the treatment of stress urinary incontinent women--2-year follow-up.	Int Urogynecol J. 2012 Oct;23(10):1403-12.
May, J, et al	M429 OUTCOME OF OBTRYX	Poster presentations / Inte J Gynecol & Obstet 11953 (2012) S531-S867
Mazouni, Chafika	Urinary complications and sexual function after the tension-free vaginal tape procedure	Acta Obstetricia et Gynecologica Scandinavica volume 83 Issue 10 pages 955-961 October 2004
McCracken, G.R.	Five Year Follow-Up Comparing Tension-Free Vaginal Tape and Colposuspension	Ulster Med J 2007; 76(3) 146-149
McNanley AR, et al	Poster Presentation Abstracts 108 Recovery after robotic sacrocolpopexy: the patient's perspective	J Pelvic Med & Surg (2009) 12(5) 353
Meier, B	Group Faults the FDA on Oversight of Devices	New York Times 04/12/2011
Mellier G, Benayed B, et al	Suburethral tape via the obturator route: is the TOT a simplification of the TVT?	Int Urogynecol J (2004) 15: 227-232
Merritt, K.	Factors Influencing Bacterial Adherence to Biomaterials	J Biomater Appl. 1991;5:185-203
Meschia M, Bertozzi R, et al	Peri-operative morbidity and early results of a randomised trial comparing TVT and TVT-O	Int Urogynecol J (2007) 18: 1257-1261



Meschia M, et al	TVT-secur: a minimally invasive procedure for the treatment of primary stress urinary incontinence. One year data from a multi-centre prospective trial	Int Urogynecol J (2009) 20:313-317
Meschia M, et al	Tension-Free Vaginal Tape: Analysis of Outcomes and Complications in 404 Stress Incontinent Women	Int Urogynecol J (2001) (Suppl 2): S24-S27
Meschia M, et al	Single-incision mid-urethral sling: impact of obesity on outcome.	Eur J Obstet Gynecol Reprod Biol 2013;170:571- 574
Meschia M, Pifarotti P, et al	Tension-free Vaginal Tape (TVT) and Intravaginal Slingplasty (IVS) for Stress Urinary Incontinence: a Multicenter Randomized Trial.	Am J Obstet Gynecol. 2006 Nov;195(5):1338-42. Epub 2006 Jun 12
Meschia M, Pifarotti P, et al	A randomized comparison of tension-free vaginal tape and endopelvic fascia plication in women with genital prolapse and occult stress urinary incontinence	Am J Obstet Gynecol. 2004 Mar;190(3):609-13.
Meschia M, Pifarotti P, et al	A Multicenter Randomized Comparison of Tension free Vaginal Tape (TVT) and Trans-obturator in- out Technique (TVT-O) for the Treatment of Stress Urinary Incontinence: One Year Results.	International Urogynecology Journal
Miklos JR, Chinthakana O, et al	The IUGA_ICS classification of synthetic mesh complications in femals pelvic floor reconstructive surgery: A multicenter study2	Int Urogynecol J (2014) 25 (Suppl 1):S140-S141
Milani R, Salvatore S, et al	Functional and anatomical outcome of anterior and posterior vaginal prolapse repair with prolene mesh.	BJOG (2005)112:107-111
Miller D, Lucente V, et al	Prospective clinical assessment of the transvaginal mesh technique for treatment of pelvic organ prolapse-5-year results.	Female Pelvic Med Reconstr Surg 2011; 17: 139-143
Miller JR, Botros SM, et al	Comparing stress urinary outcomes of tension free vaginal tape with transobturator tape sling: a retrospective cohort study	Presented at ICS 2005, 35th annual meeting of the International Continence Society, Aug. 31- Sept. 2, 2005, Montreal, Canada.
Mirosh M and Epp A.	TVT vs laparoscopic Burch colposuspension for the treatment of stress urinary incontinence.	<a href="http://www.ics.org/Abstracts/Publish/43/00640.pdf">http://www.ics.org/Abstracts/Publish/43/00640. pdf</a>
Mischinger J, et al	Different surgical approaches for stress urinary incontinence in women	Minerva Ginecol 2013; 65: 21-8

Misrai V, Roupret M, et al	Surgical resection for suburethral sling complications after treatment for stress urinary incontinence	J Urol 2009 May;181(5):2198-202; discussion 2203
Moalli, P., et al	Tensile Properties of Five Commonly Used Mid- Urethral Slings Relative to the TVT	Int Urogynecol J (2008) 19:655-663
Montoya TI, et al	Anatomic relationships of pudendal nerve branches	Am J Obstet Gynecol 2011; 205: 504.e1-5
Moon, HB., et al	Occurrence and Accumulation Patterns of Polycyclic Aromatic Hydrocarbons and Synthetic Musk Compounds in Adipose Tissues of Korean Females.	Chemosphere. (2012) 86:485-90
Moore R, Mitchell G, et al	"ICS Abs 827 MiniArc single incision sling: 1 year follow-up on a new minimally invasive treatment for female SUI."	International Continence Society Mtg 2009
Moore RD, et al	Minimally Invasive Treatment for Female Stress Urinary Incontinence	Expert Rev Obstet Gynecol 3(2), 257-272 (2008)
Moore RD, Miklos JR.	Single incision mini-sling. 1 year follow-up on a new minimally invasive treatment for female SUI.	Int Urogyn J 2009;20(Suppl 3):S312-313 IUGA Abs 298
Moore WS, Hall AD	Late Suture Failure in the Pathogenesis of Anastomotic False Aneurysms	Annals of Surgery 172.6 (1970):1064-1068
Morey AF, Medendorp AR, et al	Transobturator versus transabdominal mid urethral slings: a multi-institutional comparison of obstructive voiding complications	J Urology, (2006)175:1014-1017
Moride Y, et al	Under-reporting of adverse drug reactions in general practice	Br J Clin Pharmacol 1997; 43: 177-181
Morley R and Nethercliffe J.	Minimally invasive surgical techniques for stress incontinence surgery.	Best Practice & Research Clinical Obstetrics & Gynaecology ,19:925-40,07-Nov-05
Morton HC, et al	Urethral injury associated with minimally invasive mid-urethral sling procedures for the treatment of stress urinary incontinence: a case series and systematic literature search	BJOG 2009;1 16:1120-1126
Mostafa A, Agur W, et al	A multicentre prospective randomised study of single-incision mini-sling (Ajust®) versus tension- free vaginal tape-obturator (TVT-O) in the management of female stress urinary incontinence: pain profile and short-term outcomes.	Eur J Obstet Gynecol Reprod Biol. (2012);165:115-21.



Mostafa A, et al	Single-incision mini-slugs versus standard midurethral slugs in surgical management of female stress urinary incontinence: An updated systematic review and meta-analysis of effectiveness and complications	Eur Urol 2014;65:4-2-427.
Mouzoon, N, Garome, M, et al.	Substantially Unsafe Medical Devices Pose Great Threat to Patients; Safeguards Must be Strengthened, Not Weakened	Public Citizen, February 2012
Mueller, ER	Retropubic Bladder Neck Suspensions	In H.B. Goldman (Ed.), Complications of Female Incontinence and Pelvic Reconstructive Surgery (pp. 107-114): Humana Press
Muffy TM, et al	Interventional Radiologic Treatment of pelvic hemorrhage after placement of mesh for reconstructive pelvic surgery	Obstet Gynecol 2012; 119:459-62
Muir TW, Tulikangas PK, et al	The Relationship of Tension-Free Vaginal Tape Insertion and the Vascular Anatomy	Obstetrics & Gynecology, 101:933-6, 01-May-03
Murphy M, van Raalte H, et al	"Incontinence-related quality of life and sexual function following the tension-free vaginal tape versus the ""inside-out"" tension-free vaginal tape obturator."	Int Urogynecol J (2008) 19:481-487
Murray S, et al	Urethral distortion after placement of synthetic mid urethral sling	J Urol 2011; 185(4):1321-1326
Mustafa M, et al	Bladder erosion of tension-free vaginal tape presented as vesical stone: management and review of literature	Int Urol Nephrol 2007; 39:453-455
N/A	Discrepancies between trial results reported on clinical trial registry and in journals	Medical Press medicalxpress.com March 11, 2014
N/A	IAGE News Letter 2011-2012	
N/A	Regulatory Issues: FDA urged to replace 510(k) process	OR Manager (2011) 27;9: 1
N/A	2005 IUGA Grafts Roundtable International Urogynecological Association: The Usage of Grafts in Pelvic Reconstructive Surgery Symposium 2005 July 8-10, 2005, Lago Mar Resort, Fort Lauderdale FL, USA	Int Urogynecol J (2006) 17: S1-S3
N/A	Hazardous Substances	CAW Health, Safety & Environment Fact Sheet, Aug 2011
N/A	Medical Devices and the Public's Health The FDA 510(k) Clearance Process at 35 Years	IOM Report Brief July 2011

N/A	"IUGA Sexual Dysfunction in Women Roundtable, June 6-8, 2008, Lago Mar Resort, Fort Lauderdale, Florida, USA"	Int Urogynecol J (2009) 20 (Suppl 1): S1-S2
N/A	Emedicine – Medscape.com (2013) Vaginal sling procedures	
N/A	Netdoctor.com (2013) Stress Urinary Incontinence – Pelvic Floor exercise	
N/A	FDA.gov (2013) Stress Urinary Incontinence	
N/A	Wikipedia.org (2013) Urinary Incontinence	
N/A	"American National Standard Biological evaluation of medical devices--Part 7: Ethylene oxide sterilization residuals"	ANSI/AAMI/ISO 10993-7:2008
N/A	Emedicine – Medscape.com (2013) Burch colposuspension	
N/A	WebMD.com (2013) Mechanical devices for Urinary Incontinence in Women	
N/A	Womensdoctor.com (2013) Burch Procedure and Paravaginal Repair	
Narang S, Han HC	Initial Experience of TVT-Abbrevio at a Tertiary Care Hospital	ICS Abs. 682
Narang, S, et al	TVT Abbrevio for management of female stress urinary incontinence - A prospective analysis over 22 months in a tertiary care hospital	Narang abstract poster
Naumann G, Lobodasch K, et al	Tension free vaginal tape (TVT) vs less invasive free tape (LIFT)-a randomized multicentric study of suburethral sling surgery.	<a href="http://www.ics.org/Abstracts/Publish/44/000481.pdf">http://www.ics.org/Abstracts/Publish/44/000481.pdf</a>
Nazemi TM, Yamada B, et al	Minimum 24-Month Followup of the Sling for the Treatment of Stress Urinary Incontinence	The Journal of Urology, 179:596-9, 21-Dec-07
Neuman M	TVT and TVT Obturator comparison of two operative procedures	European J Obstet & Gynecol and Reproduct Biology 131 (2007) 89-92
Neuman M, Sosnovski V, et al	Transobturator vs single-incision suburethral mini- slings for treatment of female stress urinary incontinence: early postoperative pain and 3-year follow-up.	J Minim Invasive Gynecol. (2011) 18: 769-773
Neuman M.	TVT-obturator: short-term data on an operative procedure for the cure of female stress urinary incontinence performed on 300 patients.	Eur Urol. (2007) 51: 1083-1088.

Neuman M.	Perioperative Complications and Early Follow Up with 100 TVT-SECUR Procedures	The Journal of Minimally Invasive Gynecology (2008) 15: 480-484
Neuman, M, Sosnovski V, et al	Comparison of two inside-out transobturator suburethral sling techniques for stress incontinence: Early postoperative thigh pain and 3-year outcomes	Int J Urol (2012) 19: 1103-1107
Nezhat FR, et al	Robotic-assisted laparoscopic transection and repair of an obturator nerve during pelvic lymphadenectomy for endometrial cancer	Obstet Gynecol 2012; 119: 462-4
Nguyen JN and Burchette RJ.	Outcome after anterior vaginal prolapsed repair: a randomized controlled trial.	Obstetrics & Gynecology (2008),111 (4):891-898
Nguyen JN, Jakus-Waldman SM, et al	Perioperative complications and reoperations after incontinence and prolapse surgeries using prosthetic implants.	Obstetrics & Gynecology, (2012)119(3):539-546
NICE	Urinary Incontinence. The management of urinary incontinence in women Issued: September 2013	NICE clinical guideline 171
NICE	Eighth Report of Session 2012-13 Volume II	
NICE	2013-NICE-urinary incontinence p24-Synthetic tapes	NICE clinical guideline 171 pg 24
NICE	Clinical Guideline 171 Urinary Incontinence: The management of urinary incontinence in women	NICE clinical guideline 171 (2013)
NICE	Urinary incontinence: The management of urinary incontinence in women	NICE clinical guideline 171 (2013)
NICE Guidance	<a href="http://guidance.nice.org.uk/index.jsp?action=article&amp;o=32572">http://guidance.nice.org.uk/index.jsp?action=article&amp;o=32572</a>	
Nieminen K, Hiltunen R, et al	"Outcomes after anterior vaginal wall repair with mesh: a randomized, controlled trial with a 3 year follow-up"	Am J Obstet Gynecol,(2010)203:235.e1-8
Nilsson C.G., et al	Reprint: Long-term Results of the Tension-Free Vaginal Tape (TVT) Procedure for Surgical Treatment of Female Stress Urinary Incontinence	Int Urogynecol J (2001)(Supp. 2):S5-S8 Eth.Mesh.00159481-00159484
Nilsson CG, et al	Seventeen years' follow-up of the tension-free vaginal tape procedure for female stress urinary incontinence	Int Urogynecol J 24, 1265-1269 (2013)
Nilsson CG, et al	Seven year follow up of the tension free vaginal tape procedure for treatment of urinary incontinence	Obstet Gynecol 2004; 104: 1259-62

Nilsson, C	Creating a gold standard surgical procedure: the development and implementation of TVT	Int Urogynecol J DOI 10.1007/s00192-014-2616- 2 (2015)
Nilsson, CG, et al	Eleven years prospective follow-up of the tension- free vaginal tape procedure for treatment of stress urinary incontinence2	Int Urogynecol J (2008) 19:1043-1047
Nilsson, CG., et al	Long-term Results of the Tension-Free Vaginal Tape (TVT) Procedure for Surgical Treatment of Female Stress Urinary Incontinence	Int Urogynecol J (2001) (Suppl 2): S5-S8
Noblett KL, Shen B, et al	Lynx midurethral sling system: a 1-year prospective study on efficacy and safety	5479411
Norton, P	New technology in gynecologic surgery. Is new necessarily better?	Obstet & Gynecol (2006)108(3)2:707-708
Nossier S, et al	Presentation Number: Poster 30 Safety and efficacy of the Solyx™ single-incision sling system	Female Pelvic Medicine & Reconstructive Surgery (2010) 16(5)2:S110-S1111
Novara G, et al	Tension-free midurethral slings in the treatment of female stress urinary incontinence: a systematic review and meta-analysis of randomized controlled trials of effectiveness	European Urol 52 (2007) 663-679
Novara G, Galfano A, et al	Complication rates of tension-free midurethral slings in the treatment of female stress urinary incontinence: a systematic review and meta- analysis of randomized controlled trials comparing tension-free midurethral tapes to other surgical procedures and d	Eur Urol (2008) 53: 288-309
Novara G, Galfano A, et al	Critical Assessment of Pelvic Floor Surgical Reconstruction Outcome	EAU-EBU Update Series 4 (2006) 202-213
Nygaard I, Barber MD, et al	Prevalence of symptomatic pelvic floor disorders in US women.	JAMA (2008); 300(11):1311-1316
Nygaard I, Brubaker L, et al	Long-term Outcomes Following Abdominal Sacrocolpopexy for Pelvic Organ Prolapse	JAMA (2013);309(19):2016-2024,
Nygaard IE, McCreery R, et al	Abdominal sacrocolpopexy: a comprehensive review.	Obstet & Gynecol (2004)104(4):805-823
Nygaard, I	Approval Process for Devices and Mesh for Surgical Treatment of Pelvic Organ Prolapse and Urinary Incontinence	Clinical Obstet and Gynecol (2013) 56(2): 229-231
Ogah J, Cody DJ, and Rogerson L.	Minimally invasive synthetic suburethral sling operations for stress urinary incontinence in women	Cochrane Database of Systematic Reviews ,CD006375,07-Oct-09

Ogah J, Cody DJ, and Rogerson L.	Minimally invasive synthetic suburethral sling operations for stress urinary incontinence in women: a short version Cochrane Review	Neurourology and Urodynamics(2011)30:284- 291
OHSU	Clinical trial results inconsistently reported among journals, government website _ News _ OHSU	03/31/14 Portland, Ore.
Okulu E, Kayigil O, et al	Use of three types of synthetic mesh material in sling surgery - a prospective randomized clinical trial evaluating effectiveness and complications	Scandinavian J Urology 2013; 47: 217-224
Oliveira R, Botelho F, et al	Exploratory Study Assessing Efficacy and Complications of TVT-O, TVT-Secur, and Mini-Arc: Results at 12-Month Follow-up.	European Urology (2011) 59: 940-44
Ortega-Castillo V and Neri-Ruz ES.	Chapter 15: Surgical Complications with Synthetic Materials	Urinary Incontinence (book published by InTech),241-62,02-Apr-12
Osterberg, B., et al	Effect of Suture Materials on Bacterial Survival in Infected Wounds. An Experimental Study	Acta Chir Scand. 1979; 145(7):431-4
Ostergard DR	Degradation, infection and heat effects on polypropylene mesh for pelvic implantation: what was known and when it was known	Int Urogynecol J (2011) 22:771-774
Ostergard DR	Lessons from the past: directions for the future Do new marketed surgical procedures and grafts produce ethical, personal liability, and legal concerns for physicians?	Int Urogynecol J (2007) 18:591-598
Ostergard, DR	Vaginal mesh grafts and the Food and Drug Administration	Int Urogynecol J (2010) 21: 1181-1183
Ozog Y., et al	Weight of Polypropylene Mesh is Not the Only Property Defining in Vivo Mesh Biomechanics	IUGA Abstract 47
Ozog Y., et al	Shrinkage and biomechanical evaluation of lightweight synthetics in a rabbit model for primary fascial repair	Int Urogynecol J published online: 12 May 2011
Palma P, et al	Monoprosthesis for anterior vaginal prolapse and stress urinary incontinence: midterm results of an international multicentre prospective study	Int Urogynecol J (2011) 22:1535-1541
Palma P, et al	Arcus to acrus microsling: technique and preliminary results	Int Urogynecol J Accepted 23 February 2008
Palva K, Rinne K, et al	A randomized trial comparing tension-free vaginal tape with tension-free vaginal tape-obturator: 36- month results	Int Urogynecol J(2010) 21: 1049-1055

Paraiso MF, Walters MD, et al	Laparoscopic Burch Colposuspension versus Tension-free vaginal tape: a randomized trial	Obstetrics & Gynecology (2004)104(6):1249- 1258
Parden AM, Gleason JL, et al	Incontinence outcomes in women undergoing primary and repeat midurethral sling procedures.	Obstet Gynecol (2013);121(201):273-278
Pariente, JL.	Mechanical evaluation of various suburethral tapes used for the treatment of stress urinary incontinence.	(2005) Progres en Urologie, 15 (6), pp. 1106- 1109 (French - abstract only)
Parnell, BA, et al	Gentiofemoral and Perineal Neuralgia After Transobturator Midurethral Sling	Obstet Gynecol. 2012 Feb; 119(2 Pt 2):428-31
Peoples, AJ., et al	Determination of Volatile Purgeable Alogenated Hydrocarbons in Human Adipose Tissue and Blood Serum	Bull. Environm. Contam. Toxicol. 23, 244-249 (1979)
Persson J, Teleman P, et al	"Cost-analyzes based on a prospective, randomized study comparing laparoscopic colposuspension with a tension-free vaginal tape procedure. "	Acta Obstetricia et Gynecologica Scand 2002 Nov; 81(11):1066-73
Peters A, et al	Referral Patterns for Pelvic Floor Surgical Prosthesis Complications: From Symptom Onset and Initial Treatments to Evaluation at a Tertiary Care Center	Female Pelvic Med Reconstr Surg 2015;21: 116- 120
Petri E, Ashok K	Complications of synthetic slings used in female stress urinary incontinence and applicability of the new IUGA-ICS classification <sup>2</sup>	Euro Obstet & Gynecol and Reproductive Biology J 165 (2012) 347-351
Petri E, Ashok K	Comparison of late complications of retropubic and transobturator slings in stress urinary incontinence	Int Urogynecol J (2012) 23:321-325
Petri E, et al	Comparison of late complications of retropubic and transobturator slings in stress urinary incontinence	Int Urogynecol J (2012) 23:321-325
Petros P	Letter to the Editor - A basis for long-term midurethral tape complications	Neurology and Urodynamics 2011;130: 199-200
Petros P, et al	Evolution of midurethral and other mesh slings – a critical analysis.	Neurology Urodynamics 2012;9999:1-7
Petros P, et al	A basis for long-term midurethral tape complications	Neurourology and Urodynamics (2011) 30: 199-200
Petros P, Richardson PA.	Midurethral tissue fixation system sling – a “micromethod” for cure of stress urinary incontinence – preliminary report.	Aust NZ J Obstet Gynaecol 2005;45:372-375

Petros PE and Ulmsten UI.	An integral theory of female urinary incontinence: Experimental and clinical considerations	Acta Obstet Gynecol Scand (1990) 69 Suppl 153:7-31
Petros PEP, Richardson PA.	Midurethral tissue fixation system (TFS) sling for cure of stress incontinence – 3 year results.	Int Urogynecol J 2008;19:869-871.
Pham S, Rodeheaver GT, et al	Ease of Continuous Dermal Suture Removal	J Emergence Med 1990; 8:539-543
Phé V, Zimmern P, et al	Outcome measures for SUI can we minimally agree?	World J Urol published online 20 March 2014
Phillips L, Flood CG, Schulz JA.	Case report of tension-free vaginal tape associated bowel obstruction and relationship to body habitus.	Int Urogyn J 2009;20:367-368
Pickett SD, et al	The significant morbidity of removing pelvic mesh from multiple vaginal compartments	Obstetric & Gynecol 2015;125(6): 1418-1422
Pierre, Gounon	Abstract "Histological analysis of peri prosthetic tissues of mesh explanted for complication after SUI or POP surgery"	Universite de Nice Sophia Antipolis France
Pikaart DP, et al	Laparoscopic removal of pubovaginal polypropylene tension-free tape slings	JSLS (2006) 10:220-225
Pocock ST	The pros and cons of noninferiority trials	Fundamental & Clinical Pharmacol 17 (2003) 483-490
Polichetti M, et al	"IUGA Abs. 557 SUS(suburethral support): a new technique for short suburethral sling application"	Int Urogyn J 2009;20(Suppl 3):S477-478
Porena M, Costantini E, et al	Tension-free vaginal tape versus transobturator tape as surgery for stress urinary incontinence: results of a multicentre randomised trial.	Eur Urology, (2007)52:1481-1490
Portney S	Advertising and Promotion of Medical Devices	J Health Law (2006) 39(2) 265-282
Postlethwait RW	Long-Term Comparative Study of Nonabsorbable Sutures	Ann Surg (1970) 171(6): 892-898
Postlethwait RW	Five Year Study of Tissue Reaction to Synthetic Sutures	Ann Surg 190(1):54-57 (1979)
Presthus JB, VanDrie D, et al	Short-Term Assessment of MiniArc Single Incision for Treatment of Stress Urinary Incontinence	The Journal of Minimally Invasive Gynecology,15:28S-9S Abstract 103,01-Nov-08
Propex Geotextile Systems	Propex EB-405 Durability of Polypropylene	Propex EB-405



Public Citizen	Substantially Unsafe Medical Devices Pose Great Threat to Patients: Safeguards Must be Strengthened, Not Weakened	Public Citizen, February 2012
Pukall CF, et al	Neural correlates of painful genital touch in women with vulvar vestibulitis syndrome	Pain 2005 May; 115(1-2):118-27
Rardin CR, Moore R, et al	Recurrent Thigh Abscess with Necrotizing Fasciitis from a Retained Transobturator Sling Segment	The Journal of Minimally Invasive Gynecology,16:84-7,01-Jan-09
Ravve, C.	Principles of Polymer Chemistry	2000, Revised edition
Rechberger T, et al	A randomized comparison between monofilament and multifilament tapes for stress incontinence surgery	Int Urogynecol J Pelvic Floor Dysfunct. 2003 Dec;14(6):432-6. Epub 2003 Nov 25
Rechberger T, Wrobel A, et al	"The tissue reaction to polypropylene mono-et multifilamentous tape used in surgical techniques of stress urinary incontinence treatment."	Gin Pol 2003;74(9):1008-1013
Rees PM, et al	Sexual function in men and women with neurological disorders2	Lancet 2007; 369: 512-525
Reisenauer C, Kirschniak A, et al	Transobturator tape inside-out. A minimally invasive treatment of stress urinary incontinence: surgical procedure and Anatomical considerations	European Journal of Obstetrics & Gynecology and Reproductive Biology(2006) 127: 123-129
Reisenauer C, Shiozawa T, et al	Anatomic study of prolapse surgery with nonanchored mesh and a vaginal support device	Obstet Gynecol 2010; 203:1.e1-1.e7
Renganathan A, Basu M, Duckett J.	A series of Advantage suburethral slings	J Obstet Gynaecol Aug 2011;31:521-523
Reynolds WS, et al	Obturator foramen dissection for excision of symptomatic transobturator mesh	J Urology (2012) 187: 1680-1684
Rezapour M, et al.	Tension-Free Vaginal Tape (TVT) in Stress Incontinent Women with Intrinsic Sphincter Deficiency (ISD) -- A Long-Term Follow-Up	Int Urogynecol J (2001) (Suppl 2): S12-S14
Rezapour M., et al	Tension-Free Vaginal Tape (TVT) in Women with Recurrent Stress Urinary Incontinence – A Long- term Follow up	Int Urogynecol J (2001) (Suppl 2):S9-S11
Rezapour, M., et al	Tension-Free Vaginal Tape (TVT) in Women with Mixed Urinary Incontinence -- A Long-Term Follow-Up	Int Urogynecol J (2001) (Suppl 2): S15-S18
Riachi L	A New Minimally Invasive Treatment Option for Stress Urinary Incontinence in Woman: TVT Abbrevio, a Shorter Sling with an Inside-out Transobturator Approach	Surg Technol Int (2013) 23: 176-180 ETH.MESH.12840997

Rice NT, Hu Y, et al	Pelvic mesh complications in women before and after the 2011 FDA public health notification	Female Pelvic Med Reconstr Surg (2013);19:33- 338
Richter HE, et al	Retropubic versus Transobturator Midurethral Slings for Stress Incontinence	N Engl J Med 2010 362;22
Ridgeway B, Arias B, and Barber M.	Variation of the obturator foramen and pubic arch of the female bony pelvis	The American Journal of Obstetrics & Gynecology,198:546.e1-4,01-May-08
Rigaud J, Delavierre D,et al	Management of chronic pelvic and perineal pain after suburethral tape placement for urinary incontinence	Prog Urol. 2010 Nov;20(12):1166-74. Epub 2010 Oct 20. Review. French
Rigaud J, Pothin P, et al	Functional Results After Tape Removal for Chronic Pelvic Pain Following Tension-Free Vaginal Tape or Transobturator Tape	J Urology (2010) 184: 610-615
Rinne K, Laurikainen E, et al	A randomized trial comparing TVT with TVT-O: 12- month results.	The International Urogynecology Journal,19:1049-54,29-Mar-08
Riva D, Sacca V, Tonta A, et al.	T.V.T. versus T.O.T.: A Randomized Study at 1- year Follow-up.	International Urogynecology Journal, 2006 Sept 17(2) Supp: 57-100
Robert M, et al	Patient expectations, subjective improvement and objective cure: is there a difference between the transobturator tape and the tension free vaginal tape procedure? (Abstract)	Abstract 217
Robinson D., et al	What Women Want - Their Interpretation of the Concept of Cure	Journal of Pelvic Medicine & Surgery, Vol. 9 Issue 6, pp. 273-277, 2003
Robinson GA, et al	Motor neuron target selectivity and survival after prolonged axotomy	Restor Neurol Neurosci. 2013 Jan 1;31(4):451-60
Rogers AS, et al	Physician knowledge, attitudes, and behavior related to reporting adverse drug events	Arch Intern Med 1988 Jul: 148(7): 1596-600
Rogers RG	Urinary Stress Incontinence in Women	N Engl J Med 2008; 358:1029-36
Rogo-Gupta, L., Raz, S.	Pain Complications of Mesh Surgery.	In H.B. Goldman (Ed.), Complications of Female Incontinence and Pelvic Reconstructive Surgery (pp. 87-105): Humana Press. (2013)
Rogowski A, Bienkowski, et al	Mesh retraction correlates with vaginal pain and overactive bladder symptoms after anterior vaginal mesh repair <sup>2</sup>	Int Urogynecol J (2013) 24:2087-2092
Rosch R, et al	Mesh implants In hernia repair. Inflammatory cell response in a rat model.	Eur Surg Res (2003)35(3): 161-6
Ross S, et al	Incontinence-specific quality of life measures used in trials of treatments for female urinary incontinence: a systematic review	Int Urogynecol J (2006) 17(3): 272-85

Ross S, et al	Ethical issues associated with the introduction of new surgical devices, or just because we can, doesn't mean we should	June JOGC Juin 2008 508-513
Ross S, Robert M, Lier D, et al	Surgical Management of Stress Urinary Incontinence in Women: Safety, Effectiveness, and Cost-Utility of Trans-Obturator Tape (TOT) versus Tension-free Vaginal Tape (TVT) Five Years After a Randomized Surgical Trial.	BMC Women's Health 2011, 11:34
Ross S, Robert M, Swaby C, et al	Transobturator Tape Compared With Tension- Free Vaginal Tape for Stress Incontinence <sup>2</sup>	Obstetrics & Gynecology, 114:1287-94, 01-Dec-09
Roth, Ted M.	Management of persistent groin pain after transobturator sling	Int Urogynecol J (2007) 18:1371-1373 doi: 10.1007/s00192-007-0365-1
Roumeguère T, Quackels T, et al	Trans-obturator vaginal tape (TOT) for female stress incontinence: one year follow-up in 120 patients	European Urology (2005) 48: 805-809
Russell R	Zimmern cites shortcomings in published outcomes for anti-incontinence surgery	Newsroom Center Times, March 28-31, 2011
Sand PK	Sumit Trial Outcomes: Clinical Insights Into Percutaneous Tibial Nerve Stimulation	Abstracts Female Pelvic Med & Reconstr Surg (2011) 17(5): S2
Santoro M, Gorrie TM	Ethics and the Pharmaceutical Industry	Eds New York: Cambridge University Press 2005
Sarsotti C, et al	The transobturatoric tape procedure for stress urinary incontinence: results of an Argentinean multicenter experience	Int Urogynecol J (2007) 18 (Suppl 1):S107-S244 Abst 375
Sayer T, Lim J., et al	Medium-term clinical outcomes following surgical repair for vaginal prolapse with tension-free mesh and vaginal support device	Int Urogynecol J DOI 10.1007/s00192-011-1600- 3
Scales, J.T.	Tissue reactions to synthetic materials.	Proc R Soc Med, 46:647-652.
Scheiner D, Betschart C, et al	Retropubic TVT vs. Transobturator Outside-in TOT and TVT-O: One-year results from our prospective randomized study.	Neurourology and Urodynamics Eth.Mesh.00587443-00587444
Scheiner DA, et al	Twelve months effect on voiding function of retropubic compared with outside-in and inside- out transobturator midurethral slings	Int Urogynecol J 2012 Feb;23(2):197-206
Schierlitz L, Dwyer P, et al	A prospective randomised controlled trial comparing vaginal prolapse repair with and without tensionfree vaginal tape (TVT) in women with severe genital prolapse and occult stress incontinence: long term follow up.	Int Urogynecol J (2010) 21 (Suppl 1):S1-S428

Schierlitz L, Dwyer P, et al	A randomized controlled study to compare tension free vaginal tape (TVT) and Monarc trans- obturator tape in the treatment of women with urodynamic stress incontinence (USI) and intrinsic sphincter deficiency (ISD): the three year follow up.	Int Urogynecol J (2010) 21 (Suppl 1):S1-S428
Schierlitz L, Dwyer PL, et al	Effectiveness of tension-free vaginal tape compared with transobturator tape in women with stress urinary incontinence and intrinsic sphincter deficiency: a randomized controlled trial	Obstet Gynecol (2008)112:1253-1261
Schierlitz LHE, Dwyer PL,et al	A randomised controlled study to compare tension free vaginal tape (TVT) and Monarc transobturator tape in the treatment of women with urodynamic stress incontinence (USI) and intrinsic sphincter deficiency (Abstract no. 032).	Int Urogynecol J (2010) 21 (Suppl 1):S1-S428
Schneider, H.	Long-Term Performance of Polypropylene Geosynthetics	Durability and Aging of Geosynthetics, edited by R.M. Koerner, Elsevier Applied Science, 1989, pp. 95-109
Scottish News	Mesh surgery scandal: Controversial implants to be banned by Dumfries and Galloway health board	Scottish News article
Sekiguchi Y, et al	Outpatient mid urethral tissue fixation system sling for urodynamic stress urinary incontinence: 1-year results.	J Urol 2009;182:2810-283.
Serati M, et al	TVT-O for the treatment of pure urodynamic stress incontinence: efficacy, adverse effects, and prognostic factors at 5-year follow-up	European Urology 63 (2013) 872-878
Serati M, Salvatore S, et al	Surgical treatment for female stress incontinence: What is the gold-standard procedure?	Int Urogynecol J 2009;20:619-621
Serati, M.	Tension-free Vaginal Tape for the Treatment of Urodynamic Stress Incontinence: Efficacy and Adverse Effects at 10- Year Follow-Up	Eur Urol61 (2012) 939-946
Serati, M.	The Impact of the Mid-Urethral Slings for the Treatment of Stress Urinary Incontinence on Female Sexuality.	J Sex Med 2009; 6:1534-1542

Serels S, et al	Safety and efficacy of the Solyx single-incision sling for the treatment of stress urinary incontinence. Preliminary results	Urotoday Int J. 2011 Feb;4(1):art5
Serels S, et al	Long term follow up of the Solyx single incision sling in the treatment of female stress urinary incontinence (SUI)	Open J Urology 2014, 4, 13-17
Serels S, et al	Preliminary findings with the Solyx™ single-incision sling system in female stress urinary incontinence	Int Urogynecol J (2010) 21:557-561
Serels S.	Cadaveric assessment of synthetic mid-urethral sling placement.	Open J Urol 2011;1:19-24
Shah, K, Nikolavsky, D, et al.	Surgical management of lower urinary mesh perforation after mid-urethral polypropylene mesh sling: mesh excision, urinary tract reconstruction and concomitant pubovaginal sling with autologous rectus fascia	Int Urogynecol J DOI 10.1007/s00192-013-2146- 3 (2013)
Shah, S.M.	Impact of Vaginal Surgery For Stress Urinary Incontinence On Female Sexual Function: Is The Use Of Polypropylene Mesh Detrimental?	Urology (2005) 65: 270-274
Sharifiaghdas F and Mortazavi N.	Tension-free vaginal tape and autologous rectus fascia pubovaginal sling for the treatment of urinary stress incontinence: a medium-term follow-up	Med Princ Pract 2008; 17:209-214
Shaw JS, Jeppson PC, et al	Incidence of Postoperative Thigh Pain after TVT Obturator and TVT Abbrevio	J Min Invas Gynecol Abstracts (2014) 21: S52
Shaw JS, Rardin PC, et al	Effect of TVT-O Abbrevio on Post-Operative Groin Pain	J Min Invas Gynecol Abstracts (2014) 21: S27-28
Shepherd J, Feola A, et al	Uniaxial Tensile Properties of Seven Vaginally Implanted Meshes for Pelvic Organ Prolapse	IUGA Abstract
Shepherd JP, et al	Uniaxial biomechanical properties of seven different vaginally implanted meshes for pelvic organ prolapse	Int Urogynecol J (2012) 23:613-620
Shindel AW, et al	Urethral slings placed by the transobturator approach: evolution in the technique and review of the literature	Curr Urol Rep 2005; 6(5):385-92
Simsiman AJ, Powell CR, Stratford	Suburethral sling materials: best outcome with autologous tissue	The American Journal of Obstetrics & Gynecology,193:2112-6,01-Dec-05
Sivanesan K	Comment on "tvf and tvf-obturator: comparison of two operative procedures"	Eur J Obstet Gynecol Reprod Biol 131 (2007) 87-90

Sivaslioglu AA, et al	A prospective randomized controlled trial of the transobturator tape and tissue fixation mini-sling in patients with stress urinary incontinence: 5- year results	J urology (2012) 188: 194-199
Sivaslioglu AA, et al	A prospective randomized controlled trial of the transobturator tape and tissue fixation system minisling in 80 patients with stress urinary incontinence – 3 year results	Pelviperrineology 2010;29:56-59.
Skala C, Renezeder K, et al	The IUGA ICS classification of complications of prosthesis and graft insertion	Int Urogynecol J (2011) 22:1429-1435
Skala CE, Renezeder K, et al	Mesh-complications following prolapse surgery: management and outcome	Euro Obstet & Tynecol and Reprod Biology J 159 (2011) 453-456
Skypunch, OW	Giant Papillary Conjunctivitis from an Exposed Prolene Suture	Can. J. Ophthalmology, 1986; 21(5:189-192)
Smith JJ, et al	Long-term outcomes and review of complications in 75 patients with Boston Scientific Advantage mesh in mid-urethral slings	www.bostonscientific.com
Smith PP, et al	Comparison of single-incision mid-urethral tape (Ophira™) and transobturator tape (Obtryx™) sub- urethral sling procedures for female stress urinary incontinence	JCMR 2013 5(5) 58-61
Song YF, Huang HJ, Xu B, et al	[Comparative study of tension-free vaginal tape and fascia lata for stress urinary incontinence]. [Chinese].	Zhonghua Fu Chan Ke Za Zhi. 2004 Oct;39(10):658-61. Chinese.
Songara, RK et al	Need for harmonization of labeling of medical devices: a review	J Adv Pharm Technol Res. 1020 Apr-Jun; 1(2): 127-144
Sound Urological Associates	Women's Health webpage	Sound Urology Webpage <a href="http://soundurology.com/womens-health/2/16/2015">http://soundurology.com/womens-health/2/16/2015</a>
Spinosa J, Dubuis P, and Riederer	Transobturator surgery for female stress incontinence: a comparative anatomical study of outside-in vs inside-out techniques	BJU International,100:1097-102,14-Sep-07
Stanton SL et al.	Some Reflections on Tension-Free Vaginal Tape – A New Surgical Procedure for Treatment of Female Urinary Incontinence	Int Urogynecol J (2001) (Suppl 2): S1-S2
Starr DS, Weatherford ST, et al	Suture Material as a Factor in the Occurrence of Anastomotic False Aneurysms	Arch Surg April 1979; 114: 412-415
Staskin DR, Plzak L	Synthetic Slings: Pros and Cons	Current Urology Reports 2002, 3:414-417



Steege, J.F.	Evaluation and Treatment of Dyspareunia	Obstetrics Y Gynecology: May 2009- Volume 113- Issue 5- pp 1124-1136
Stephens C, Zimmern PE	Expansion of role of web based networks to mesh complications	World J Urol published online 20 July 2014
Sternschuss, G, et al	Post-Implantation Alternations of Polypropylene in the Human	J Uro 2012;188: 27-32
Strasberg SM, et al	The accordion severity grading system of surgical complications	Annals of Surgery 2009; 250(2):177-186
Strus, M., et al	Hydrogen Peroxide Produced by Lactobacillus Species as a Regulatory Molecule for Vaginal Microflora	Med Dosw Mikrobiol 2004;56:67-77
Strus, M., et al	The In Vitro Effect of Hydrogen Peroxide on Vaginal Microbial Communities	FEMS Immunol Med Microbiol. 2006 Oct;48(1):56-63
Sung VW, Schleinitz MD,et al	Comparison of retropubic vs transobturator approach to midurethral slings: a systematic review and meta-analysis	The American Journal of Obstetrics & Gynecology,197:3-11,01-Jul-07
Svabik K, Martan A, et al	Ultrasound appearances after mesh implantation-- evidence of mesh contraction or folding?	Int Urogynecol J (2011) 22:529-533
Svenningsen, Rune	Long-term follow-up of the retropubic tension- free vaginal tape procedure	Int Urogynecol J (2013) 24:1271-1278 DOI 10.1007/s00192-013-2058-2
Sweat, S., et al	Polypropylene Mesh Tape for Stress Urinary Incontinence: Complication of Urethral Erosion and Outlet Obstruction	The Journal of Urology, Vol. 168, 144- 146, (July 2002)
Szarnicki RJ	Polypropylene Suture Fracture	Ann Thorac Surg 1985 April; 39(4):400
Tahseen S, et al	Effect of transobturator tape on overactive bladder symptoms and urge urinary incontinence in women with mixed urinary incontinence	Obstet Gynecol 2009; 113:617-23
Tamai A, et al	TVt and TOT: a comparison between these two techniques based on our clinical experience	Urologia 2008 October-December; 75(4): 232- 236
Tamussino K, Hanzal E, et al	Transobturator tapes for stress urinary incontinence: Results of the Austrian registry	Am J Obstet Gynecol 2007; 197:634e1- 634e5
Tamussino K, Tammaa A, et al	TVT vs. TVT-O for primary stress incontinence: a randomized clinical trial	Int Urogynecol J (2008) 19 (Suppl 1):S1- S166
Tamussino KF, et al	Tension-Free Vaginal Tape Operation: Results of the Austrian Registry	Obstet Gynecol 2001;98:732-6
Tarcan T, et al	Safety and Efficacy of retropubic or transobturator midurethral slings in a randomized cohort of Turkish women	Urol Int Published online Aug 20, 2014



Taub, D., et al	Complications Following Surgical Intervention for Stress Urinary Incontinence: A National Perspective	Neurourology and Urodynamics 24:659-605 (2005)
Téllez Martínez-Fornés M, Fernánde	A three year follow-up of a prospective open randomized trial to compare tension-free vaginal tape with Burch colposuspension for treatment of female stress urinary incontinence <sup>2</sup>	Actas Urologicas Españolas 2009 Nov;[Deleted Object](10):1088-96
Teo R, Moran P, Mayne C, and Tince	Randomized Trial of Tension-free Vaginal Tape and Tension-free Vaginal Tape-Obturator for Urodynamic Stress Incontinence in Women.	Journal of Urology ,185: 1350-5 ,22-Feb-11
The American College of Obstetricia	"Grants, Sponsorships, and Other Support - 2009 "	The American Congress of Obstetricians and Gynecologists website,01-Jan-09
The American College of Obstetricia	Committee Opinion No. 513: Vaginal Placement of Synthetic Mesh for Pelvic Organ Prolapse	Obstetrics & Gynecology,118:1459-64,01-Dec-11
The American College of Obstetricia	"Grants, Sponsorships, and Other Support - 2012"	The American Congress of Obstetricians and Gynecologists website,01-Mar-13
The American College of Obstetricia	"Grants, Sponsorships, and Other Support - 2010"	The American Congress of Obstetricians and Gynecologists website,01-Aug-12
The American College of Obstetricia	"Grants, Sponsorships, and Other Support - 2011"	The American Congress of Obstetricians and Gynecologists website,01-Mar-13
The British Association of Urologic	Synthetic Vaginal Tapes for Stress Incontinence Procedure-Specific Information for Patients	
Tijdink MM, Vierhout ME, et al	Surgical management of mesh-related complications after prior pelvic floor reconstructive surgery with mesh	Int Urogynecol J (2011) 22:1395-1404
Tincello D, Kenyon S, et al	Colposuspension or TVT with anterior repair for urinary incontinence and prolapse: results of and lessons from a pilot randomised patient- preference study (CARPET 1)	BJOG 2009; 116:1809-1814
Tincello DG, Botha T, et al	The TVT Worldwide Observational Registry for Long-Term Data: Safety and Efficacy of Suburethral Sling Insertion Approaches for Stress Urinary Incontinence in Women	Journal of Urology,186: 2310-5,01-Dec-11
Tipton, JS	Obturator neuropathy	Curr Rev Musculoskelet Med. 2008 Dec;1(3- 4):234-7

Tirlapur SA, Vlismas A, et al	Nerve stimulation for chronic pelvic pain and bladder pain syndrome: a systematic review	Acta Obstetricia et Gynecologica Scandinavica,,24-May-13
Tommaselli GA, D'Afiero A, et al	ICS Abs 791 Single incision tension-free vaginal tape (TVT-Secur®) in the treatment of female stress urinary incontinence.	International Continence Society Mtg 2010
Tommaselli GA, D'Afiero A, et al	Effect of Modified Surgical Technique for the Positioning of TVT-O on Post-Operative Pain	Int Urogynecol J (2011) 22 (Suppl 1) S110
Tommaselli GA, D'Afiero A, et al	Tension-free vaginal tape-O and -Secur for the treatment of stress urinary incontinence: a thirty- six-month follow-up single-blind, double-arm, randomized study.	J Minim Invasive Gynecol (2013) 20, 198-204
Tommaselli GA, D'Afiero A, et al	Efficacy of a modified technique for TVT-O positioning: a twelve-month, randomized, single- blind, multicenter, non-inferiority study.	Eur J Obstet Gynecol Reprod Biol. (2013) 167: 225-229
Tommaselli GA, Di Carlo C, et al	Efficacy and Safety of TVT-Secur in the Treatment of Female Stress Urinary Incontinence: 1-year follow-up	The International Urogynecology Journal,21:1211-7,26-May-10
Tommaselli GA, et al	Comparison of TVT-O and TVT-Abbrevio for the surgical management of female stress urinary incontinence: a 12-months preliminary study	Int J Gyencol Obstet (2012) 119: Supp 3: S504
Tommaselli GA, Formisano C, Di Car	Effects of a modified technique for TVT-O positioning on postoperative pain: single-blind randomized study.	Int Urogynecol J. (2012) 23:1293-9.
Touboul, C.	Perneal approach to vascular anatomy during transobturator cystocele repair	BJOG 2009: 116:708-712
Trivedi, et al	Understanding Female Urinary Incontinence and Master Management	S. Narayan & Sons 2014
Tseng LH, Wang AC, Lin YH, et al	Randomized comparison of the suprapubic arc sling procedure vs tension-free vaginal taping for stress incontinent women	The International Urogynecology Journal,16:230- 5,27-Oct-04
Tunn, R.	Sonomorphological Evaluation of Polypropylene Mesh Implants After Vaginal Mesh Repair in Women With Cystocele or Rectocele	Ultrasound Obstetrics Cynecol 2007; 29:449-452
Turut P, Florin P, et al	Les Complications Dues Au Prolene	Bull.Soc.Opht.France 1981: 8-9
Twiss C and Raz S.	Complications of synthetic mid-urethral slings	laborie.com,,01-Mar-08
Tzartzeva K, et al	In-Depth Nano-Investigation of Vaginal Mesh and Tape Fiber Explants in Women2	ICS Abstract 366, 2014

Tzartzeva K, Lingam D, et al	In-Depth Nano-Investigation of Vaginal Mesh and Tape Fiber Explants in Women	Study: UT SW Med Center, UT Dallas
UCLA	Clinical Updates - Surgical mesh for pelvic organ prolapse repair the subject of FDA warning	UCLA Clinical Updates webpage 9/19/2012
Ulmsten U, et al	A multicenter study of tension-free vaginal tape (TVT) for surgical treatment of stress urinary incontinence	Int Urogynecol J (1998) 9:2010-213
Ulmsten U, Henriksson L, et al	An Ambulatory Surgical Procedure Under Local Anesthesia for Treatment of Female Urinary Incontinence	The International Urogynecology Journal, 7:81- 6, 01-Jan-96
Ulmsten U.	An Introduction to Tension-Free Vaginal Tape (TVT) – A New Surgical Procedure for Treatment of Female Urinary Incontinence	Int Urogynecol J (2001)(Suppl 2):S3-S4
Ulmsten U., et al	Reprint: A Multicenter Study of Tension-Free Vaginal Tape (TVT) for Surgical Treatment of Stress Urinary Incontinence	Int Urogynecol J (1998) 9:210-213 Eth.Mesh.00145084-00145088
Ulmsten, et al	A three-year follow up of tension free vaginal tape for surgical treatment of female stress urinary incontinence	Br J Obstet Gynaecol (1999) 106: 345-350
Unger, CA, et al	Outcomes following treatment for pelvic floor mesh complications	Int Urogynecol J (2014) 25:745-749
Ustün Y, Engin-Ustün Y, et al	Tension-free vaginal tape compared with laparoscopic Burch urethropexy	J Am Assoc Gynecol Laparosc. 2003 Aug; 10(3):386-9
Valentim-lourenco A, et al	"TORP- comparing the efficacy, execution and early complications of TVT and TVT-0"	International Urogynecology Journal
Valpas A, Kivelä A, et al	Tension-free vaginal tape and laparoscopic mesh colposuspension for stress urinary incontinence	Obstetrics & Gynecology, 2004 July; 104(1):42-49
Vassallo, BJ., et al	Management of Iatrogenic Vaginal Constriction	Obstet Gynecol. 2003 Sept; 102(3): 512-20
Velemir L, Amblard J, Jet al	Urethral erosion after suburethral synthetic slings: risk factors, diagnosis, and functional outcome after surgical management	Int Urogynecol J (2008) 19: 999-1006
Vervest HAM, Bisseling TM, et al	"The prevalence of voiding difficulty after TVT, its impact on quality of life, and related risk factors."	Int Urogynecol J 2007; 18:173-182
Viereck, Volker ; Rautenberg, et al	Midurethral Sling Incision: Indications and Outcomes	Int Urogynecol J (2013) 24:645-653
Voeller, GR	New Developments in Hernia Repair	Surg Technol XI: 111-116

Vollebregt, A., et al	Bacterial Colonisation of Collagen-Coated Polypropylene Vaginal Mesh: Are Additional Intraoperative Serility Procedures Useful?	Int Urogynecol J Pelvic Floor Dysfunct. 2009 Nov;20(11):1345-51
Von Theobald P, Zimmerman CW, et al	New Techniques in Genital Prolapse Surgery	2011 XII, 310 p, Hardcover ISBN: 978-1-84885-135-4
Wadie BS, Edwan A, and Nabeeh AM.	Autologous fascial sling vs polypropylene tape at short-term followup: a prospective randomized study	J Urology, 2005 Sept; 174:990-993
Wadie BS, Mansour A, et al	Minimum 2-year follow-up of mid-urethral slings, effect on quality of life, incontinence impact and sexual function	Int Urogynecol J (2010) 21:1485-1490
Wai CY.	Surgical treatment for stress and urge urinary incontinence.	Obstet Gynecol Clin N Amer 2009;36:509-519
Wai, C.Y.	Patient Satisfaction After Midurethral Sling Surgery for Stress Urinary Incontinence.	Obstet Gynecol (2013) 121:1009-1016
Walsh CA.	TVT-Secur mini-sling for stress urinary incontinence: a review of outcomes at 12 months.	BJU International,108:652-7,14-Jul-11
Walter JE.	SOGC Technical Update No. 254: Transvaginal Mesh Procedures for Pelvic Organ Prolapse. Society of Obstetricians and Gynaecologists of Canada	Journal of Obstetrics and Gynaecology Canada,[Deleted Object]:168-74 ,01-Feb-11
Waltregny D,	New Surgical Technique for Treatment of Stress Urinary Incontinence TVT-Abbrevio: From Development to Clinical Experience	Surg Technol Int (2012) 22:149-157
Waltregny D, de Leval J.	The TVT-obturator surgical procedure for the treatment of female stress urinary incontinence: a clinical update.	Int Urogynecol J (2009) 20:337-348
Waltregny D, Gaspar Y, et al	TVT-O for the treatment of female stress urinary incontinence: Results of a prospective study after a 3-year minimum follow-up	European Urology,53:401-10,21-Aug-07
Waltregny D, Reul O, et al	Inside out transobturator vaginal tape for the treatment of female stress urinary incontinence: interim results of a prospective study after a 1- year minimum followup.	J Urol. 2006 Jun;175(6):2191-5.
Waltregny D, Thomas A, et al	Three year results of a prospective randomized trial comparing the original inside-out transobturator (TVT-O™) procedure with a modified version using a shortened tape and reduced dissection for the treatment of female SUI	Poster #254

Waltregny D, Thomas A, et al	Three year results of a prospective randomized trial comparing the original inside-out transobuturator (TVT-O™) Procedure with a modified version of using a shortened tape and reduced dissection for the treatment of female stress urinary incontinence	ICU Abstract 254
Waltregny TA, DeLeval J	One year results of a prospective randomized trial comparing the original inside-out transobuturator (TVT-O™) procedure and a modified version using a shortened tape and reduced dissection for the treatment of female stress urinary incontinence [abstract]	Int Urogynecol J (2010) 21 (Suppl 1) S219-S221 IUGA, Abst 153
Wang AC and Chen MC.	Comparison of tension-free vaginal taping versus modified Burch colposuspension on urethral obstruction: a randomized controlled trial	Neurourology and Urodynamics 22:185-190 (2003)
Wang AC, Lee L, Lin C, et al	A histologic and immunohistochemical analysis of defective vaginal healing after continence raping procedures: A prospective case-controlled pilot study	American Journal of Obstetrics and Gynecology (2004) 191, 1868-74
Wang F, Song Y and Huang H.	Prospective randomized trial of TVT and TOT as primary treatment for female stress urinary incontinence with or without pelvic organ prolapse in Southeast China	Arch Gynecol Obstet. 2010 Feb; 281(2) 279-86
Wang W, Zhu L and Lang J.	Transobturator tape procedure versus tension-free vaginal tape for treatment of stress urinary incontinence	International Journal of Gynecology & Obstetrics 104 (2009) 113-116
Wang YJ, Li FP, Wang Q, et al	Comparison of three mid-urethral tension-free tapes (TVT, TVT-O, and TVT-Secur) in the treatment of female stress urinary incontinence: 1-year follow-up.	Int Urogynecol J. 2011 Nov;22(11):1369-74. doi: 10.1007/s00192-011-1445-9. Epub 2011 May 13
Wang, AC	A microbiological and immunohistochemical analysis of periurethral and vaginal tissue in women with de novo urge symptoms after mid- urethral sling procedures—a prospective case- controlled study.	Int Urogynecol J (2008) 19:1145-1150
Ward KL and Hilton P et al	Prospective multicenter randomized trial of tension-free vaginal tape and colposuspension for primary urodynamic stress incontinence2	BMJ,325:1-7,13-Jul-02

Ward KL and Hilton P et al	A prospective multicenter randomized trial of tension-free vaginal tape and colposuspension for primary urodynamic stress incontinence: Two- year follow-up	The American Journal of Obstetrics & Gynecology,190:324-31,01-Feb-04
Ward KL and Hilton P.	Tension-free vaginal tape versus colposuspension for primary urodynamic stress incontinence: 5- year follow up	BJOG 2008;115:226-233
Weber AM, et al	The standardization of terminology for researchers in female pelvic floor disorders	Int Urogynecol J (2001) 12: 178-186
Westermann L, Brown A, et al	Delayed Presentation of an Enterocutaneous Fistula After Tension-Free Vaginal Tape Sling	Female Pelvic Reconstr Surg 2011; 17: 258-259
White , R.A., Hirose, F.M., et al.	(1981). Histopathologic observations after short- term implantation of two porous elastomers in dogs.	Biomaterials, 2:171-176.
Whiteside JL and Walters MD.	Anatomy of the obturator region: relations to a trans-obturator sling	The International Urogynecology Journal,15:223- 6,24-Feb-04
Willard, FH, et al	The Neuroanatomy of Female Pelvic Pain	Pain in Women: A Clinical Guide, Chapter 2
Williams D.	Review Biodegradation of surgical polymers	Journal of Materials Science. 1982; 17:1233-1246
Wilson, C, et al	Short-term efficacy of transobturator sling in women veterans with a history of sexual trauma	MAAUA 68th Annual Meeting Abstracts (2010)
Withagen MI, Milani AL, et al	Trocar-guided mesh compared with conventional vaginal repair in recurrent prolapsed: a randomized controlled trial.	Obstetrics & Gynecology,117:242-50 ,01-Feb-11
Wolter CE, Starkman JS, et al	Removal of transobturator midurethral sling for refractory thigh pain.	Urology (2008) 72: 461.e1-461.e3
Wood,AJ	Materials characterization and histological analysis of explanted polypropylene, PTFE, and PET hernia meshes from an individual patient	J Mater Sci: Mater Med (2013) 24:1113-1122
Wu JM, Kawasaki A, et al	"Predicting the number of women who will undergo incontinence and prolapse surgery, 2010 to 2050. "	The American Journal of Obstetrics & Gynecology,205:230.e1-5 ,02-Apr-11
Wyczolkowski, M., et al	Reoperation After Complicated Tension-Free Vaginal Tape Procedures	The Journal of Urology, Vol 166, 1004-1005 (Sept 2001)
Yang C, et al	Presentation Number: Oral Poster 18 Anchor extraction forces for single-incision slings: strength comparison in a rabbit model	Female Pelvic Med & Reconstr Surg (2011) 17(5)2:S92



Yang CH, Chan PH, Lai SK, et al	A retrospective study comparing tension-free vaginal tape and transobturator suburethral tape for surgical treatment of female stress urinary incontinence --- a preliminary report.	J Chin Med Assoc. 2007 Dec;70(12):541-4. doi: 10.1016/S1726-4901(08)70057-1.
Zahn C et al	Anatomic comparison of two transobturator tape procedures.	Obstet Gynecol 2007;109:701-6.
Zhang Y, Jiang M, Tong XW, et al	The comparison of an inexpensive-modified transobturator vaginal tape versus TVT-O procedure for the surgical treatment of female stress urinary incontinence.	Taiwan J Obstet Gynecol. 2011 Sep;50(3):318- 21. doi: 10.1016/j.tjog.2011.01.033
Zhong C, Yuan C, Guang- hui D, et al	Comparison of three kinds of midurethral slings for surgical treatment of female stress urinary incontinence.	Urologia 2010;77(1):37-42
Zhu L, Lang J, Hai N, and Wong F.	Comparing vaginal tape and transobturator tape for the treatment of mild and moderate stress incontinence <sup>2</sup>	International Journal of Gynecology & Obstetrics (2007) 99, 14-17
Zuckerman DM., et al	Medical Device Recalls and the FDA Approval Process	Arch Intern Med 2011;17(11): 1006-1011 Published Online February 14, 2011
Zugor V, Labanaris AP, et al	"TVT vs. TOT: a comparison in terms of continence results, complications and quality of life after a median follow-up of 48 months. "	Int Urol Nephrol 42, 915-20 (2010)
Zullo MA, Plotti F, et al	One-year follow-up of tension-free vaginal tape (TVT) and trans-obturator suburethral tape from inside to outside (TVT-O) for surgical treatment of female stress urinary incontinence: a prospective randomised trial.	European Urology,51:1376-82,07-Nov-06
Zumbe J., Porres D., Degiorgis PL.	Obturator and thigh abscess after transobturator tape implantation for stress urinary incontinence.	Urol Int. (2008); 81:483-485
Zwarenstein M, et al	Improving the reporting of pragmatic trials: an extension of the CONSORT statement	BMJ 2008; 337:12390
Zycynski HM, Carey MP, et al	One-year clinical outcomes after prolapse surgery with nonanchored mesh and vaginal support device	Am J Obstet Gynecol 2010; 203:587.e1-8



**Documents**

ETH.MESH.00079636
ETH.MESH.00142270
ETH.MESH.00145084
ETH.MESH.00160615
ETH.MESH.00211038
ETH.MESH.00301741
ETH.MESH.00371496
ETH.MESH.00371503
ETH.MESH.00409674
ETH.MESH.00409674
ETH.MESH.00528636
ETH.MESH.00584291
ETH.MESH.00584488
ETH.MESH.00658421
ETH.MESH.00844331
ETH.MESH.00844341
ETH.MESH.00860239
ETH.MESH.008612291
ETH.MESH.00862206
ETH.MESH.00862227
ETH.MESH.00862284
ETH.MESH.00875647
ETH.MESH.01217926
ETH.MESH.01219629
ETH.MESH.0122104
ETH.MESH.01221055
ETH.MESH.01511661
ETH.MESH.02091873
ETH.MESH.02134271
ETH.MESH.02180128
ETH.MESH.02180833
ETH.MESH.02237103
ETH.MESH.02340306
ETH.MESH.02340471
ETH.MESH.02340504
ETH.MESH.02340568

ETH.MESH.02340756
ETH.MESH.02340829
ETH.MESH.02340902
ETH.MESH.02340974
ETH.MESH.02341203
ETH.MESH.02620465
ETH.MESH.02620469
ETH.MESH.02620490
ETH.MESH.02620556
ETH.MESH.02620593
ETH.MESH.02620686
ETH.MESH.02620705
ETH.MESH.02620909
ETH.MESH.02620914
ETH.MESH.02620918
ETH.MESH.02620964
ETH.MESH.02621143
ETH.MESH.02621147
ETH.MESH.02621151
ETH.MESH.02621155
ETH.MESH.02621179
ETH.MESH.02621288
ETH.MESH.02621292
ETH.MESH.02621453
ETH.MESH.02621462
ETH.MESH.02621596
ETH.MESH.02621670
ETH.MESH.02621834
ETH.MESH.02621931
ETH.MESH.02621936
ETH.MESH.02621966
ETH.MESH.02622043
ETH.MESH.02622081
ETH.MESH.02622267
ETH.MESH.02622276
ETH.MESH.02622374
ETH.MESH.02622420
ETH.MESH.02622556
ETH.MESH.02622581

ETH.MESH.02622600
ETH.MESH.02622612
ETH.MESH.02622627
ETH.MESH.02622686
ETH.MESH.02622691
ETH.MESH.02622725
ETH.MESH.02622755
ETH.MESH.02622778
ETH.MESH.02622891
ETH.MESH.02623003
ETH.MESH.02623139
ETH.MESH.02623159
ETH.MESH.02623164
ETH.MESH.02623207
ETH.MESH.02623250
ETH.MESH.02623269
ETH.MESH.02623282
ETH.MESH.02623355
ETH.MESH.02623406
ETH.MESH.02623442
ETH.MESH.02623462
ETH.MESH.02623497
ETH.MESH.02623516
ETH.MESH.02623558
ETH.MESH.02623595
ETH.MESH.02623646
ETH.MESH.02623714
ETH.MESH.02623743
ETH.MESH.02625032
ETH.MESH.02625074
ETH.MESH.02625079
ETH.MESH.02625089
ETH.MESH.02625120
ETH.MESH.02625125
ETH.MESH.02625150
ETH.MESH.02625159
ETH.MESH.02625185
ETH.MESH.02625246
ETH.MESH.02625267

ETH.MESH.02625316
ETH.MESH.02625397
ETH.MESH.02625453
ETH.MESH.02625497
ETH.MESH.02625669
ETH.MESH.02625674
ETH.MESH.02625679
ETH.MESH.02625684
ETH.MESH.02625710
ETH.MESH.02625716
ETH.MESH.02625722
ETH.MESH.02625726
ETH.MESH.02625737
ETH.MESH.02625742
ETH.MESH.02625752
ETH.MESH.02625826
ETH.MESH.02625869
ETH.MESH.02625889
ETH.MESH.02625959
ETH.MESH.02626015
ETH.MESH.02626049
ETH.MESH.02626075
ETH.MESH.02626081
ETH.MESH.02626092
ETH.MESH.02626137
ETH.MESH.02626309
ETH.MESH.02626369
ETH.MESH.02626387
ETH.MESH.02626430
ETH.MESH.02626493
ETH.MESH.02626533
ETH.MESH.02626537
ETH.MESH.02626541
ETH.MESH.02626545
ETH.MESH.02626549
ETH.MESH.02626558
ETH.MESH.02626678
ETH.MESH.02626821
ETH.MESH.02626864

ETH.MESH.02626940
ETH.MESH.02626960
ETH.MESH.02627023
ETH.MESH.02627074
ETH.MESH.02627096
ETH.MESH.02627130
ETH.MESH.02627135
ETH.MESH.02627141
ETH.MESH.02627146
ETH.MESH.02627156
ETH.MESH.02627161
ETH.MESH.02627177
ETH.MESH.02627183
ETH.MESH.02627204
ETH.MESH.02627209
ETH.MESH.02627224
ETH.MESH.02627242
ETH.MESH.02627294
ETH.MESH.02651570
ETH.MESH.03358217
ETH.MESH.03427878
ETH.MESH.03458123
ETH.MESH.03459088
ETH.MESH.03535750
ETH.MESH.03671138
ETH.MESH.03907327
ETH.MESH.03909708
ETH.MESH.03910183
ETH.MESH.03932909
ETH.MESH.04093125
ETH.MESH.04193990
ETH.MESH.05222673
ETH.MESH.05225354
ETH.MESH.05299520
ETH.MESH.05315240
ETH.MESH.05315244
ETH.MESH.05316775
ETH.MESH.05367676
ETH.MESH.05428995

ETH.MESH.05500891
ETH.MESH.05514963
ETH.MESH.05549696
ETH.MESH.05560961
ETH.MESH.05561179
ETH.MESH.05799233
ETH.MESH.05815791
ETH.MESH.05815797
ETH.MESH.05918082
ETH.MESH.05972834
ETH.MESH.06003173
ETH.MESH.06087471
ETH.MESH.06087513
ETH.MESH.06214296
ETH.MESH.06239100
ETH.MESH.06261965
ETH.MESH.06851860
ETH.MESH.06861946
ETH.MESH.07483151
ETH.MESH.07730291
ETH.MESH.07876870
ETH.MESH.07876890
ETH.MESH.07876905
ETH.MESH.08003181
ETH.MESH.08003197
ETH.MESH.08003215
ETH.MESH.08003231
ETH.MESH.08003247
ETH.MESH.08003263
ETH.MESH.08003279
ETH.MESH.08003295
ETH.MESH.08003303
ETH.MESH.08218336
ETH.MESH.08218337
ETH.MESH.08692673
ETH.MESH.08696050
ETH.MESH.09744840
ETH.MESH.09744848
ETH.MESH.09744858

ETH.MESH.09744866
ETH.MESH.09744870
ETH.MESH.00002162
ETH.MESH.00004755
ETH.MESH.00006636
ETH.MESH.00007091
ETH.MESH.00008631
ETH.MESH.00010743
ETH.MESH.00019863
ETH.MESH.00029963
ETH.MESH.00030025
ETH.MESH.00030098
ETH.MESH.00030372
ETH.MESH.00031538
ETH.MESH.00057335
ETH.MESH.00057336
ETH.MESH.00057513
ETH.MESH.00057515
ETH.MESH.00066960
ETH.MESH.00073089
ETH.MESH.00074068
ETH.MESH.00074499
ETH.MESH.00077727
ETH.MESH.00130934



ETH.MESH.00134029
ETH.MESH.00134498
ETH.MESH.00136359
ETH.MESH.00139845
ETH.MESH.00140431
ETH.MESH.00143697
ETH.MESH.00143700
ETH.MESH.00143842
ETH.MESH.00144270
ETH.MESH.00144304
ETH.MESH.00155598
ETH.MESH.00158286
ETH.MESH.00158559
ETH.MESH.00159634
ETH.MESH.00160612
ETH.MESH.00160813
ETH.MESH.00161129
ETH.MESH.00161131
ETH.MESH.00161444
ETH.MESH.00161512
ETH.MESH.00162420
ETH.MESH.00163582
ETH.MESH.00164023
ETH.MESH.00164643

ETH.MESH.00166287
ETH.MESH.00167104
ETH.MESH.00167119
ETH.MESH.00169748
ETH.MESH.00203456
ETH.MESH.00206973
ETH.MESH.00206974
ETH.MESH.00207012
ETH.MESH.00208897
ETH.MESH.00209295
ETH.MESH.00209965
ETH.MESH.00211038
ETH.MESH.00211259
ETH.MESH.00211263
ETH.MESH.00212665
ETH.MESH.00222366
ETH.MESH.00222779
ETH.MESH.00223634
ETH.MESH.00223640
ETH.MESH.00223800
ETH.MESH.00223801
ETH.MESH.00259047
ETH.MESH.00259269
ETH.MESH.00259634

ETH.MESH.00260591
ETH.MESH.00260739
ETH.MESH.00261818
ETH.MESH.00270802
ETH.MESH.00271641
ETH.MESH.00301367
ETH.MESH.00301741
ETH.MESH.00301977
ETH.MESH.00302105
ETH.MESH.00302181
ETH.MESH.00302390
ETH.MESH.00303084
ETH.MESH.00303944
ETH.MESH.00304013
ETH.MESH.00311792
ETH.MESH.00311802
ETH.MESH.00311832
ETH.MESH.00312179
ETH.MESH.00318311
ETH.MESH.00321804
ETH.MESH.00326865
ETH.MESH.00326882
ETH.MESH.00328895
ETH.MESH.00329112

ETH.MESH.00330760
ETH.MESH.00339053
ETH.MESH.00339083
ETH.MESH.00339437
ETH.MESH.00340835
ETH.MESH.00340836
ETH.MESH.00340839
ETH.MESH.00340990
ETH.MESH.00343129
ETH.MESH.00343379
ETH.MESH.00345289
ETH.MESH.00345842
ETH.MESH.00346157
ETH.MESH.00346194
ETH.MESH.00346227
ETH.MESH.00346427
ETH.MESH.00346665
ETH.MESH.00349122
ETH.MESH.00349508
ETH.MESH.00350696
ETH.MESH.00350720
ETH.MESH.00351439
ETH.MESH.00351697
ETH.MESH.00353476

ETH.MESH.00354234
ETH.MESH.00354724
ETH.MESH.00354725
ETH.MESH.00355003
ETH.MESH.00355435
ETH.MESH.00371496
ETH.MESH.00371572
ETH.MESH.00371587
ETH.MESH.00394544
ETH.MESH.00396836
ETH.MESH.00400954
ETH.MESH.00400957
ETH.MESH.00401213
ETH.MESH.00407285
ETH.MESH.00409657
ETH.MESH.00409659
ETH.MESH.00409670
ETH.MESH.00409674
ETH.MESH.00412260
ETH.MESH.00427910
ETH.MESH.00440005
ETH.MESH.00493332
ETH.MESH.00499024
ETH.MESH.00524059

ETH.MESH.00524444
ETH.MESH.00524907
ETH.MESH.00525322
ETH.MESH.00525573
ETH.MESH.00526473
ETH.MESH.00526484
ETH.MESH.00528636
ETH.MESH.00533025
ETH.MESH.00533250
ETH.MESH.00533283
ETH.MESH.00539862
ETH.MESH.00540449
ETH.MESH.00540629
ETH.MESH.00541379
ETH.MESH.00574783
ETH.MESH.00575160
ETH.MESH.00576887
ETH.MESH.00581033
ETH.MESH.00581482
ETH.MESH.00581483
ETH.MESH.00584175
ETH.MESH.00584291
ETH.MESH.00584527
ETH.MESH.00584811

ETH.MESH.00585220
ETH.MESH.00585672
ETH.MESH.00585802
ETH.MESH.00585823
ETH.MESH.00585842
ETH.MESH.00586018
ETH.MESH.00586019
ETH.MESH.00589494
ETH.MESH.00590896
ETH.MESH.00591127
ETH.MESH.00592915
ETH.MESH.00600985
ETH.MESH.00602025
ETH.MESH.00607406
ETH.MESH.00632655
ETH.MESH.00640394
ETH.MESH.00647404
ETH.MESH.00658058
ETH.MESH.00658177
ETH.MESH.00658421
ETH.MESH.00658453
ETH.MESH.00659430
ETH.MESH.00681364
ETH.MESH.00687819



ETH.MESH.00700344
ETH.MESH.00700348
ETH.MESH.00719198
ETH.MESH.00722339
ETH.MESH.00746204
ETH.MESH.00746209
ETH.MESH.00746210
ETH.MESH.00748275
ETH.MESH.00748310
ETH.MESH.00750880
ETH.MESH.00751159
ETH.MESH.00754439
ETH.MESH.00756887
ETH.MESH.00756984
ETH.MESH.00758412
ETH.MESH.00764215
ETH.MESH.00764323
ETH.MESH.00766347
ETH.MESH.00766975
ETH.MESH.00772228
ETH.MESH.00772231
ETH.MESH.00790545
ETH.MESH.00791766
ETH.MESH.00796051

ETH.MESH.00823660
ETH.MESH.00826026
ETH.MESH.00826028
ETH.MESH.00826046
ETH.MESH.00827376
ETH.MESH.00832937
ETH.MESH.00844341
ETH.MESH.00846523
ETH.MESH.00847536
ETH.MESH.00857821
ETH.MESH.00858080
ETH.MESH.00858110
ETH.MESH.00858252
ETH.MESH.00858891
ETH.MESH.00862206
ETH.MESH.00862227
ETH.MESH.00862284
ETH.MESH.00863391
ETH.MESH.00863405
ETH.MESH.00863692
ETH.MESH.00863841
ETH.MESH.00864085
ETH.MESH.00864101
ETH.MESH.00864131

ETH.MESH.00864413
ETH.MESH.00864503
ETH.MESH.00865147
ETH.MESH.00865322
ETH.MESH.00866317
ETH.MESH.00870466
ETH.MESH.00870481
ETH.MESH.00875647
ETH.MESH.00994917
ETH.MESH.00995657
ETH.MESH.00995835
ETH.MESH.00996929
ETH.MESH.00998286
ETH.MESH.00998292
ETH.MESH.00998347
ETH.MESH.00998349
ETH.MESH.01066916
ETH.MESH.01068862
ETH.MESH.01147115
ETH.MESH.01147116
ETH.MESH.01186613
ETH.MESH.01188589
ETH.MESH.01201387
ETH.MESH.01201955

ETH.MESH.01201957
ETH.MESH.01202101
ETH.MESH.01203957
ETH.MESH.01212090
ETH.MESH.01216122
ETH.MESH.01216123
ETH.MESH.01216125
ETH.MESH.01216820
ETH.MESH.01216831
ETH.MESH.01217285
ETH.MESH.01217673
ETH.MESH.01218019
ETH.MESH.01218099
ETH.MESH.01218361
ETH.MESH.01219629
ETH.MESH.01219984
ETH.MESH.01220661
ETH.MESH.01220693
ETH.MESH.01221024
ETH.MESH.01221618
ETH.MESH.01221735
ETH.MESH.01222075
ETH.MESH.01222584
ETH.MESH.01222617

ETH.MESH.01226442
ETH.MESH.01226446
ETH.MESH.01238538
ETH.MESH.01238551
ETH.MESH.01238552
ETH.MESH.01239065
ETH.MESH.01247379
ETH.MESH.01250926
ETH.MESH.01250962
ETH.MESH.01252509
ETH.MESH.01264260
ETH.MESH.01265223
ETH.MESH.01265511
ETH.MESH.01268264
ETH.MESH.01279975
ETH.MESH.01310061
ETH.MESH.01310476
ETH.MESH.01310482
ETH.MESH.01317508
ETH.MESH.01317510
ETH.MESH.01317515
ETH.MESH.01320395
ETH.MESH.0137272
ETH.MESH.01419741

ETH.MESH.01538120
ETH.MESH.01592467
ETH.MESH.01592899
ETH.MESH.01673341
ETH.MESH.01674264
ETH.MESH.01675805
ETH.MESH.01745568
ETH.MESH.01752532
ETH.MESH.01758770
ETH.MESH.01775242
ETH.MESH.01782867
ETH.MESH.01784428
ETH.MESH.01784823
ETH.MESH.01785259
ETH.MESH.01795909
ETH.MESH.01805958
ETH.MESH.01805963
ETH.MESH.01808310
ETH.MESH.01808311
ETH.MESH.01809080
ETH.MESH.01809082
ETH.MESH.01811770
ETH.MESH.01813975
ETH.MESH.01814740

ETH.MESH.01816436
ETH.MESH.01816990
ETH.MESH.01822361
ETH.MESH.01945854
ETH.MESH.02010834
ETH.MESH.02013947
ETH.MESH.02017152
ETH.MESH.02019485
ETH.MESH.02020712
ETH.MESH.02026591
ETH.MESH.02089392
ETH.MESH.02090196
ETH.MESH.02106741
ETH.MESH.02106803
ETH.MESH.02108293
ETH.MESH.02114101
ETH.MESH.02122903
ETH.MESH.02178872
ETH.MESH.02180737
ETH.MESH.02180759
ETH.MESH.02180826
ETH.MESH.02180828
ETH.MESH.02180833
ETH.MESH.02182839



ETH.MESH.02201463
ETH.MESH.02219162
ETH.MESH.02219202
ETH.MESH.02222437
ETH.MESH.02229061
ETH.MESH.02231537
ETH.MESH.02235375
ETH.MESH.02236580
ETH.MESH.02236784
ETH.MESH.02237103
ETH.MESH.02237665
ETH.MESH.02238117
ETH.MESH.02248778
ETH.MESH.02249435
ETH.MESH.02250914
ETH.MESH.02252005
ETH.MESH.02253078
ETH.MESH.02254087
ETH.MESH.02254165
ETH.MESH.02265802
ETH.MESH.02265803
ETH.MESH.02293673
ETH.MESH.02309289
ETH.MESH.02310653

ETH.MESH.02316434
ETH.MESH.02322544
ETH.MESH.02340306
ETH.MESH.02340471
ETH.MESH.02340504
ETH.MESH.02340829
ETH.MESH.02340902
ETH.MESH.02341203
ETH.MESH.02342102
ETH.MESH.02342194
ETH.MESH.02589032
ETH.MESH.02592466
ETH.MESH.02592467
ETH.MESH.02596464
ETH.MESH.02596794
ETH.MESH.02599695
ETH.MESH.02599885
ETH.MESH.02607272
ETH.MESH.02613804
ETH.MESH.02614396
ETH.MESH.02614599
ETH.MESH.02619504
ETH.MESH.02619601
ETH.MESH.02620354

ETH.MESH.02620681
ETH.MESH.02621559
ETH.MESH.02621946
ETH.MESH.02621961
ETH.MESH.02623743
ETH.MESH.02625055
ETH.MESH.02625060
ETH.MESH.02625065
ETH.MESH.02625419
ETH.MESH.02626097
ETH.MESH.02627331
ETH.MESH.02627466
ETH.MESH.02628698
ETH.MESH.02630134
ETH.MESH.02652179
ETH.MESH.02653001
ETH.MESH.02654027
ETH.MESH.02656825
ETH.MESH.02877814
ETH.MESH.03160750
ETH.MESH.03162936
ETH.MESH.03208548
ETH.MESH.03208738
ETH.MESH.03259032

ETH.MESH.03259439
ETH.MESH.03274663
ETH.MESH.03294572
ETH.MESH.03334244
ETH.MESH.03358217
ETH.MESH.03358398
ETH.MESH.03419391
ETH.MESH.03422160
ETH.MESH.03427878
ETH.MESH.03432766
ETH.MESH.03456775
ETH.MESH.03458123
ETH.MESH.03459088
ETH.MESH.03459211
ETH.MESH.03460640
ETH.MESH.03483690
ETH.MESH.03488556
ETH.MESH.03497846
ETH.MESH.03497878
ETH.MESH.03499528
ETH.MESH.03502981
ETH.MESH.03509755
ETH.MESH.03509909
ETH.MESH.03535750

ETH.MESH.03571983
ETH.MESH.03574916
ETH.MESH.03575061
ETH.MESH.03589219
ETH.MESH.03605398
ETH.MESH.03605457
ETH.MESH.03614158
ETH.MESH.03618587
ETH.MESH.03624321
ETH.MESH.03625982
ETH.MESH.03627114
ETH.MESH.03642659
ETH.MESH.03643186
ETH.MESH.03643726
ETH.MESH.03644217
ETH.MESH.03652924
ETH.MESH.03654499
ETH.MESH.03654649
ETH.MESH.03656697
ETH.MESH.03658577
ETH.MESH.03671138
ETH.MESH.03685918
ETH.MESH.03715571
ETH.MESH.03715869

ETH.MESH.03715978
ETH.MESH.03721328
ETH.MESH.03722384
ETH.MESH.03730703
ETH.MESH.03735432
ETH.MESH.03736538
ETH.MESH.03736578
ETH.MESH.03736932
ETH.MESH.03736967
ETH.MESH.03738466
ETH.MESH.03738468
ETH.MESH.03750903
ETH.MESH.03751168
ETH.MESH.03753682
ETH.MESH.03832685
ETH.MESH.03898831
ETH.MESH.03905059
ETH.MESH.03905472
ETH.MESH.03905968
ETH.MESH.03906527
ETH.MESH.03907327
ETH.MESH.03907468
ETH.MESH.03907853
ETH.MESH.03908029

ETH.MESH.03909708
ETH.MESH.03909721
ETH.MESH.03910175
ETH.MESH.03910183
ETH.MESH.03910208
ETH.MESH.03910418
ETH.MESH.03910799
ETH.MESH.03911107
ETH.MESH.03911712
ETH.MESH.03913651
ETH.MESH.03914629
ETH.MESH.03914631
ETH.MESH.03914631E
ETH.MESH.03914909
ETH.MESH.03915380
ETH.MESH.03915567
ETH.MESH.03915588
ETH.MESH.03916715
ETH.MESH.03916716
ETH.MESH.03916905
ETH.MESH.03917298
ETH.MESH.03917309
ETH.MESH.03917375
ETH.MESH.03918253

ETH.MESH.03918352
ETH.MESH.03918552
ETH.MESH.03919143
ETH.MESH.03923426
ETH.MESH.03924530
ETH.MESH.03928696
ETH.MESH.03929173
ETH.MESH.03932909
ETH.MESH.03932912
ETH.MESH.03934952
ETH.MESH.03941617
ETH.MESH.03941621
ETH.MESH.03941623
ETH.MESH.03965159
ETH.MESH.03966039
ETH.MESH.03981288
ETH.MESH.03985932
ETH.MESH.04005863
ETH.MESH.04015102
ETH.MESH.04020134
ETH.MESH.04037600
ETH.MESH.04044797
ETH.MESH.04048515
ETH.MESH.04049320



ETH.MESH.04050265
ETH.MESH.04061003
ETH.MESH.04062405
ETH.MESH.04077109
ETH.MESH.04081189
ETH.MESH.04081871
ETH.MESH.04082973
ETH.MESH.04090122
ETH.MESH.04092868
ETH.MESH.04093125
ETH.MESH.04094863
ETH.MESH.04099233
ETH.MESH.04101014
ETH.MESH.04101817
ETH.MESH.04101824
ETH.MESH.04122262
ETH.MESH.04126728
ETH.MESH.04127133
ETH.MESH.04193990
ETH.MESH.04205632
ETH.MESH.04314739
ETH.MESH.04321393
ETH.MESH.04321397
ETH.MESH.04321401

ETH.MESH.04321405
ETH.MESH.04321409
ETH.MESH.04321413
ETH.MESH.04321418
ETH.MESH.04321436
ETH.MESH.04321454
ETH.MESH.04321472
ETH.MESH.04321488
ETH.MESH.04321670
ETH.MESH.04321682
ETH.MESH.04321694
ETH.MESH.04321706
ETH.MESH.04384185
ETH.MESH.04385229
ETH.MESH.04474763
ETH.MESH.04927339
ETH.MESH.04933406
ETH.MESH.04938298
ETH.MESH.04941016
ETH.MESH.04985249
ETH.MESH.05119622
ETH.MESH.05120364
ETH.MESH.05125293
ETH.MESH.05163323

ETH.MESH.05183409
ETH.MESH.05210364
ETH.MESH.05217971
ETH.MESH.05222673
ETH.MESH.05224295
ETH.MESH.05236223
ETH.MESH.05237034
ETH.MESH.05237336
ETH.MESH.05238373
ETH.MESH.05238382
ETH.MESH.05240144
ETH.MESH.05245427
ETH.MESH.05276086
ETH.MESH.05276184
ETH.MESH.05337217
ETH.MESH.05337225
ETH.MESH.05347751
ETH.MESH.05352721
ETH.MESH.05367673
ETH.MESH.05403236
ETH.MESH.05442881
ETH.MESH.05442973
ETH.MESH.05446127
ETH.MESH.05447475

ETH.MESH.05447481
ETH.MESH.05453719
ETH.MESH.05454207
ETH.MESH.05456117
ETH.MESH.05469908
ETH.MESH.05475773
ETH.MESH.05479411
ETH.MESH.05479535
ETH.MESH.05479717
ETH.MESH.05493965
ETH.MESH.05494064
ETH.MESH.05500891
ETH.MESH.05514963
ETH.MESH.05519476
ETH.MESH.05529274
ETH.MESH.05529653
ETH.MESH.05548122
ETH.MESH.05549696
ETH.MESH.05553782
ETH.MESH.05560961
ETH.MESH.05572526
ETH.MESH.05572669
ETH.MESH.05585033
ETH.MESH.05588123

ETH.MESH.05600730
ETH.MESH.05600916
ETH.MESH.05604390
ETH.MESH.05620358
ETH.MESH.05639356
ETH.MESH.05641096
ETH.MESH.05644163
ETH.MESH.05764101
ETH.MESH.05768705
ETH.MESH.05791132
ETH.MESH.05793690
ETH.MESH.05795299
ETH.MESH.05795309
ETH.MESH.05795322
ETH.MESH.05806931
ETH.MESH.05815791
ETH.MESH.05822684
ETH.MESH.05916450
ETH.MESH.05918082
ETH.MESH.05918776
ETH.MESH.05920616
ETH.MESH.05961197
ETH.MESH.05961204
ETH.MESH.05961212

ETH.MESH.05961304
ETH.MESH.05972834
ETH.MESH.05973195
ETH.MESH.05998775
ETH.MESH.05998803
ETH.MESH.05998805
ETH.MESH.05998807
ETH.MESH.05998811
ETH.MESH.05998816
ETH.MESH.05998819
ETH.MESH.05998821
ETH.MESH.05998835
ETH.MESH.06001408
ETH.MESH.06015196
ETH.MESH.06015198
ETH.MESH.06015227
ETH.MESH.06016054
ETH.MESH.06040171
ETH.MESH.06040657
ETH.MESH.06087471
ETH.MESH.06087513
ETH.MESH.06164409
ETH.MESH.06171801
ETH.MESH.06195201

ETH.MESH.06214296
ETH.MESH.06239100
ETH.MESH.06260647
ETH.MESH.06261965
ETH.MESH.06372356
ETH.MESH.06378084
ETH.MESH.06696367
ETH.MESH.06696589
ETH.MESH.06767981
ETH.MESH.06806078
ETH.MESH.06806078
ETH.MESH.06828907
ETH.MESH.06852118
ETH.MESH.06852120
ETH.MESH.06857127
ETH.MESH.06859904
ETH.MESH.06861473
ETH.MESH.06861946
ETH.MESH.06866919
ETH.MESH.06866920
ETH.MESH.06866921
ETH.MESH.06869163
ETH.MESH.06873447
ETH.MESH.06880021

ETH.MESH.06880472
ETH.MESH.06881576
ETH.MESH.06884249
ETH.MESH.06884516
ETH.MESH.06884728
ETH.MESH.06892171
ETH.MESH.06917699
ETH.MESH.06921531
ETH.MESH.06921562
ETH.MESH.06923868
ETH.MESH.06927231
ETH.MESH.06927248
ETH.MESH.06928076
ETH.MESH.06928168
ETH.MESH.07181044
ETH.MESH.07192929
ETH.MESH.07205369
ETH.MESH.07218087
ETH.MESH.07219684
ETH.MESH.07226377
ETH.MESH.07226914
ETH.MESH.07236294
ETH.MESH.07318311
ETH.MESH.07383730



ETH.MESH.07402878
ETH.MESH.07455220
ETH.MESH.07455424
ETH.MESH.07474296
ETH.MESH.07506983
ETH.MESH.07724068
ETH.MESH.07730291
ETH.MESH.07808480
ETH.MESH.07808484
ETH.MESH.07877085
ETH.MESH.07903520
ETH.MESH.07903682
ETH.MESH.07931874
ETH.MESH.07937826
ETH.MESH.07937979
ETH.MESH.07954703
ETH.MESH.07954867
ETH.MESH.08003197
ETH.MESH.08003215
ETH.MESH.08003231
ETH.MESH.08003247
ETH.MESH.08003263
ETH.MESH.08003279
ETH.MESH.08003295

ETH.MESH.08003303
ETH.MESH.08004035
ETH.MESH.08005908
ETH.MESH.08007248
ETH.MESH.08007502
ETH.MESH.08014324
ETH.MESH.08041930
ETH.MESH.08050183
ETH.MESH.08065931
ETH.MESH.08066401
ETH.MESH.08073801
ETH.MESH.08078799
ETH.MESH.08107153
ETH.MESH.08107933
ETH.MESH.08164248
ETH.MESH.08167644
ETH.MESH.08170224
ETH.MESH.08315779
ETH.MESH.08334245
ETH.MESH.08344659
ETH.MESH.08345895
ETH.MESH.08376560
ETH.MESH.08426862
ETH.MESH.08505071

ETH.MESH.08570968
ETH.MESH.08581280
ETH.MESH.08581412
ETH.MESH.08584142
ETH.MESH.08614017
ETH.MESH.08664680
ETH.MESH.08692660
ETH.MESH.08692670
ETH.MESH.08692673
ETH.MESH.08695896
ETH.MESH.08696050
ETH.MESH.08696084
ETH.MESH.08696085
ETH.MESH.08776231
ETH.MESH.08776497
ETH.MESH.08792102
ETH.MESH.08792936
ETH.MESH.08793207
ETH.MESH.08793552
ETH.MESH.08793554
ETH.MESH.08793646
ETH.MESH.08793648
ETH.MESH.08798099
ETH.MESH.08968369

ETH.MESH.08969368
ETH.MESH.09004550
ETH.MESH.09004554
ETH.MESH.09004555
ETH.MESH.09052531
ETH.MESH.09133724
ETH.MESH.09155883
ETH.MESH.09155909
ETH.MESH.09158424
ETH.MESH.09161482
ETH.MESH.09164480
ETH.MESH.09170211
ETH.MESH.09183969
ETH.MESH.09214438
ETH.MESH.09214439
ETH.MESH.09218058
ETH.MESH.09218059
ETH.MESH.09227438
ETH.MESH.09227440
ETH.MESH.09234951
ETH.MESH.09234953
ETH.MESH.09235084
ETH.MESH.09264884
ETH.MESH.09264884

ETH.MESH.09264945
ETH.MESH.09266465
ETH.MESH.09266659
ETH.MESH.09266668
ETH.MESH.09268506
ETH.MESH.09273600
ETH.MESH.09275875
ETH.MESH.09279161
ETH.MESH.09293114
ETH.MESH.09294125
ETH.MESH.09306898
ETH.MESH.09346417
ETH.MESH.09346419
ETH.MESH.09478633
ETH.MESH.09478684
ETH.MESH.09504558
ETH.MESH.09504568
ETH.MESH.09656790
ETH.MESH.09664947
ETH.MESH.09671620
ETH.MESH.09732998
ETH.MESH.09744840
ETH.MESH.09744848
ETH.MESH.09744858

ETH.MESH.09744866
ETH.MESH.09744870
ETH.MESH.09746373
ETH.MESH.09746615
ETH.MESH.09747632
ETH.MESH.09747724
ETH.MESH.09747728
ETH.MESH.09748842
ETH.MESH.09748848
ETH.MESH.09888187
ETH.MESH.09905181
ETH.MESH.09905193
ETH.MESH.09909020
ETH.MESH.09909642
ETH.MESH.09909830
ETH.MESH.09911296
ETH.MESH.09922406
ETH.MESH.09932848
ETH.MESH.09932902
ETH.MESH.09932908
ETH.MESH.09936426
ETH.MESH.09936503
ETH.MESH.09951087
ETH.MESH.09951106

ETH.MESH.09951746
ETH.MESH.09952163
ETH.MESH.09952168
ETH.MESH.09952714
ETH.MESH.09954485
ETH.MESH.09955374
ETH.MESH.09955464
ETH.MESH.09955474
ETH.MESH.09956434
ETH.MESH.09956613
ETH.MESH.09957926
ETH.MESH.09958050
ETH.MESH.09960437
ETH.MESH.09970762
ETH.MESH.09977270
ETH.MESH.09984519
ETH.MESH.09985777
ETH.MESH.10003595
ETH.MESH.10039685
ETH.MESH.10051284
ETH.MESH.10051331
ETH.MESH.10132609
ETH.MESH.10133116
ETH.MESH.10150515

ETH.MESH.10181793
ETH.MESH.10181921
ETH.MESH.10183005
ETH.MESH.10216874
ETH.MESH.10224489
ETH.MESH.10227358
ETH.MESH.10232709
ETH.MESH.10233144
ETH.MESH.10237693
ETH.MESH.10287104
ETH.MESH.10302266
ETH.MESH.10302268
ETH.MESH.10372554
ETH.MESH.10525611
ETH.MESH.10591939
ETH.MESH.10630324
ETH.MESH.10630803
ETH.MESH.10630809
ETH.MESH.10632641
ETH.MESH.10632650
ETH.MESH.10633520
ETH.MESH.10635251
ETH.MESH.10818812
ETH.MESH.10818814



ETH.MESH.10818815
ETH.MESH.10984358
ETH.MESH.11175841
ETH.MESH.11175843
ETH.MESH.11175844
ETH.MESH.11175863
ETH.MESH.11175864
ETH.MESH.11205022
ETH.MESH.11283949
ETH.MESH.11283974
ETH.MESH.11333804
ETH.MESH.11336474
ETH.MESH.11353422
ETH.MESH.11434264
ETH.MESH.11434367
ETH.MESH.11445493
ETH.MESH.11445930
ETH.MESH.11474337
ETH.MESH.11529265
ETH.MESH.11770891
ETH.MESH.11773498
ETH.MESH.11790162
ETH.MESH.11917445
ETH.MESH.11920108

ETH.MESH.11985160
ETH.MESH.12002262
ETH.MESH.12002601
ETH.MESH.12002845
ETH.MESH.12002847
ETH.MESH.12006763
ETH.MESH.12009027
ETH.MESH.12009095
ETH.MESH.1218594
ETH.MESH.12627553
ETH.MESH.12730858
ETH.MESH.12831407
ETH.MESH.12877116
ETH.MESH.12877118
ETH.MESH.12906504
ETH.MESH.12907174
ETH.MESH.12907175
ETH.MESH.12910023
ETH.MESH.12910030
ETH.MESH.12910111
ETH.MESH.12913351
ETH.MESH.12933182
ETH.MESH.12939705
ETH.MESH.13204508

ETH.MESH.13206130
ETH.MESH.13208194
ETH.MESH.13210344
ETH.MESH.13213760
ETH.MESH.13259844
ETH.MESH.13274846
ETH.MESH.13274855
ETH.MESH.13284086
ETH.MESH.13292806
ETH.MESH.13296239
ETH.MESH.13374555
ETH.MESH.13374559
ETH.MESH.13376661
ETH.MESH.13376756
ETH.MESH.13376759
ETH.MESH.13681528
ETH.MESH.13681529
ETH.MESH.13700031
ETH.MESH.13700033
ETH.MESH.13704931
ETH.MESH.13739540
ETH.MESH.13797826
ETH.MESH.13860322
ETH.MESH.13869615

ETH.MESH.13896042
ETH.MESH.13907355
ETH.MESH.14221357
ETH.MESH.14273633
ETH.MESH.14348386
ETH.MESH.14409737
ETH.MESH.14410478
ETH.MESH.14410703
ETH.MESH.14410846
ETH.MESH.14411026
ETH.MESH.14415287
ETH.MESH.14416076
ETH.MESH.14416182
ETH.MESH.14416898
ETH.MESH.14442958
ETH.MESH.14450438
ETH.MESH.14450971
ETH.MESH.14451057
ETH.MESH.14451060
ETH.MESH.14471186
ETH.MESH.14819286
ETH.MESH.14852589
ETH.MESH.14852591
ETH.MESH.14852593

ETH.MESH.14896228
ETH.MESH.14896230
ETH.MESH.14896233
ETH.MESH.14901753
ETH.MESH.14901754
ETH.MESH.14901756
ETH.MESH.14901758
ETH.MESH.14908783
ETH.MESH.14908784
ETH.MESH.14913573
ETH.MESH.14913576
ETH.MESH.14967283
ETH.MESH.14967284
ETH.MESH.14967286
ETH.MESH.14994654
ETH.MESH.14994657
ETH.MESH.15034561
ETH.MESH.15143734
ETH.MESH.15144988
ETH.MESH.15144996
ETH.MESH.15285672
ETH.MESH.15363068
ETH.MESH.15406846
ETH.MESH.15406857

ETH.MESH.15406860
ETH.MESH.15406862
ETH.MESH.15406864
ETH.MESH.15406867
ETH.MESH.15406869
ETH.MESH.15406871
ETH.MESH.15406874
ETH.MESH.15406877
ETH.MESH.15406880
ETH.MESH.15406882
ETH.MESH.15406884
ETH.MESH.15406886
ETH.MESH.15406888
ETH.MESH.15406890
ETH.MESH.15406893
ETH.MESH.15406895
ETH.MESH.15406897
ETH.MESH.15406900
ETH.MESH.15406903
ETH.MESH.15406906
ETH.MESH.15406910
ETH.MESH.15406913
ETH.MESH.15406916
ETH.MESH.15406920

ETH.MESH.15406922
ETH.MESH.15406924
ETH.MESH.15406927
ETH.MESH.15406929
ETH.MESH.15406931
ETH.MESH.15406933
ETH.MESH.15406935
ETH.MESH.15406937
ETH.MESH.15406939
ETH.MESH.15406942
ETH.MESH.15406944
ETH.MESH.15406946
ETH.MESH.15406948
ETH.MESH.15406950
ETH.MESH.15406952
ETH.MESH.15406954
ETH.MESH.15406956
ETH.MESH.15406958
ETH.MESH.15406961
ETH.MESH.15406963
ETH.MESH.15406965
ETH.MESH.15406967
ETH.MESH.15406969
ETH.MESH.15406971

ETH.MESH.15406972
ETH.MESH.15406973
ETH.MESH.15406974
ETH.MESH.15406975
ETH.MESH.15406976
ETH.MESH.15406977
ETH.MESH.15406978
ETH.MESH.15406979
ETH.MESH.15406982
ETH.MESH.15406985
ETH.MESH.15406987
ETH.MESH.15406989
ETH.MESH.15406990
ETH.MESH.15406992
ETH.MESH.15406994
ETH.MESH.15406998
ETH.MESH.15426052
ETH.MESH.15433760
ETH.MESH.15955438
ETH.MESH.15958336
ETH.MESH.15958396
ETH.MESH.15958400
ETH.MESH.15958405
ETH.MESH.15958408



ETH.MESH.15958410
ETH.MESH.15958433
ETH.MESH.15958445
ETH.MESH.15958452
ETH.MESH.15958470
ETH.MESH.15958478
ETH.MESH.15958481
ETH.MESH.15958486
ETH.MESH.15958492
ETH.MESH.15958495
ETH.MESH.15958503
ETH.MESH.15958508
ETH.MESH.15958510
ETH.MESH.15958512
ETH.MESH.15958518
ETH.MESH.15958524
ETH.MESH.16260624
ETH.MESH.16263696
ETH.MESH.16289560
ETH.MESH.16308087
ETH.MESH.16354541
ETH.MESH.16357097
ETH.MESH.16359412
ETH.MESH.16359413

ETH.MESH.16359598
ETH.MESH.16416002
ETH.MESH.16426660
ETH.MESH.16432550
ETH.MESH.16433747
ETH.MESH.16434349
ETH.MESH.17556486
ETH.MESH.17556496
ETH.MESH.17556498
ETH.MESH.17556511
ETH.MESH.17556512
ETH.MESH.17556513
ETH.MESH.17556538
ETH.MESH.17556556
ETH.MESH.17556577
ETH.MESH.17556578
ETH.MESH.17556580
ETH.MESH.17556582
ETH.MESH.17556583
ETH.MESH.17556591
ETH.MESH.17556598
ETH.MESH.17556601
ETH.MESH.17556602
ETH.MESH.17606501

ETH.MESH.17619399
ETH.MESH.17640736
ETH.MESH.17642669
ETH.MESH.17661336
ETH.MESH.17661347
ETH.MESH.17777759
ETH.MESH.17777763
ETH.MESH.18846146
ETH.MESH.19223769
ETH.MESH.19306944
ETH.MESH.19354118
ETH.MESH.19356913
ETH.MESH.20006789
ETH.MESH03919404
ETH-03531
ETH-03877
ETH-50330
ETH-53294
ETHMESH.CHAHAL.00000 001
ETHMESH.CHAHAL.00000 006
ETHMESH.CHAHAL.00000 028
ETHMESH.CHAHAL.00000 044
ETHMESH.CHAHAL.00000 049
ETHMESH.CHAHAL.00000 051

ETHMESH.CHAHAL.00000 052
ETHMESH.CHAHAL.00000 064
ETHMESH.OHARA.000000 01
ETHMESH.OHARA.000001 57
ETHMESH.OHARA.000003 04
ETHMESH.OHARA.000003 13
ETHMESH.OHARA.000003 15
ETHMESH.OHARA.000003 22
ETHMESH.OHARA.000003 28
ETHMESH.OHARA.000003 34
ETHMESH.OHARA.000003 40
ETHMESH.OHARA.000003 47
ETHMESH.OHARA.000003 54
ETHMESH.OHARA.000003 60
HMESH_ETH.00958003
HMESH_ETH.07269753
HMESH_ETH_00228961
HMESH_ETH_00660369
HMESH_ETH_00946830
HMESH_ETH_00958014
HMESH_ETH_01840151
HMESH_ETH_01876389
HMESH_ETH_01881060
HMESH_ETH_02512521

HMESH_ETH_02571221
HMESH_ETH_02781707
HMESH_ETH_02944363
HMESH_ETH_06033196
HMESH_ETH_06509815
HMESH_ETH_06509816
HMRDH_ETH_00133261
MIL000268
MIL00282
PM.00003.m4v
PM.00004.m4v
PM.00011.m4v
T-1499
T-3137

**Deposition Transcripts**

Arnaud (July 19 & 20, 2013) and all exhibits thereto
Barbolt (August 14 & 15, 2013) and all exhibits thereto
Batke (August 1 & 2, 2013) and all exhibits thereto
Beach (June 17, 2013) and all exhibits thereto
Beath (July 11 & 12, 2013) and all exhibits thereto
Bogardus (August 30, 2013) and all exhibits thereto
Brown (September 11 & 12, 2013) and all exhibits thereto
Burkley (May 22 & 23, 2013) and all exhibits thereto
Finley (May 13 & 14, 2013) and all exhibits thereto
Henderson (May 30 & 31, 2013) and all exhibits thereto
Hinoul (June 26 & 27, 2013) and all exhibits thereto
Honjoski (April 16 & 17, 2013) and all exhibits thereto
Holste (July 29 & 30, 2013) and all exhibits thereto
Kammerer (June 12, 2013) and all exhibits thereto
Lamont (April 3 & 4, 2013) and all exhibits thereto
Lamont (September 11, 2013) and all exhibits thereto
Lin (March 12 & 13, May 2 & 3, August 1, 2013) and all exhibits thereto
Lisa (April 25 & 26, 2013) and all exhibits thereto
Luscombe (July 29 & 30, 2013) and all exhibits thereto
Mahar (June 17 & 18, 2013) and all exhibits thereto
Mahmoud (July 15 & 16, 2013) and all exhibits thereto
Maree (July 22 & 23, 2013) and all exhibits thereto
Mittenthal (August 13 & 14, 2013) and all exhibits thereto
O'Bryan (June 6 & 7, 2013) and all exhibits thereto
Owens (June 19 & 20, 2013) and all exhibits thereto
Paine (June 13 & 14, 2013) and all exhibits thereto
Parisi (June 5 & 6, 2013) and all exhibits thereto
Pattysen (July 10 & 11, 2013) and all exhibits thereto
Robinson (July 24 & 25, 2013) and all exhibits thereto
Selman (June 20 & 21, 2013) and all exhibits thereto
Smith (May 15 & 16, June 4 & 5, August 20 & 21, 2013) and all exhibits thereto
St. Hilaire (July 11 & 12, 2013) and all exhibits thereto
Vailhe (June 20 & 21, 2013) and all exhibits thereto
Weisberg (May 30 & 31, August 9, 2013) and all exhibits thereto
Yale (August 7 & 8, 2013) and all exhibits thereto

DEPONENT	DATE
Angelini, Laura, Transcripts and Exhibits	All dates
Arnaud, Axel, MD Transcripts and Exhibits	All dates
Barbolt, Thomas A., Ph.D Transcripts and Exhibits	10/10/2012; 08/04/2013; 08/15/2013; 01/07/2014; 01/08/2014
Batke, Boris Transcripts and Exhibits	8/1-2/2013
Beath, Catherine Transcripts and Exhibits	07/11-12/2013
Burkley, Dan Transcripts and Exhibits	5/22/2013; 5/23/2013
Chen, Meng, MD Transcripts and Exhibits	10/29-30/2013
London-Brown, Allison Transcripts and Exhibits	All dates
Hart, James D., MD Transcripts and Exhibits	09/17/2013; 12/20/2013
Hellhammer, Brigitte, MD Transcripts and Exhibits	09/11-12/2013
Hinoul, Piet Transcripts and Exhibits	04/05/2012; 06/26-27/2013; 01/13-15/2014
Holste, Joerg Transcripts and Exhibits	07/29-30-2013
Horton, Ron Transcripts and Exhibits	7/1/2015
Isenberg, Richard, MD Transcripts and Exhibits	11/5/13 and 11/6/13
Divilio, Thomas Transcripts and Exhibits	All dates
Kirkemo, Aaron, Transcripts and Exhibits	All dates
Kammerer, Gene, Transcript and Exhibits	All dates
Lin, Susan, Transcripts and Exhibits	3/12-13/2013; 05/3,6/2013; 08/01/2013
Lamont, Daniel J. Transcript	4/3-4/2013; 9/11/2013
Owens, Charlotte Transcript and Exhibits	9/12/2012; 6/20/2013
Robinson, David Transcripts and Exhibits	07/24-25/2013; 09/11/2013
Selman, Renee Transcript and Exhibits	6/21/2013
Smith, Dan, Transcripts and Exhibits	05/15-16/2013; 06/04- 05/2013; 08/20-21/2013
Vailhe, Christophe, Ph.D., Transcripts and Exhibits	06/20-21/2013
Weisberg, Martin, MD Transcripts and Exhibits	05/30-31/2013; 08/09/2013
McCoy, Sheri Transcripts and Exhibits	All dates
Yale, Mark, Transcript and Exhibits	8/7/2013
Scheich, Martina Transcript and Exhibits	All dates
Trial Testimony of Piet Hinoul - Batiste v. Ethicon	3/26/14; 3/27/14; 3/28/14
Jones, Scot Transcript and Exhibits	6/9/2014
Rovner, Eric Transcript and Exhibits	All dates
Elbert, Katrina Transcript and Exhibits	12/23/2014
Elbert, Katrina Trial Transcript and Exhibits From Perry v. Ethicon	
Testimony and Exhibits from Batiste v. Ethicon Trial	

All materials contained and/or referenced in the body of this expert report.